

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082025\
 Data File : P0113088.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2025 14:27
 Operator : YP/AJ
 Sample : Q2901-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EXTERIOR WINDOW CAULK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 14:49:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 19 06:08:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.664	3.655	183.4E6	109.0E6	22.399	21.545
2)	SA Decachlor...	8.682	8.627	153.4E6	42044328	22.124	22.189
Target Compounds							
3)	L1 AR-1016-1	4.747	4.733	965298	685512	3.616	4.019
4)	L1 AR-1016-2	4.762	4.748	1118473	759539	2.786	2.948
5)	L1 AR-1016-3	4.819	4.922	3417759	1077356	13.402	7.717 #
6)	L1 AR-1016-4	4.947	4.964	1827897	8569962	8.382	73.437 #
7)	L1 AR-1016-5	5.199	5.176	16496073	8308923	74.431	58.296
8)	L2 AR-1221-1	3.863	3.869	226505	38846	2.421	0.615 #
9)	L2 AR-1221-2	3.966	3.954	2565994	2332560	37.986	49.976 #
10)	L2 AR-1221-3	4.055	4.025	2446520	1348244	11.053	9.081
11)	L3 AR-1232-1	4.055	4.025	2446520	1348244	14.060	11.484
12)	L3 AR-1232-2	4.531	4.748	1558391	759539	15.607	6.198 #
13)	L3 AR-1232-3	4.762	4.922	1118473	1077356	5.903	17.008 #
14)	L3 AR-1232-4	4.947	5.005	1827897	3883268	19.411	69.071 #
15)	L3 AR-1232-5	4.986	5.176	13029234	8308923	210.342	138.798 #
16)	L4 AR-1242-1	4.747	4.733	965298	685512	4.198	4.744
17)	L4 AR-1242-2	4.762	4.748	1118473	759539	3.255	3.502
18)	L4 AR-1242-3	4.819	4.922	3417759	1077356	15.403	9.211 #
19)	L4 AR-1242-4	4.947	5.005	1827897	3883268	10.538	33.010 #
20)	L4 AR-1242-5	5.595	5.525	16530614	58926327	82.586	395.126 #
21)	L5 AR-1248-1	4.747	4.733	965298	685512	5.533	6.147
22)	L5 AR-1248-2	4.986	4.964	13029234	8569962	53.870	53.329
23)	L5 AR-1248-3	5.199	5.005	16496073	3883268	51.734	22.888 #
24)	L5 AR-1248-4	5.553	5.176	102.3E6	8308923	204.813	41.707 #
25)	L5 AR-1248-5	5.595	5.565	16530614	10083977	48.407	47.500
26)	L6 AR-1254-1	5.553	5.525	102.3E6	58926327	189.499	186.858
27)	L6 AR-1254-2	5.702	5.673	132.7E6	74473784	268.999	265.656
28)	L6 AR-1254-3	6.106	6.074	376.6E6	205.3E6	509.633	492.456
29)	L6 AR-1254-4	6.333	6.301	185.4E6	94806282	332.860	344.956
30)	L6 AR-1254-5	6.752	6.716	818.0E6	365.5E6	1166.877	1022.437
31)	L7 AR-1260-1	6.236	6.203	358.0E6	190.0E6	802.705	745.062
32)	L7 AR-1260-2	6.426	6.392	671.7E6	308.0E6	978.142	890.953
33)	L7 AR-1260-3	6.792	6.543	799.5E6	257.5E6	1294.867	865.588 #
34)	L7 AR-1260-4	7.051	7.012	608.7E6	243.9E6	1359.196	1176.980
35)	L7 AR-1260-5	7.294	7.254	1796.0E6	599.9E6	1483.453	1280.759
36)	L8 AR-1262-1	6.792	6.755	799.5E6	342.4E6	873.093	825.864
37)	L8 AR-1262-2	7.294	7.254	1796.0E6	599.9E6	1187.704	1158.933
38)	L8 AR-1262-3	7.576	7.535	418.4E6	133.6E6	755.083	743.799
39)	L8 AR-1262-4	7.640	7.599	1090.2E6	337.0E6	1114.801	1162.452
40)	L8 AR-1262-5	8.136	8.092	348.3E6	90291430	835.536	833.475
41)	L9 AR-1268-1	7.576	7.535	418.4E6	133.6E6	243.129	262.195

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082025\
Data File : P0113088.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2025 14:27
Operator : YP/AJ
Sample : Q2901-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 14:49:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 19 06:08:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.640	7.599	1090.2E6	337.0E6	752.851	811.253
43) L9	AR-1268-3	7.846	7.803	43092595	14356072	35.723	43.968
44) L9	AR-1268-4	8.136	8.092	348.3E6	90291430	756.119	749.726
45) L9	AR-1268-5	8.429	8.378	163.7E6	41536141	51.349	53.186



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082025\
Data File : PO113088.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2025 14:27
Operator : YP/AJ
Sample : Q2901-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 14:49:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 19 06:08:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.664 min
Delta R.T.: 0.000 min
Response: 183372918
Conc: 22.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#1 Tetrachloro-m-xylene

R.T.: 3.655 min
Delta R.T.: 0.000 min
Response: 109031119
Conc: 21.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.682 min
Delta R.T.: 0.000 min
Response: 153350351
Conc: 22.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.627 min
Delta R.T.: -0.001 min
Response: 42044328
Conc: 22.19 ng/ml



#3 AR-1016-1


R.T.: 4.747 min
Delta R.T.: -0.001 min
Response: 965298
Conc: 3.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#3 AR-1016-1

R.T.: 4.733 min
Delta R.T.: 0.004 min
Response: 685512
Conc: 4.02 ng/ml




#4 AR-1016-2

R.T.: 4.762 min
Delta R.T.: -0.004 min
Response: 1118473
Conc: 2.79 ng/ml



#4 AR-1016-2


R.T.: 4.748 min
Delta R.T.: 0.001 min
Response: 759539
Conc: 2.95 ng/ml



#5 AR-1016-3


R.T.: 4.819 min
Delta R.T.: -0.005 min
Response: 3417759
Conc: 13.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#5 AR-1016-3

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 1077356
Conc: 7.72 ng/ml




#6 AR-1016-4

R.T.: 4.947 min
Delta R.T.: 0.004 min
Response: 1827897
Conc: 8.38 ng/ml



#6 AR-1016-4


R.T.: 4.964 min
Delta R.T.: 0.000 min
Response: 8569962
Conc: 73.44 ng/ml



#7 AR-1016-5


R.T.: 5.199 min
Delta R.T.: -0.001 min
Response: 16496073
Conc: 74.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#7 AR-1016-5

R.T.: 5.176 min
Delta R.T.: 0.000 min
Response: 8308923
Conc: 58.30 ng/ml




#8 AR-1221-1

R.T.: 3.863 min
Delta R.T.: -0.011 min
Response: 226505
Conc: 2.42 ng/ml



#8 AR-1221-1


R.T.: 3.869 min
Delta R.T.: 0.005 min
Response: 38846
Conc: 0.62 ng/ml



#9 AR-1221-2


R.T.: 3.966 min
Delta R.T.: 0.006 min
Response: 2565994
Conc: 37.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#9 AR-1221-2

R.T.: 3.954 min
Delta R.T.: 0.004 min
Response: 2332560
Conc: 49.98 ng/ml




#10 AR-1221-3

R.T.: 4.055 min
Delta R.T.: 0.019 min
Response: 2446520
Conc: 11.05 ng/ml



#10 AR-1221-3


R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 1348244
Conc: 9.08 ng/ml



#11 AR-1232-1


R.T.: 4.055 min
Delta R.T.: 0.019 min
Response: 2446520
Conc: 14.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#11 AR-1232-1

R.T.: 4.025 min
Delta R.T.: 0.002 min
Response: 1348244
Conc: 11.48 ng/ml




#12 AR-1232-2

R.T.: 4.531 min
Delta R.T.: 0.004 min
Response: 1558391
Conc: 15.61 ng/ml



#12 AR-1232-2


R.T.: 4.748 min
Delta R.T.: 0.000 min
Response: 759539
Conc: 6.20 ng/ml



#13 AR-1232-3


R.T.: 4.762 min
Delta R.T.: -0.004 min
Response: 1118473
Conc: 5.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#13 AR-1232-3

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 1077356
Conc: 17.01 ng/ml




#14 AR-1232-4

R.T.: 4.947 min
Delta R.T.: 0.004 min
Response: 1827897
Conc: 19.41 ng/ml



#14 AR-1232-4


R.T.: 5.005 min
Delta R.T.: -0.002 min
Response: 3883268
Conc: 69.07 ng/ml



#15 AR-1232-5


R.T.: 4.986 min
Delta R.T.: 0.000 min
Response: 13029234
Conc: 210.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#15 AR-1232-5

R.T.: 5.176 min
Delta R.T.: 0.000 min
Response: 8308923
Conc: 138.80 ng/ml




#16 AR-1242-1

R.T.: 4.747 min
Delta R.T.: -0.001 min
Response: 965298
Conc: 4.20 ng/ml



#16 AR-1242-1


R.T.: 4.733 min
Delta R.T.: 0.003 min
Response: 685512
Conc: 4.74 ng/ml



#17 AR-1242-2


R.T.: 4.762 min
Delta R.T.: -0.005 min
Response: 1118473
Conc: 3.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#17 AR-1242-2

R.T.: 4.748 min
Delta R.T.: 0.000 min
Response: 759539
Conc: 3.50 ng/ml




#18 AR-1242-3

R.T.: 4.819 min
Delta R.T.: -0.005 min
Response: 3417759
Conc: 15.40 ng/ml



#18 AR-1242-3


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 1077356
Conc: 9.21 ng/ml



#19 AR-1242-4


R.T.: 4.947 min
Delta R.T.: 0.004 min
Response: 1827897
Conc: 10.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#19 AR-1242-4

R.T.: 5.005 min
Delta R.T.: -0.002 min
Response: 3883268
Conc: 33.01 ng/ml




#20 AR-1242-5

R.T.: 5.595 min
Delta R.T.: 0.000 min
Response: 16530614
Conc: 82.59 ng/ml



#20 AR-1242-5


R.T.: 5.525 min
Delta R.T.: 0.000 min
Response: 58926327
Conc: 395.13 ng/ml



#21 AR-1248-1


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 965298
Conc: 5.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#21 AR-1248-1

R.T.: 4.733 min
Delta R.T.: 0.003 min
Response: 685512
Conc: 6.15 ng/ml




#22 AR-1248-2

R.T.: 4.986 min
Delta R.T.: 0.000 min
Response: 13029234
Conc: 53.87 ng/ml



#22 AR-1248-2


R.T.: 4.964 min
Delta R.T.: 0.000 min
Response: 8569962
Conc: 53.33 ng/ml



#23 AR-1248-3


R.T.: 5.199 min
Delta R.T.: 0.000 min
Response: 16496073
Conc: 51.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#23 AR-1248-3

R.T.: 5.005 min
Delta R.T.: -0.001 min
Response: 3883268
Conc: 22.89 ng/ml




#24 AR-1248-4

R.T.: 5.553 min
Delta R.T.: 0.000 min
Response: 102325322
Conc: 204.81 ng/ml



#24 AR-1248-4


R.T.: 5.176 min
Delta R.T.: 0.000 min
Response: 8308923
Conc: 41.71 ng/ml



#25 AR-1248-5


R.T.: 5.595 min
Delta R.T.: 0.002 min
Response: 16530614
Conc: 48.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#25 AR-1248-5

R.T.: 5.565 min
Delta R.T.: -0.001 min
Response: 10083977
Conc: 47.50 ng/ml




#26 AR-1254-1

R.T.: 5.553 min
Delta R.T.: 0.000 min
Response: 102325322
Conc: 189.50 ng/ml



#26 AR-1254-1


R.T.: 5.525 min
Delta R.T.: 0.000 min
Response: 58926327
Conc: 186.86 ng/ml



#27 AR-1254-2


R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 132738487
Conc: 269.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#27 AR-1254-2

R.T.: 5.673 min
Delta R.T.: 0.000 min
Response: 74473784
Conc: 265.66 ng/ml




#28 AR-1254-3

R.T.: 6.106 min
Delta R.T.: 0.001 min
Response: 376556451
Conc: 509.63 ng/ml



#28 AR-1254-3


R.T.: 6.074 min
Delta R.T.: 0.000 min
Response: 205316010
Conc: 492.46 ng/ml



#29 AR-1254-4


R.T.: 6.333 min
Delta R.T.: -0.001 min
Response: 185403126
Conc: 332.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#29 AR-1254-4

R.T.: 6.301 min
Delta R.T.: 0.000 min
Response: 94806282
Conc: 344.96 ng/ml




#30 AR-1254-5

R.T.: 6.752 min
Delta R.T.: 0.000 min
Response: 817950239
Conc: 1166.88 ng/ml



#30 AR-1254-5


R.T.: 6.716 min
Delta R.T.: -0.001 min
Response: 365511598
Conc: 1022.44 ng/ml



#31 AR-1260-1


R.T.: 6.236 min
Delta R.T.: 0.000 min
Response: 357990777
Conc: 802.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#31 AR-1260-1

R.T.: 6.203 min
Delta R.T.: 0.000 min
Response: 190042105
Conc: 745.06 ng/ml




#32 AR-1260-2

R.T.: 6.426 min
Delta R.T.: 0.000 min
Response: 671710944
Conc: 978.14 ng/ml



#32 AR-1260-2


R.T.: 6.392 min
Delta R.T.: 0.000 min
Response: 308012813
Conc: 890.95 ng/ml



#33 AR-1260-3


R.T.: 6.792 min
Delta R.T.: 0.000 min
Response: 799546164
Conc: 1294.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#33 AR-1260-3

R.T.: 6.543 min
Delta R.T.: -0.001 min
Response: 257455052
Conc: 865.59 ng/ml




#34 AR-1260-4

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 608749055
Conc: 1359.20 ng/ml



#34 AR-1260-4


R.T.: 7.012 min
Delta R.T.: 0.000 min
Response: 243928583
Conc: 1176.98 ng/ml



#35 AR-1260-5


R.T.: 7.294 min
Delta R.T.: 0.000 min
Response: 1795992708
Conc: 1483.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#35 AR-1260-5

R.T.: 7.254 min
Delta R.T.: -0.001 min
Response: 599945158
Conc: 1280.76 ng/ml




#36 AR-1262-1

R.T.: 6.792 min
Delta R.T.: 0.002 min
Response: 799546164
Conc: 873.09 ng/ml



#36 AR-1262-1


R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 342415001
Conc: 825.86 ng/ml



#37 AR-1262-2


R.T.: 7.294 min
Delta R.T.: 0.001 min
Response: 1795992708
Conc: 1187.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#37 AR-1262-2

R.T.: 7.254 min
Delta R.T.: -0.001 min
Response: 599945158
Conc: 1158.93 ng/ml




#38 AR-1262-3

R.T.: 7.576 min
Delta R.T.: 0.001 min
Response: 418397250
Conc: 755.08 ng/ml



#38 AR-1262-3


R.T.: 7.535 min
Delta R.T.: 0.000 min
Response: 133579173
Conc: 743.80 ng/ml



#39 AR-1262-4


R.T.: 7.640 min
Delta R.T.: 0.000 min
Response: 1090239767
Conc: 1114.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#39 AR-1262-4

R.T.: 7.599 min
Delta R.T.: -0.001 min
Response: 336996245
Conc: 1162.45 ng/ml




#40 AR-1262-5

R.T.: 8.136 min
Delta R.T.: 0.000 min
Response: 348317029
Conc: 835.54 ng/ml



#40 AR-1262-5


R.T.: 8.092 min
Delta R.T.: 0.000 min
Response: 90291430
Conc: 833.48 ng/ml



#41 AR-1268-1


R.T.: 7.576 min
Delta R.T.: -0.001 min
Response: 418397250
Conc: 243.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#41 AR-1268-1

R.T.: 7.535 min
Delta R.T.: -0.002 min
Response: 133579173
Conc: 262.20 ng/ml




#42 AR-1268-2

R.T.: 7.640 min
Delta R.T.: -0.002 min
Response: 1090239767
Conc: 752.85 ng/ml



#42 AR-1268-2


R.T.: 7.599 min
Delta R.T.: -0.003 min
Response: 336996245
Conc: 811.25 ng/ml



#43 AR-1268-3


R.T.: 7.846 min
Delta R.T.: 0.000 min
Response: 43092595
Conc: 35.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK




#43 AR-1268-3

R.T.: 7.803 min
Delta R.T.: -0.002 min
Response: 14356072
Conc: 43.97 ng/ml




#44 AR-1268-4

R.T.: 8.136 min
Delta R.T.: 0.000 min
Response: 348317029
Conc: 756.12 ng/ml



#44 AR-1268-4


R.T.: 8.092 min
Delta R.T.: 0.000 min
Response: 90291430
Conc: 749.73 ng/ml



#45 AR-1268-5

R.T.: 8.429 min
Delta R.T.: 0.000 min
Response: 163721575
Conc: 51.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXTERIOR WINDOW CAULK



#45 AR-1268-5

R.T.: 8.378 min
Delta R.T.: -0.001 min
Response: 41536141
Conc: 53.19 ng/ml