

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082123\
 Data File : P0097334.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2023 17:39
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/22/2023
 Supervised By :Ankita Jodhani 08/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 01:50:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.409	3.614	159.5E6	62638461	48.188	47.317
2) SA Decachlor...	10.344	8.626	113.9E6	52441166	52.068	45.328
Target Compounds						
3) L1 AR-1016-1	5.595	4.697	45537818	21402192	483.500	482.727m
4) L1 AR-1016-2	5.617	4.716	65109178	30018216	486.709	490.728
5) L1 AR-1016-3	5.680	4.892	40734271	16231646	468.096	496.860
6) L1 AR-1016-4	5.781	4.934	32027509	13587650	476.810	475.319
7) L1 AR-1016-5	6.082	5.148	33155120	17522705	474.886	455.769
31) L7 AR-1260-1	7.237	6.183	57358038	31951980	456.353	454.069
32) L7 AR-1260-2	7.501	6.373	63747316	37690649	473.131	451.584
33) L7 AR-1260-3	7.870	6.526	49032242	34567304	472.490	447.649
34) L7 AR-1260-4	8.102	6.998	53974317	28510059	482.760	444.899
35) L7 AR-1260-5	8.436	7.242	104.9E6	62734887	499.580	452.521

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082123\
Data File : PO097334.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2023 17:39
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/22/2023
Supervised By :Ankita Jodhani 08/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 01:50:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

