

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082219\
 Data File : P0059652.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Aug 2019 12:40
 Operator : SM/AJ
 Sample : K4398-01DL 2X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AOC-3-SOUTH-A-5DL

Manual Integrations
 APPROVED

Ankita
 8/23/2019 2:22:08 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 15:37:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.133	3.500	452287	414407	12.918	11.513m
2)	SA Decachlor...	9.613	8.365	494124	382181	9.288	9.478m
Target Compounds							
26)	L6 AR-1254-1	6.127	5.335	1617361	550113	775.186	231.939m#
27)	L6 AR-1254-2	6.342	5.478	2768043	483540	829.687	230.443m#
28)	L6 AR-1254-3	6.710	5.875	2097568	713936	591.121	212.322m#
29)	L6 AR-1254-4	6.987	6.100	811732	436674	275.581	211.790m
30)	L6 AR-1254-5	7.402	6.509	1198463	1077362	387.552m	346.979

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082219\
Data File : PO059652.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Aug 2019 12:40
Operator : SM/AJ
Sample : K4398-01DL 2X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AOC-3-SOUTH-A-5DL

Manual Integrations
APPROVED

Ankita
8/23/2019 2:22:08 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 15:37:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

