

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082219\
 Data File : P0059657.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Aug 2019 14:06
 Operator : SM/AJ
 Sample : PB122439BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB122439BL

Manual Integrations
 APPROVED

Ankita
 8/23/2019 2:22:21 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 15:41:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.133	3.501	1011951	889382	28.903	24.708
2) SA Decachlor...	9.613	8.365	1059330	780499	19.912	19.357m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082219\
Data File : P0059657.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Aug 2019 14:06
Operator : SM/AJ
Sample : PB122439BL
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB122439BL


Manual Integrations
APPROVED

Ankita
8/23/2019 2:22:21 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 15:41:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.133 min
Delta R.T.: -0.016 min
Response: 1011951
Conc: 28.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122439BL


Manual Integrations
APPROVED

Ankita
8/23/2019 2:22:21 PM




#1 Tetrachloro-m-xylene

R.T.: 3.501 min
Delta R.T.: -0.009 min
Response: 889382
Conc: 24.71 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.613 min
Delta R.T.: -0.028 min
Response: 1059330
Conc: 19.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.365 min
Delta R.T.: -0.018 min
Response: 780499
Conc: 19.36 ng/ml m