

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082219\
 Data File : P0059678.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Aug 2019 20:34
 Operator : SM/AJ
 Sample : PB122470BL
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB122470BL

Manual Integrations
 APPROVED

Ankita
 8/23/2019 2:23:35 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 01:21:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.127	3.498	888501	789277	25.377	21.927m
2) SA Decachlor...	9.605	8.363	1090505	850039	20.498	21.082

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082219\
Data File : P0059678.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Aug 2019 20:34
Operator : SM/AJ
Sample : PB122470BL
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB122470BL


Manual Integrations
APPROVED

Ankita
8/23/2019 2:23:35 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 01:21:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.127 min
Delta R.T.: -0.022 min
Response: 888501
Conc: 25.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122470BL


Manual Integrations
APPROVED

Ankita
8/23/2019 2:23:35 PM




#1 Tetrachloro-m-xylene

R.T.: 3.498 min
Delta R.T.: -0.012 min
Response: 789277
Conc: 21.93 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.605 min
Delta R.T.: 0.002 min
Response: 1090505
Conc: 20.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.363 min
Delta R.T.: -0.021 min
Response: 850039
Conc: 21.08 ng/ml