

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082219\
 Data File : P0059681.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Aug 2019 21:22
 Operator : SM/AJ
 Sample : K4473-02
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CAT-345-BUCKET-082119

Manual Integrations
 APPROVED

Ankita
 8/23/2019 2:23:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 01:22:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.127	3.499	810554	761357	23.150m	21.151
2) SA Decachlor...	9.605	8.362	1027870	824950	19.321	20.460

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082219\
Data File : PO059681.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Aug 2019 21:22
Operator : SM/AJ
Sample : K4473-02
Misc :
ALS Vial : 36 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CAT-345-BUCKET-082119


Manual Integrations
APPROVED

Ankita
8/23/2019 2:23:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 01:22:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.127 min
Delta R.T.: -0.022 min
Response: 810554
Conc: 23.15 ng/ml m

Instrument :
ECD_O
ClientSampleId :
CAT-345-BUCKET-082119


Manual Integrations
APPROVED

Ankita
8/23/2019 2:23:39 PM




#1 Tetrachloro-m-xylene

R.T.: 3.499 min
Delta R.T.: -0.011 min
Response: 761357
Conc: 21.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.605 min
Delta R.T.: 0.002 min
Response: 1027870
Conc: 19.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.362 min
Delta R.T.: -0.021 min
Response: 824950
Conc: 20.46 ng/ml