

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082219\
 Data File : P0059712.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Aug 2019 6:15
 Operator : SM/AJ
 Sample : K4459-03 5X
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HA-11

Manual Integrations
 APPROVED

Ankita
 8/23/2019 2:24:29 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 06:49:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.126	3.498	248496	221332	7.097	6.149m
2) SA Decachlor...	9.601	8.361	257456	197057	4.839m	4.887m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082219\
Data File : PO059712.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Aug 2019 6:15
Operator : SM/AJ
Sample : K4459-03 5X
Misc :
ALS Vial : 63 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
HA-11


Manual Integrations
APPROVED

Ankita
8/23/2019 2:24:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 06:49:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.126 min
Delta R.T.: -0.023 min
Response: 248496
Conc: 7.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
HA-11


Manual Integrations
APPROVED

Ankita
8/23/2019 2:24:29 PM




#1 Tetrachloro-m-xylene

R.T.: 3.498 min
Delta R.T.: -0.012 min
Response: 221332
Conc: 6.15 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.601 min
Delta R.T.: -0.002 min
Response: 257456
Conc: 4.84 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.361 min
Delta R.T.: -0.023 min
Response: 197057
Conc: 4.89 ng/ml m