

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082220\
 Data File : P0070794.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Aug 2020 3:42
 Operator : DD\AJ
 Sample : AR1242CCC500
 Misc : FOR RT STUDY
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Manual Integrations
 APPROVED

Ankita
 8/24/2020 11:30:03 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 06:28:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 21 04:17:00 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.387f	3.561	4302025	2412526	61.756	59.929
2) SA Decachlor...	10.015f	8.749	4899103	3263569	52.329	57.955
Target Compounds						
16) L4 AR-1242-1	5.691f	4.788f	1269940	813090	593.920m	603.637m
17) L4 AR-1242-2	5.713f	4.806f	1725674	1042169	549.872m	548.690m
18) L4 AR-1242-3	5.777f	4.992f	1081467	579043	571.719	575.180
19) L4 AR-1242-4	5.883f	5.091f	957627	522024	592.439	582.489
20) L4 AR-1242-5	6.652f	5.646f	1044378	762111	499.143m	563.261

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082220\
Data File : PO070794.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Aug 2020 3:42
Operator : DD\AJ
Sample : AR1242CCC500
Misc : FOR RT STUDY
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Manual Integrations
APPROVED

Ankita
8/24/2020 11:30:03 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 06:28:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 21 04:17:00 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

