

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082223\
 Data File : P0097374.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Aug 2023 12:49
 Operator : YP/AJ
 Sample : 04099-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CONC-WC-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/23/2023
 Supervised By :Ankita Jodhani 08/23/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 01:34:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 04:34:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.408	3.612	27574382	14087894	8.330	10.642m#
2) SA Decachlor...	10.342	8.622	24883618	11762102	11.372m	10.167m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082223\
Data File : P0097374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Aug 2023 12:49
Operator : YP/AJ
Sample : 04099-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_O

ClientSampleId :

CONC-WC-1

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 08/23/2023

Supervised By :Ankita Jodhani 08/23/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 01:34:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 04:34:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.408 min
Delta R.T.: 0.001 min
Response: 27574382
Conc: 8.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONC-WC-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/23/2023
Supervised By :Ankita Jodhani 08/23/2023




#1 Tetrachloro-m-xylene

R.T.: 3.612 min
Delta R.T.: -0.009 min
Response: 14087894
Conc: 10.64 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.342 min
Delta R.T.: 0.003 min
Response: 24883618
Conc: 11.37 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.622 min
Delta R.T.: -0.022 min
Response: 11762102
Conc: 10.17 ng/ml m