

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082319\
 Data File : P0059721.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Aug 2019 8:42
 Operator : SM/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 12:23:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.126	3.498	817324	756398	23.344	21.014
2)	SA Decachlor...	9.601	8.360	1007187	818546	18.932	20.301
Target Compounds							
3)	L1 AR-1016-1	5.278	0.000	1014	0	0.632	N.D. #
4)	L1 AR-1016-2	5.301	0.000	1497	0	0.626	N.D. #
5)	L1 AR-1016-3	5.349	4.776	2199	134327	1.476	124.164 #
6)	L1 AR-1016-4	5.456	4.776	551	134327	0.456	148.758 #
7)	L1 AR-1016-5	5.757	5.031	36083	-3019	28.099	N.D. #
8)	L2 AR-1221-1	4.370	3.661	1022	6667	2.414	16.609 #
9)	L2 AR-1221-2	4.466	3.791	536	10238	1.650	33.607 #
10)	L2 AR-1221-3	4.521	0.000	1308	0	1.243	N.D. #
11)	L3 AR-1232-1	4.521	0.000	1308	0	1.049	N.D. #
12)	L3 AR-1232-2	5.054	0.000	443	0	0.627	N.D. #
13)	L3 AR-1232-3	5.301	4.776	1497	134327	0.978	197.276 #
14)	L3 AR-1232-4	5.475	4.846	581	79759	0.742	130.740 #
15)	L3 AR-1232-5	5.585	5.031	10230	-3019	17.018	N.D. #
16)	L4 AR-1242-1	5.301	0.000	1497	0	1.179	N.D. #
17)	L4 AR-1242-2	5.301	0.000	1497	0	0.807	N.D. #
18)	L4 AR-1242-3	5.349	4.776	2199	134327	1.881	160.774 #
19)	L4 AR-1242-4	5.475	4.846	581	79759	0.600	95.590 #
20)	L4 AR-1242-5	6.184	5.352	70796	148051	56.339	132.354 #
21)	L5 AR-1248-1	5.278	0.000	1014	0	1.011	N.D. #
22)	L5 AR-1248-2	5.550	4.776	6235	134327	4.298	116.490 #
23)	L5 AR-1248-3	5.757	4.846	36083	79759	21.710	67.151 #
24)	L5 AR-1248-4	6.121	5.031	249162	-3019	121.893	N.D. #
25)	L5 AR-1248-5	6.184	5.352	70796	148051	35.700	103.439 #
26)	L6 AR-1254-1	6.121	5.352	249162	148051	119.421	62.421 #
27)	L6 AR-1254-2	6.335	5.532	117315	201259	35.164	95.915 #
28)	L6 AR-1254-3	6.692	5.872	118720	156049	33.457	46.409 #
29)	L6 AR-1254-4	6.978	0.000	206426	0	70.081	N.D. #
30)	L6 AR-1254-5	7.361	6.505	113656	32738	36.753	10.544 #
31)	L7 AR-1260-1	6.876	5.995	97581	96452	36.197	42.324
32)	L7 AR-1260-2	7.111	6.184	274751	300356	74.047	102.515 #
33)	L7 AR-1260-3	7.471	6.331	100660	47355	30.723	17.876 #
34)	L7 AR-1260-4	7.693	6.794	141358	5482	46.025	2.423 #
35)	L7 AR-1260-5	7.998	7.036	740638	16121	104.596	2.929 #
36)	L8 AR-1262-1	7.471	6.537	100660	6283	20.742	1.685 #
37)	L8 AR-1262-2	7.998	7.036	740638	16121	95.989	2.742 #
38)	L8 AR-1262-3	0.000	7.312	0	14232	N.D.	5.601 #
39)	L8 AR-1262-4	0.000	7.378	0	10170	N.D.	2.301 #
40)	L8 AR-1262-5	8.968	0.000	1108434	0	431.875	N.D. #
41)	L9 AR-1268-1	0.000	7.312	0	14232	N.D.	1.999 #
42)	L9 AR-1268-2	0.000	7.378	0	10170	N.D.	1.577 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082319\
 Data File : P0059721.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Aug 2019 8:42
 Operator : SM/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 12:23:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	8.570	0.000	1823461	0	258.691	N.D. #
44) L9	AR-1268-4	8.968	0.000	1108434	0	398.711	N.D. #
45) L9	AR-1268-5	9.356	8.136	36323	7681	1.970	0.529 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082319\
 Data File : PO059721.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Aug 2019 8:42
 Operator : SM/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 12:23:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.126 min
Delta R.T.: -0.023 min
Response: 817324
Conc: 23.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.498 min
Delta R.T.: -0.012 min
Response: 756398
Conc: 21.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.601 min
Delta R.T.: -0.002 min
Response: 1007187
Conc: 18.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.360 min
Delta R.T.: -0.024 min
Response: 818546
Conc: 20.30 ng/ml



#3 AR-1016-1


R.T.: 5.278 min
Delta R.T.: -0.002 min
Response: 1014
Conc: 0.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.550 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.301 min
Delta R.T.: 0.001 min
Response: 1497
Conc: 0.63 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.570 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.349 min
Delta R.T.: -0.011 min
Response: 2199
Conc: 1.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#5 AR-1016-3

R.T.: 4.776 min
Delta R.T.: 0.036 min
Response: 134327
Conc: 124.16 ng/ml




#6 AR-1016-4

R.T.: 5.456 min
Delta R.T.: -0.004 min
Response: 551
Conc: 0.46 ng/ml



#6 AR-1016-4


R.T.: 4.776 min
Delta R.T.: -0.004 min
Response: 134327
Conc: 148.76 ng/ml



#7 AR-1016-5


R.T.: 5.757 min
Delta R.T.: 0.017 min
Response: 36083
Conc: 28.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#7 AR-1016-5

R.T.: 5.031 min
Delta R.T.: 0.041 min
Response: -3019
Conc: N.D.




#8 AR-1221-1

R.T.: 4.370 min
Delta R.T.: 0.010 min
Response: 1022
Conc: 2.41 ng/ml



#8 AR-1221-1


R.T.: 3.661 min
Delta R.T.: -0.055 min
Response: 6667
Conc: 16.61 ng/ml



#9 AR-1221-2


R.T.: 4.466 min
Delta R.T.: 0.020 min
Response: 536
Conc: 1.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 3.791 min
Delta R.T.: -0.010 min
Response: 10238
Conc: 33.61 ng/ml




#10 AR-1221-3

R.T.: 4.521 min
Delta R.T.: 0.000 min
Response: 1308
Conc: 1.24 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.874 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.521 min
Delta R.T.: 0.000 min
Response: 1308
Conc: 1.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.874 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.054 min
Delta R.T.: 0.016 min
Response: 443
Conc: 0.63 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.581 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.301 min
Delta R.T.: -0.020 min
Response: 1497
Conc: 0.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#13 AR-1232-3

R.T.: 4.776 min
Delta R.T.: 0.023 min
Response: 134327
Conc: 197.28 ng/ml




#14 AR-1232-4

R.T.: 5.475 min
Delta R.T.: -0.007 min
Response: 581
Conc: 0.74 ng/ml



#14 AR-1232-4


R.T.: 4.846 min
Delta R.T.: 0.012 min
Response: 79759
Conc: 130.74 ng/ml



#15 AR-1232-5


R.T.: 5.585 min
Delta R.T.: 0.017 min
Response: 10230
Conc: 17.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#15 AR-1232-5

R.T.: 5.031 min
Delta R.T.: 0.028 min
Response: -3019
Conc: N.D.




#16 AR-1242-1

R.T.: 5.301 min
Delta R.T.: 0.000 min
Response: 1497
Conc: 1.18 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.564 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.301 min
Delta R.T.: -0.021 min
Response: 1497
Conc: 0.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.581 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.349 min
Delta R.T.: -0.034 min
Response: 2199
Conc: 1.88 ng/ml



#18 AR-1242-3


R.T.: 4.776 min
Delta R.T.: 0.023 min
Response: 134327
Conc: 160.77 ng/ml



#19 AR-1242-4


R.T.: 5.475 min
Delta R.T.: -0.007 min
Response: 581
Conc: 0.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#19 AR-1242-4

R.T.: 4.846 min
Delta R.T.: 0.012 min
Response: 79759
Conc: 95.59 ng/ml




#20 AR-1242-5

R.T.: 6.184 min
Delta R.T.: -0.022 min
Response: 70796
Conc: 56.34 ng/ml



#20 AR-1242-5


R.T.: 5.352 min
Delta R.T.: 0.004 min
Response: 148051
Conc: 132.35 ng/ml



#21 AR-1248-1


R.T.: 5.278 min
Delta R.T.: -0.002 min
Response: 1014
Conc: 1.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.550 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.550 min
Delta R.T.: 0.010 min
Response: 6235
Conc: 4.30 ng/ml



#22 AR-1248-2


R.T.: 4.776 min
Delta R.T.: -0.004 min
Response: 134327
Conc: 116.49 ng/ml



#23 AR-1248-3


R.T.: 5.757 min
Delta R.T.: 0.017 min
Response: 36083
Conc: 21.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#23 AR-1248-3

R.T.: 4.846 min
Delta R.T.: 0.026 min
Response: 79759
Conc: 67.15 ng/ml




#24 AR-1248-4

R.T.: 6.121 min
Delta R.T.: -0.019 min
Response: 249162
Conc: 121.89 ng/ml



#24 AR-1248-4


R.T.: 5.031 min
Delta R.T.: 0.041 min
Response: -3019
Conc: N.D.



#25 AR-1248-5


R.T.: 6.184 min
Delta R.T.: 0.004 min
Response: 70796
Conc: 35.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#25 AR-1248-5

R.T.: 5.352 min
Delta R.T.: -0.018 min
Response: 148051
Conc: 103.44 ng/ml




#26 AR-1254-1

R.T.: 6.121 min
Delta R.T.: 0.011 min
Response: 249162
Conc: 119.42 ng/ml



#26 AR-1254-1


R.T.: 5.352 min
Delta R.T.: 0.022 min
Response: 148051
Conc: 62.42 ng/ml



#27 AR-1254-2


R.T.: 6.335 min
Delta R.T.: 0.005 min
Response: 117315
Conc: 35.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#27 AR-1254-2

R.T.: 5.532 min
Delta R.T.: 0.052 min
Response: 201259
Conc: 95.92 ng/ml




#28 AR-1254-3

R.T.: 6.692 min
Delta R.T.: -0.008 min
Response: 118720
Conc: 33.46 ng/ml



#28 AR-1254-3


R.T.: 5.872 min
Delta R.T.: -0.008 min
Response: 156049
Conc: 46.41 ng/ml



#29 AR-1254-4


R.T.: 6.978 min
Delta R.T.: -0.002 min
Response: 206426
Conc: 70.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.100 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.361 min
Delta R.T.: -0.029 min
Response: 113656
Conc: 36.75 ng/ml



#30 AR-1254-5


R.T.: 6.505 min
Delta R.T.: -0.005 min
Response: 32738
Conc: 10.54 ng/ml



#31 AR-1260-1


R.T.: 6.876 min
Delta R.T.: 0.026 min
Response: 97581
Conc: 36.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#31 AR-1260-1

R.T.: 5.995 min
Delta R.T.: -0.005 min
Response: 96452
Conc: 42.32 ng/ml




#32 AR-1260-2

R.T.: 7.111 min
Delta R.T.: 0.000 min
Response: 274751
Conc: 74.05 ng/ml



#32 AR-1260-2


R.T.: 6.184 min
Delta R.T.: 0.004 min
Response: 300356
Conc: 102.51 ng/ml



#33 AR-1260-3


R.T.: 7.471 min
Delta R.T.: 0.001 min
Response: 100660
Conc: 30.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#33 AR-1260-3

R.T.: 6.331 min
Delta R.T.: 0.000 min
Response: 47355
Conc: 17.88 ng/ml




#34 AR-1260-4

R.T.: 7.693 min
Delta R.T.: 0.003 min
Response: 141358
Conc: 46.03 ng/ml



#34 AR-1260-4


R.T.: 6.794 min
Delta R.T.: -0.006 min
Response: 5482
Conc: 2.42 ng/ml



#35 AR-1260-5


R.T.: 7.998 min
Delta R.T.: -0.002 min
Response: 740638
Conc: 104.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#35 AR-1260-5

R.T.: 7.036 min
Delta R.T.: -0.004 min
Response: 16121
Conc: 2.93 ng/ml




#36 AR-1262-1

R.T.: 7.471 min
Delta R.T.: -0.018 min
Response: 100660
Conc: 20.74 ng/ml



#36 AR-1262-1


R.T.: 6.537 min
Delta R.T.: -0.022 min
Response: 6283
Conc: 1.69 ng/ml



#37 AR-1262-2


R.T.: 7.998 min
Delta R.T.: -0.025 min
Response: 740638
Conc: 95.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#37 AR-1262-2

R.T.: 7.036 min
Delta R.T.: -0.021 min
Response: 16121
Conc: 2.74 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.303 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.312 min
Delta R.T.: -0.020 min
Response: 14232
Conc: 5.60 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 8.386 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#39 AR-1262-4

R.T.: 7.378 min
Delta R.T.: -0.019 min
Response: 10170
Conc: 2.30 ng/ml




#40 AR-1262-5

R.T.: 8.968 min
Delta R.T.: -0.018 min
Response: 1108434
Conc: 431.87 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 7.889 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T. : 8.302 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#41 AR-1268-1

R.T.: 7.312 min
Delta R.T.: -0.018 min
Response: 14232
Conc: 2.00 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.390 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.378 min
Delta R.T.: -0.019 min
Response: 10170
Conc: 1.58 ng/ml



#43 AR-1268-3


R.T.: 8.570 min
Delta R.T.: -0.019 min
Response: 1823461
Conc: 258.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.592 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 8.968 min
Delta R.T.: -0.018 min
Response: 1108434
Conc: 398.71 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 7.888 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.356 min
Delta R.T.: 0.006 min
Response: 36323
Conc: 1.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK



#45 AR-1268-5

R.T.: 8.136 min
Delta R.T.: -0.021 min
Response: 7681
Conc: 0.53 ng/ml