

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0082319\  
 Data File : P0059772.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 23 Aug 2019 22:41  
 Operator : SM/AJ  
 Sample : K4461-12  
 Misc :  
 ALS Vial : 48 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 SB-06-COMP

Manual Integrations  
 APPROVED

Ankita  
 8/26/2019 3:12:06 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 23 23:24:01 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0081719.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Sat Aug 17 03:09:44 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1 | Resp#2 | ng/ml   | ng/ml  |
|-----------------------------|-------|-------|--------|--------|---------|--------|
| -----                       |       |       |        |        |         |        |
| System Monitoring Compounds |       |       |        |        |         |        |
| 1) SA Tetrachlo...          | 4.122 | 3.497 | 975314 | 903418 | 27.856m | 25.098 |
| 2) SA Decachlor...          | 9.596 | 8.359 | 987835 | 791265 | 18.568  | 19.624 |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO082319\  
Data File : PO059772.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 23 Aug 2019 22:41  
Operator : SM/AJ  
Sample : K4461-12  
Misc :  
ALS Vial : 48 Sample Multiplier: 1



Instrument :  
ECD\_O  
ClientSampleId :  
SB-06-COMP


Manual Integrations  
APPROVED

Ankita  
8/26/2019 3:12:06 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Aug 23 23:24:01 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO081719.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat Aug 17 03:09:44 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.122 min  
Delta R.T.: -0.027 min  
Response: 975314  
Conc: 27.86 ng/ml m

Instrument :  
ECD\_O  
ClientSampleId :  
SB-06-COMP


Manual Integrations  
APPROVED

Ankita  
8/26/2019 3:12:06 PM




#1 Tetrachloro-m-xylene

R.T.: 3.497 min  
Delta R.T.: -0.013 min  
Response: 903418  
Conc: 25.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.596 min  
Delta R.T.: -0.007 min  
Response: 987835  
Conc: 18.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.359 min  
Delta R.T.: -0.025 min  
Response: 791265  
Conc: 19.62 ng/ml