

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0082418\
 Data File : P0048472.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Aug 2018 19:05
 Operator : SM/SJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 01:39:09 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 24 01:39:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.388	3.630	9521290	13216593	50.000	50.000
2) SA Decachlor...	10.072	8.606	8182264	7869736	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.095	4.300	2282804	2929271	500.000	500.000
4) L1 AR-1016-2	5.288	4.710	2273989	4258952	500.000	500.000
5) L1 AR-1016-3	5.556	4.729	3268259	6237926	500.000	500.000
6) L1 AR-1016-4	5.639	4.784	2944912	4192332	500.000	500.000
7) L1 AR-1016-5	6.031	5.157	2421400	3443486	500.000	500.000
31) L7 AR-1260-1	7.153	6.184	4632557	6456139	500.000	500.000
32) L7 AR-1260-2	7.408	6.371	5543795	7980937	500.000	500.000
33) L7 AR-1260-3	7.692	6.525	6394758	7084917	500.000	500.000
34) L7 AR-1260-4	7.992	6.994	4545507	5360107	500.000	500.000
35) L7 AR-1260-5	8.306	7.234	8792069	12472054	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082418\
Data File : P0048472.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Aug 2018 19:05
Operator : SM/SJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 01:39:09 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082418.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 24 01:39:11 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.388 min
Delta R.T.: 0.000 min
Response: 9521290
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.630 min
Delta R.T.: 0.000 min
Response: 13216593
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.072 min
Delta R.T.: 0.000 min
Response: 8182264
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.606 min
Delta R.T.: 0.000 min
Response: 7869736
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.095 min
Delta R.T.: 0.000 min
Response: 2282804
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.300 min
Delta R.T.: 0.000 min
Response: 2929271
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.288 min
Delta R.T.: 0.000 min
Response: 2273989
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.710 min
Delta R.T.: 0.000 min
Response: 4258952
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.556 min
Delta R.T.: 0.000 min
Response: 3268259
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.729 min
Delta R.T.: 0.000 min
Response: 6237926
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.639 min
Delta R.T.: 0.000 min
Response: 2944912
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 4192332
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.031 min
Delta R.T.: 0.000 min
Response: 2421400
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.157 min
Delta R.T.: 0.000 min
Response: 3443486
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.153 min
Delta R.T.: 0.000 min
Response: 4632557
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.184 min
Delta R.T.: 0.000 min
Response: 6456139
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.408 min
Delta R.T.: 0.000 min
Response: 5543795
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.371 min
Delta R.T.: 0.000 min
Response: 7980937
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.692 min
Delta R.T.: 0.000 min
Response: 6394758
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.525 min
Delta R.T.: 0.000 min
Response: 7084917
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 4545507
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.994 min
Delta R.T.: 0.000 min
Response: 5360107
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.306 min
Delta R.T.: 0.000 min
Response: 8792069
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.234 min
Delta R.T.: 0.000 min
Response: 12472054
Conc: 500.00 ng/ml