

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0082420\  
 Data File : P0070810.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 24 Aug 2020 12:16  
 Operator : DD\AJ  
 Sample : PB131153BL  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 PB131153BL

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 24 13:37:06 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0082020.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Fri Aug 21 04:17:00 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1  | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-------|-------|---------|---------|--------|--------|
| -----                       |       |       |         |         |        |        |
| System Monitoring Compounds |       |       |         |         |        |        |
| 1) SA Tetrachlo...          | 4.370 | 3.555 | 1459330 | 847233  | 20.949 | 21.046 |
| 2) SA Decachlor...          | 9.998 | 8.740 | 1818426 | 1183429 | 19.423 | 21.016 |

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO082420\  
Data File : PO070810.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 24 Aug 2020 12:16  
Operator : DD\AJ  
Sample : PB131153BL  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
PB131153BL

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Aug 24 13:37:06 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO082020.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Aug 21 04:17:00 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.370 min  
Delta R.T.: 0.000 min  
Response: 1459330  
Conc: 20.95 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
PB131153BL




#1 Tetrachloro-m-xylene

R.T.: 3.555 min  
Delta R.T.: 0.002 min  
Response: 847233  
Conc: 21.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.998 min  
Delta R.T.: 0.000 min  
Response: 1818426  
Conc: 19.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.740 min  
Delta R.T.: -0.002 min  
Response: 1183429  
Conc: 21.02 ng/ml