

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082619\
 Data File : P0059846.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Aug 2019 10:16
 Operator : SM/AJ
 Sample : K4490-03RE
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 14:33:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.119	3.495	846329	846441	24.172	23.515
2)	SA Decachlor...	9.585	8.413	1334483	2810	25.084	0.070 #
Target Compounds							
3)	L1 AR-1016-1	5.271	4.607	5008	14591	3.123	10.793 #
4)	L1 AR-1016-2	5.290	4.607	3300	14591	1.379	7.284 #
5)	L1 AR-1016-3	5.390	4.774	29114	5798	19.546	5.360 #
6)	L1 AR-1016-4	5.474	4.774	-813	5798	N.D.	6.421 #
7)	L1 AR-1016-5	5.788	4.982	532	6612	0.414	5.740 #
8)	L2 AR-1221-1	4.379	0.000	5762	0	13.606	N.D. #
9)	L2 AR-1221-2	4.417	3.787	17449	25575	53.723	83.955 #
10)	L2 AR-1221-3	4.501	3.854	22584	17764	21.470	18.156
11)	L3 AR-1232-1	4.501	3.854	22584	17764	18.119	15.581
12)	L3 AR-1232-2	5.045	4.607	1423	14591	2.014	11.457 #
13)	L3 AR-1232-3	5.290	4.774	3300	5798	2.155	8.516 #
14)	L3 AR-1232-4	5.474	4.846	-813	6212	N.D.	10.183 #
15)	L3 AR-1232-5	5.577	4.982	5797	6612	9.643	9.713
16)	L4 AR-1242-1	5.290	4.607	3300	14591	2.600	14.184 #
17)	L4 AR-1242-2	5.290	4.607	3300	14591	1.778	9.580 #
18)	L4 AR-1242-3	5.390	4.774	29114	5798	24.908	6.940 #
19)	L4 AR-1242-4	5.474	4.846	-813	6212	N.D.	7.445 #
20)	L4 AR-1242-5	6.206	5.365	2136	10169	1.700	9.091 #
21)	L5 AR-1248-1	5.271	4.607	5008	14591	4.996	17.621 #
22)	L5 AR-1248-2	5.534	4.774	13676	5798	9.426	5.028 #
23)	L5 AR-1248-3	5.788	4.814	532	6335	0.320	5.334 #
24)	L5 AR-1248-4	6.134	4.982	31982	6612	15.646	4.550 #
25)	L5 AR-1248-5	6.170	5.365	24189	10169	12.198	7.105 #
26)	L6 AR-1254-1	6.106	5.325	28066	31376	13.452	13.229
27)	L6 AR-1254-2	6.323	5.470	54798	21516	16.425	10.254 #
28)	L6 AR-1254-3	6.714	5.866	16623	38260	4.685	11.379 #
29)	L6 AR-1254-4	6.970	6.091	29043	14997	9.860	7.274 #
30)	L6 AR-1254-5	7.385	6.499	54826	71812	17.729	23.128 #
31)	L7 AR-1260-1	6.844	5.990	27504	22660	10.203	9.943
32)	L7 AR-1260-2	7.099	6.178	60290	27637	16.248	9.433 #
33)	L7 AR-1260-3	7.452	6.326	31957	24698	9.754	9.323
34)	L7 AR-1260-4	7.681	6.786	49889	7861	16.244	3.475 #
35)	L7 AR-1260-5	7.987	7.032	95778	36487	13.526	6.630 #
36)	L8 AR-1262-1	7.514	6.532	19312	19582	3.979	5.253 #
37)	L8 AR-1262-2	8.029	7.077	48639	2121	6.304	0.361 #
38)	L8 AR-1262-3	8.286	7.353	65006	63898	12.741	25.147 #
39)	L8 AR-1262-4	8.412	7.353	55713	63898	14.453	14.460
40)	L8 AR-1262-5	8.997	0.000	52278	0	20.369	N.D. #
41)	L9 AR-1268-1	8.286	7.353	65006	63898	6.908	8.973 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082619\
 Data File : P0059846.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Aug 2019 10:16
 Operator : SM/AJ
 Sample : K4490-03RE
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 14:33:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.412	7.353	55713	63898	7.068	9.910 #
43) L9	AR-1268-3	8.595	7.592	45443	867	6.447	0.158 #
44) L9	AR-1268-4	8.997	0.000	52278	0	18.805	N.D. #
45) L9	AR-1268-5	9.342	8.217	14105	1678	0.765	0.115 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082619\
Data File : P0059846.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Aug 2019 10:16
Operator : SM/AJ
Sample : K4490-03RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 14:33:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.119 min
Delta R.T.: -0.030 min
Response: 846329
Conc: 24.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.495 min
Delta R.T.: -0.015 min
Response: 846441
Conc: 23.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.585 min
Delta R.T.: -0.018 min
Response: 1334483
Conc: 25.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.413 min
Delta R.T.: 0.029 min
Response: 2810
Conc: 0.07 ng/ml



#3 AR-1016-1


R.T.: 5.271 min
Delta R.T.: -0.009 min
Response: 5008
Conc: 3.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.607 min
Delta R.T.: 0.057 min
Response: 14591
Conc: 10.79 ng/ml




#4 AR-1016-2

R.T.: 5.290 min
Delta R.T.: -0.010 min
Response: 3300
Conc: 1.38 ng/ml



#4 AR-1016-2


R.T.: 4.607 min
Delta R.T.: 0.037 min
Response: 14591
Conc: 7.28 ng/ml



#5 AR-1016-3


R.T.: 5.390 min
Delta R.T.: 0.030 min
Response: 29114
Conc: 19.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.774 min
Delta R.T.: 0.034 min
Response: 5798
Conc: 5.36 ng/ml




#6 AR-1016-4

R.T.: 5.474 min
Delta R.T.: 0.014 min
Response: -813
Conc: N.D.



#6 AR-1016-4


R.T.: 4.774 min
Delta R.T.: -0.006 min
Response: 5798
Conc: 6.42 ng/ml



#7 AR-1016-5


R.T.: 5.788 min
Delta R.T.: 0.048 min
Response: 532
Conc: 0.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 4.982 min
Delta R.T.: -0.008 min
Response: 6612
Conc: 5.74 ng/ml




#8 AR-1221-1

R.T.: 4.379 min
Delta R.T.: 0.019 min
Response: 5762
Conc: 13.61 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.717 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.417 min
Delta R.T.: -0.029 min
Response: 17449
Conc: 53.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.787 min
Delta R.T.: -0.014 min
Response: 25575
Conc: 83.95 ng/ml




#10 AR-1221-3

R.T.: 4.501 min
Delta R.T.: -0.021 min
Response: 22584
Conc: 21.47 ng/ml



#10 AR-1221-3


R.T.: 3.854 min
Delta R.T.: -0.020 min
Response: 17764
Conc: 18.16 ng/ml



#11 AR-1232-1


R.T.: 4.501 min
Delta R.T.: -0.021 min
Response: 22584
Conc: 18.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.854 min
Delta R.T.: -0.020 min
Response: 17764
Conc: 15.58 ng/ml




#12 AR-1232-2

R.T.: 5.045 min
Delta R.T.: 0.008 min
Response: 1423
Conc: 2.01 ng/ml



#12 AR-1232-2


R.T.: 4.607 min
Delta R.T.: 0.026 min
Response: 14591
Conc: 11.46 ng/ml



#13 AR-1232-3


R.T.: 5.290 min
Delta R.T.: -0.032 min
Response: 3300
Conc: 2.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.774 min
Delta R.T.: 0.021 min
Response: 5798
Conc: 8.52 ng/ml




#14 AR-1232-4

R.T.: 5.474 min
Delta R.T.: -0.008 min
Response: -813
Conc: N.D.



#14 AR-1232-4


R.T.: 4.846 min
Delta R.T.: 0.011 min
Response: 6212
Conc: 10.18 ng/ml



#15 AR-1232-5


R.T.: 5.577 min
Delta R.T.: 0.009 min
Response: 5797
Conc: 9.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 4.982 min
Delta R.T.: -0.022 min
Response: 6612
Conc: 9.71 ng/ml




#16 AR-1242-1

R.T.: 5.290 min
Delta R.T.: -0.012 min
Response: 3300
Conc: 2.60 ng/ml



#16 AR-1242-1


R.T.: 4.607 min
Delta R.T.: 0.043 min
Response: 14591
Conc: 14.18 ng/ml



#17 AR-1242-2


R.T.: 5.290 min
Delta R.T.: -0.033 min
Response: 3300
Conc: 1.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.607 min
Delta R.T.: 0.026 min
Response: 14591
Conc: 9.58 ng/ml




#18 AR-1242-3

R.T.: 5.390 min
Delta R.T.: 0.007 min
Response: 29114
Conc: 24.91 ng/ml



#18 AR-1242-3


R.T.: 4.774 min
Delta R.T.: 0.020 min
Response: 5798
Conc: 6.94 ng/ml



#19 AR-1242-4


R.T.: 5.474 min
Delta R.T.: -0.008 min
Response: -813
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.846 min
Delta R.T.: 0.012 min
Response: 6212
Conc: 7.45 ng/ml




#20 AR-1242-5

R.T.: 6.206 min
Delta R.T.: 0.000 min
Response: 2136
Conc: 1.70 ng/ml



#20 AR-1242-5


R.T.: 5.365 min
Delta R.T.: 0.017 min
Response: 10169
Conc: 9.09 ng/ml



#21 AR-1248-1


R.T.: 5.271 min
Delta R.T.: -0.009 min
Response: 5008
Conc: 5.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.607 min
Delta R.T.: 0.057 min
Response: 14591
Conc: 17.62 ng/ml




#22 AR-1248-2

R.T.: 5.534 min
Delta R.T.: -0.006 min
Response: 13676
Conc: 9.43 ng/ml



#22 AR-1248-2


R.T.: 4.774 min
Delta R.T.: -0.006 min
Response: 5798
Conc: 5.03 ng/ml



#23 AR-1248-3


R.T.: 5.788 min
Delta R.T.: 0.048 min
Response: 532
Conc: 0.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.814 min
Delta R.T.: -0.006 min
Response: 6335
Conc: 5.33 ng/ml




#24 AR-1248-4

R.T.: 6.134 min
Delta R.T.: -0.006 min
Response: 31982
Conc: 15.65 ng/ml



#24 AR-1248-4


R.T.: 4.982 min
Delta R.T.: -0.008 min
Response: 6612
Conc: 4.55 ng/ml



#25 AR-1248-5


R.T.: 6.170 min
Delta R.T.: -0.010 min
Response: 24189
Conc: 12.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.365 min
Delta R.T.: -0.005 min
Response: 10169
Conc: 7.10 ng/ml




#26 AR-1254-1

R.T.: 6.106 min
Delta R.T.: -0.004 min
Response: 28066
Conc: 13.45 ng/ml



#26 AR-1254-1


R.T.: 5.325 min
Delta R.T.: -0.005 min
Response: 31376
Conc: 13.23 ng/ml



#27 AR-1254-2


R.T.: 6.323 min
Delta R.T.: -0.007 min
Response: 54798
Conc: 16.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.470 min
Delta R.T.: -0.010 min
Response: 21516
Conc: 10.25 ng/ml




#28 AR-1254-3

R.T.: 6.714 min
Delta R.T.: 0.014 min
Response: 16623
Conc: 4.68 ng/ml



#28 AR-1254-3


R.T.: 5.866 min
Delta R.T.: -0.014 min
Response: 38260
Conc: 11.38 ng/ml



#29 AR-1254-4


R.T.: 6.970 min
Delta R.T.: -0.010 min
Response: 29043
Conc: 9.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.091 min
Delta R.T.: -0.009 min
Response: 14997
Conc: 7.27 ng/ml




#30 AR-1254-5

R.T.: 7.385 min
Delta R.T.: -0.005 min
Response: 54826
Conc: 17.73 ng/ml



#30 AR-1254-5


R.T.: 6.499 min
Delta R.T.: -0.011 min
Response: 71812
Conc: 23.13 ng/ml



#31 AR-1260-1


R.T.: 6.844 min
Delta R.T.: -0.006 min
Response: 27504
Conc: 10.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 5.990 min
Delta R.T.: -0.010 min
Response: 22660
Conc: 9.94 ng/ml




#32 AR-1260-2

R.T.: 7.099 min
Delta R.T.: -0.011 min
Response: 60290
Conc: 16.25 ng/ml



#32 AR-1260-2


R.T.: 6.178 min
Delta R.T.: -0.002 min
Response: 27637
Conc: 9.43 ng/ml



#33 AR-1260-3


R.T.: 7.452 min
Delta R.T.: -0.018 min
Response: 31957
Conc: 9.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.326 min
Delta R.T.: -0.004 min
Response: 24698
Conc: 9.32 ng/ml




#34 AR-1260-4

R.T.: 7.681 min
Delta R.T.: -0.009 min
Response: 49889
Conc: 16.24 ng/ml



#34 AR-1260-4


R.T.: 6.786 min
Delta R.T.: -0.014 min
Response: 7861
Conc: 3.47 ng/ml



#35 AR-1260-5


R.T.: 7.987 min
Delta R.T.: -0.013 min
Response: 95778
Conc: 13.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.032 min
Delta R.T.: -0.008 min
Response: 36487
Conc: 6.63 ng/ml




#36 AR-1262-1

R.T.: 7.514 min
Delta R.T.: 0.025 min
Response: 19312
Conc: 3.98 ng/ml



#36 AR-1262-1


R.T.: 6.532 min
Delta R.T.: -0.026 min
Response: 19582
Conc: 5.25 ng/ml



#37 AR-1262-2


R.T.: 8.029 min
Delta R.T.: 0.006 min
Response: 48639
Conc: 6.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.077 min
Delta R.T.: 0.019 min
Response: 2121
Conc: 0.36 ng/ml




#38 AR-1262-3

R.T.: 8.286 min
Delta R.T.: -0.017 min
Response: 65006
Conc: 12.74 ng/ml



#38 AR-1262-3


R.T.: 7.353 min
Delta R.T.: 0.021 min
Response: 63898
Conc: 25.15 ng/ml



#39 AR-1262-4


R.T.: 8.412 min
Delta R.T.: 0.025 min
Response: 55713
Conc: 14.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.353 min
Delta R.T.: -0.044 min
Response: 63898
Conc: 14.46 ng/ml




#40 AR-1262-5

R.T.: 8.997 min
Delta R.T.: 0.011 min
Response: 52278
Conc: 20.37 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 7.889 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.286 min
Delta R.T.: -0.016 min
Response: 65006
Conc: 6.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.353 min
Delta R.T.: 0.023 min
Response: 63898
Conc: 8.97 ng/ml




#42 AR-1268-2

R.T.: 8.412 min
Delta R.T.: 0.022 min
Response: 55713
Conc: 7.07 ng/ml



#42 AR-1268-2


R.T.: 7.353 min
Delta R.T.: -0.044 min
Response: 63898
Conc: 9.91 ng/ml



#43 AR-1268-3


R.T.: 8.595 min
Delta R.T.: 0.006 min
Response: 45443
Conc: 6.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.592 min
Delta R.T.: 0.000 min
Response: 867
Conc: 0.16 ng/ml




#44 AR-1268-4

R.T.: 8.997 min
Delta R.T.: 0.011 min
Response: 52278
Conc: 18.80 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 7.888 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.342 min
Delta R.T.: -0.008 min
Response: 14105
Conc: 0.76 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.217 min
Delta R.T.: 0.061 min
Response: 1678
Conc: 0.12 ng/ml