

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0082621\
 Data File : P0080832.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Aug 2021 13:31
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 04:11:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 27 12:15:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	82.068	81.083 E3	1.2	99	0.00
2 SA	Decachlorobiphenyl	62.702	112.093 E3	-78.8#	180#	0.00
41 L9	AR-1268-1	8.885	9.311 E3	-4.8	101	0.00
42 L9	AR-1268-2	8.238	8.532 E3	-3.6	101	0.00
43 L9	AR-1268-3	6.977	7.212 E3	-3.4	101	0.00
44 L9	AR-1268-4	3.034	3.195 E3	-5.3	101	0.00
45 L9	AR-1268-5	22.025	23.329 E3	-5.9	103	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	26.949	26.359 E3	2.2	98	0.00
2 SA	Decachlorobiphenyl	27.288	45.143 E3	-65.4#	168#	0.00
41 L9	AR-1268-1	4.009	4.040 E3	-0.8	100	0.00
42 L9	AR-1268-2	3.526	3.585 E3	-1.7	101	0.00
43 L9	AR-1268-3	3.049	3.074 E3	-0.8	101	0.00
44 L9	AR-1268-4	1.406	1.441 E3	-2.5	101	0.00
45 L9	AR-1268-5	9.043	9.270 E3	-2.5	102	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1	AR-1016-1	3.092	0.000 E3	100.0#	0#	-6.27#
4 L1	AR-1016-2	4.325	0.000 E3	100.0#	0#	-6.29#
5 L1	AR-1016-3	2.737	0.000 E3	100.0#	0#	-6.36#
6 L1	AR-1016-4	2.236	0.000 E3	100.0#	0#	-6.46#
7 L1	AR-1016-5	2.183	0.000 E3	100.0#	0#	-6.78#
8 L2	AR-1221-1	928.204	0.000	100.0#	0#	-5.19#
9 L2	AR-1221-2	707.509	0.000	100.0#	0#	-5.29#
10 L2	AR-1221-3	2.207	0.000 E3	100.0#	0#	-5.38#
11 L3	AR-1232-1	1.700	0.000 E3	100.0#	0#	-5.38#
12 L3	AR-1232-2	991.883	0.000	100.0#	0#	-5.97#
13 L3	AR-1232-3	1.804	0.000 E3	100.0#	0#	-6.29#
14 L3	AR-1232-4	926.036	0.000	100.0#	0#	-6.46#
15 L3	AR-1232-5	668.934	0.000	100.0#	0#	-6.56#
16 L4	AR-1242-1	2.306	0.000 E3	100.0#	0#	-6.27#
17 L4	AR-1242-2	3.259	0.000 E3	100.0#	0#	-6.29#
18 L4	AR-1242-3	2.082	0.000 E3	100.0#	0#	-6.36#
19 L4	AR-1242-4	1.643	0.000 E3	100.0#	0#	-6.46#
20 L4	AR-1242-5	1.780	0.000 E3	100.0#	0#	-7.25#
21 L5	AR-1248-1	1.837	0.000 E3	100.0#	0#	-6.26#
22 L5	AR-1248-2	2.434	0.000 E3	100.0#	0#	-6.56#
23 L5	AR-1248-3	2.847	0.000 E3	100.0#	0#	-6.78#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0082621\
 Data File : P0080832.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Aug 2021 13:31
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 04:11:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 27 12:15:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	3.128	0.000 E3	100.0#	0#	-7.21#
25 L5 AR-1248-5	2.981	0.000 E3	100.0#	0#	-7.25#
26 L6 AR-1254-1	3.503	0.000 E3	100.0#	0#	-7.18#
27 L6 AR-1254-2	5.270	0.000 E3	100.0#	0#	-7.41#
28 L6 AR-1254-3	5.412	0.000 E3	100.0#	0#	-7.79#
29 L6 AR-1254-4	4.023	0.000 E3	100.0#	0#	-8.09#
30 L6 AR-1254-5	4.455	0.000 E3	100.0#	0#	-8.52#
31 L7 AR-1260-1	4.011	0.000 E3	100.0#	0#	-7.96#
32 L7 AR-1260-2	4.755	0.000 E3	100.0#	0#	-8.23#
33 L7 AR-1260-3	2.898	0.000 E3	100.0#	0#	-8.59#
34 L7 AR-1260-4	3.521	0.000 E3	100.0#	0#	-8.82#
35 L7 AR-1260-5	6.816	0.000 E3	100.0#	0#	-9.15#
36 L8 AR-1262-1	4.577	0.000 E3	100.0#	0#	-8.59#
37 L8 AR-1262-2	7.892	0.000 E3	100.0#	0#	-9.15#
38 L8 AR-1262-3	3.354	0.000 E3	100.0#	0#	-9.47#
39 L8 AR-1262-4	2.026	0.000 E3	100.0#	0#	-9.56#
40 L8 AR-1262-5	2.776	0.000 E3	100.0#	0#	-10.24#

Signal #2

3 L1 AR-1016-1	1.076	0.000 E3	100.0#	0#	-5.33#
4 L1 AR-1016-2	1.451	0.000 E3	100.0#	0#	-5.35#
5 L1 AR-1016-3	803.854	0.000	100.0#	0#	-5.55#
6 L1 AR-1016-4	667.549	0.000	100.0#	0#	-5.60#
7 L1 AR-1016-5	808.958	0.000	100.0#	0#	-5.83#
8 L2 AR-1221-1	312.120	0.000	100.0#	0#	-4.32#
9 L2 AR-1221-2	236.595	0.000	100.0#	0#	-4.42#
10 L2 AR-1221-3	757.473	0.000	100.0#	0#	-4.51#
11 L3 AR-1232-1	595.706	0.000	100.0#	0#	-4.51#
12 L3 AR-1232-2	616.867	0.000	100.0#	0#	-5.35#
13 L3 AR-1232-3	333.670	0.000	100.0#	0#	-5.55#
14 L3 AR-1232-4	313.624	0.000	100.0#	0#	-5.64#
15 L3 AR-1232-5	336.966	0.000	100.0#	0#	-5.83#
16 L4 AR-1242-1	791.400	0.000	100.0#	0#	-5.33#
17 L4 AR-1242-2	1.053	0.000 E3	100.0#	0#	-5.35#
18 L4 AR-1242-3	591.698	0.000	100.0#	0#	-5.55#
19 L4 AR-1242-4	589.032	0.000	100.0#	0#	-5.64#
20 L4 AR-1242-5	763.957	0.000	100.0#	0#	-6.21#
21 L5 AR-1248-1	652.198	0.000	100.0#	0#	-5.33#
22 L5 AR-1248-2	908.534	0.000	100.0#	0#	-5.60#
23 L5 AR-1248-3	926.675	0.000	100.0#	0#	-5.64#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0082621\
 Data File : P0080832.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Aug 2021 13:31
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 04:11:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 27 12:15:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.085	0.000 E3	100.0#	0#	-5.83#
25 L5 AR-1248-5	1.002	0.000 E3	100.0#	0#	-6.25#
26 L6 AR-1254-1	1.659	0.000 E3	100.0#	0#	-6.21#
27 L6 AR-1254-2	1.515	0.000 E3	100.0#	0#	-6.37#
28 L6 AR-1254-3	2.347	0.000 E3	100.0#	0#	-6.80#
29 L6 AR-1254-4	1.517	0.000 E3	100.0#	0#	-7.04#
30 L6 AR-1254-5	2.224	0.000 E3	100.0#	0#	-7.47#
31 L7 AR-1260-1	1.614	0.000 E3	100.0#	0#	-6.93#
32 L7 AR-1260-2	1.967	0.000 E3	100.0#	0#	-7.13#
33 L7 AR-1260-3	1.926	0.000 E3	100.0#	0#	-7.29#
34 L7 AR-1260-4	1.338	0.000 E3	100.0#	0#	-7.77#
35 L7 AR-1260-5	3.052	0.000 E3	100.0#	0#	-8.02#
36 L8 AR-1262-1	1.037	0.000 E3	100.0#	0#	-7.47#
37 L8 AR-1262-2	3.395	0.000 E3	100.0#	0#	-8.02#
38 L8 AR-1262-3	1.465	0.000 E3	100.0#	0#	-8.31#
39 L8 AR-1262-4	2.561	0.000 E3	100.0#	0#	-8.38#
40 L8 AR-1262-5	1.258	0.000 E3	100.0#	0#	-8.88#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0