

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082621\
 Data File : P0080821.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Aug 2021 10:01
 Operator : DD\AJ
 Sample : AR1262ICC050
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 27 10:45:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 27 10:43:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.936	4.061	394377	132766	4.921	5.102
2) SA Decachlor...	10.996	9.422	336455	152355	5.401	5.860
Target Compounds						
36) L8 AR-1262-1	8.593	7.468	267464	56760	59.578	54.657
37) L8 AR-1262-2	9.152	8.023	431197	186427	54.955	55.168
38) L8 AR-1262-3	9.466	8.311	180403	89136	53.236	62.563
39) L8 AR-1262-4	9.558	8.376	98393	143914	47.259	56.577
40) L8 AR-1262-5	10.237	8.878	141494	67999	49.428	53.706



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082621\
Data File : PO080821.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Aug 2021 10:01
Operator : DD\AJ
Sample : AR1262ICC050
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1262ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 27 10:45:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082621.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 27 10:43:13 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.936 min
Delta R.T.: 0.002 min
Response: 394377
Conc: 4.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050




#1 Tetrachloro-m-xylene

R.T.: 4.061 min
Delta R.T.: 0.000 min
Response: 132766
Conc: 5.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.996 min
Delta R.T.: 0.000 min
Response: 336455
Conc: 5.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.422 min
Delta R.T.: 0.000 min
Response: 152355
Conc: 5.86 ng/ml



#36 AR-1262-1


R.T.: 8.593 min
Delta R.T.: 0.001 min
Response: 267464
Conc: 59.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050




#36 AR-1262-1

R.T.: 7.468 min
Delta R.T.: 0.000 min
Response: 56760
Conc: 54.66 ng/ml




#37 AR-1262-2

R.T.: 9.152 min
Delta R.T.: 0.000 min
Response: 431197
Conc: 54.95 ng/ml



#37 AR-1262-2


R.T.: 8.023 min
Delta R.T.: 0.000 min
Response: 186427
Conc: 55.17 ng/ml



#38 AR-1262-3


R.T.: 9.466 min
Delta R.T.: 0.000 min
Response: 180403
Conc: 53.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050




#38 AR-1262-3

R.T.: 8.311 min
Delta R.T.: 0.000 min
Response: 89136
Conc: 62.56 ng/ml




#39 AR-1262-4

R.T.: 9.558 min
Delta R.T.: -0.004 min
Response: 98393
Conc: 47.26 ng/ml



#39 AR-1262-4


R.T.: 8.376 min
Delta R.T.: 0.000 min
Response: 143914
Conc: 56.58 ng/ml



#40 AR-1262-5

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 141494
Conc: 49.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050



#40 AR-1262-5

R.T.: 8.878 min
Delta R.T.: 0.000 min
Response: 67999
Conc: 53.71 ng/ml