

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082621\
 Data File : P0080828.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Aug 2021 12:21
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO082621AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 04:09:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 27 12:15:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.935	4.061	3992400	1268770	48.648	47.080
2) SA Decachlor...	10.996	9.423	3120642	1307253	49.770	47.905
Target Compounds						
16) L4 AR-1242-1	6.265	5.333	1147103	396096	497.489	500.501
17) L4 AR-1242-2	6.289	5.353	1587267	529438	487.020	502.690
18) L4 AR-1242-3	6.355	5.548	1005737	292505	483.092	494.349
19) L4 AR-1242-4	6.463	5.643	821601	308624	499.931	523.952
20) L4 AR-1242-5	7.249	6.212	850631	389380	477.918	509.689



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082621\
Data File : PO080828.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Aug 2021 12:21
Operator : DD\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO082621AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 04:09:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082621.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 27 12:15:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.935 min
Delta R.T.: 0.000 min
Response: 3992400
Conc: 48.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO082621AR1242




#1 Tetrachloro-m-xylene

R.T.: 4.061 min
Delta R.T.: 0.000 min
Response: 1268770
Conc: 47.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.996 min
Delta R.T.: -0.002 min
Response: 3120642
Conc: 49.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.423 min
Delta R.T.: 0.001 min
Response: 1307253
Conc: 47.91 ng/ml



#16 AR-1242-1


R.T.: 6.265 min
Delta R.T.: 0.000 min
Response: 1147103
Conc: 497.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO082621AR1242




#16 AR-1242-1

R.T.: 5.333 min
Delta R.T.: 0.001 min
Response: 396096
Conc: 500.50 ng/ml




#17 AR-1242-2

R.T.: 6.289 min
Delta R.T.: 0.000 min
Response: 1587267
Conc: 487.02 ng/ml



#17 AR-1242-2


R.T.: 5.353 min
Delta R.T.: 0.001 min
Response: 529438
Conc: 502.69 ng/ml



#18 AR-1242-3


R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 1005737
Conc: 483.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO082621AR1242




#18 AR-1242-3

R.T.: 5.548 min
Delta R.T.: 0.001 min
Response: 292505
Conc: 494.35 ng/ml




#19 AR-1242-4

R.T.: 6.463 min
Delta R.T.: 0.000 min
Response: 821601
Conc: 499.93 ng/ml



#19 AR-1242-4


R.T.: 5.643 min
Delta R.T.: 0.000 min
Response: 308624
Conc: 523.95 ng/ml



#20 AR-1242-5

R.T.: 7.249 min
Delta R.T.: -0.002 min
Response: 850631
Conc: 477.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO082621AR1242



#20 AR-1242-5

R.T.: 6.212 min
Delta R.T.: 0.000 min
Response: 389380
Conc: 509.69 ng/ml