

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082719\
 Data File : P0059923.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Aug 2019 1:14
 Operator : SM/AJ
 Sample : K4551-20MSD
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S402-K1-1.0-1.5-082619MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 04:35:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 17 03:09:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.124	3.500	1027225	845595	29.339	23.492
2)	SA Decachlor...	9.595	8.358	853425	650368	16.042	16.130
Target Compounds							
3)	L1 AR-1016-1	5.274	4.550	473709	296924	295.353	219.628 #
4)	L1 AR-1016-2	5.295	4.566	706129	435575	295.143	217.460 #
5)	L1 AR-1016-3	5.356	4.738	431067	237719	289.399	219.732
6)	L1 AR-1016-4	5.454	4.779	343199	168031	283.694	186.083 #
7)	L1 AR-1016-5	5.741	4.987	343783	227068	267.715	197.120 #
8)	L2 AR-1221-1	0.000	3.710	0	111347	N.D.	277.407 #
9)	L2 AR-1221-2	4.497	3.790	210974	177345	649.561	582.178
10)	L2 AR-1221-3	4.497	3.862	210974	199709	200.571	204.108
11)	L3 AR-1232-1	4.497	3.862	210974	199709	169.262	175.164
12)	L3 AR-1232-2	5.011	4.566	321791	435575	455.539	342.010
13)	L3 AR-1232-3	5.295	4.738	706129	237719	461.110	349.119
14)	L3 AR-1232-4	5.454	4.819	343199	187867	438.941	307.949 #
15)	L3 AR-1232-5	5.584	4.987	311480	227068	518.157	333.562 #
16)	L4 AR-1242-1	5.295	4.566	706129	435575	556.218	423.439
17)	L4 AR-1242-2	5.295	4.566	706129	435575	380.529	285.982
18)	L4 AR-1242-3	5.356	4.738	431067	237719	368.799	284.522
19)	L4 AR-1242-4	5.454	4.819	343199	187867	354.590	225.156 #
20)	L4 AR-1242-5	0.000	5.328	0	198223	N.D.	177.206 #
21)	L5 AR-1248-1	5.274	4.550	473709	296924	472.502	358.593
22)	L5 AR-1248-2	5.540	4.779	294795	168031	203.180	145.718 #
23)	L5 AR-1248-3	5.741	4.819	343783	187867	206.843	158.170
24)	L5 AR-1248-4	6.111	4.987	213474	227068	104.434	156.266 #
25)	L5 AR-1248-5	6.111	5.328	213474	198223	107.648	138.493 #
26)	L6 AR-1254-1	6.111	5.328	213474	198223	102.316	83.575
27)	L6 AR-1254-2	6.325	5.474	341207	216209	102.273	103.040
28)	L6 AR-1254-3	6.693	5.880	158856	347720	44.768	103.411 #
29)	L6 AR-1254-4	6.973	6.123	99028	117575	33.620	57.025 #
30)	L6 AR-1254-5	7.390	6.503	949938	688155	307.185	221.630 #
31)	L7 AR-1260-1	6.850	5.995	676166	458110	250.820	201.022
32)	L7 AR-1260-2	7.106	6.182	899420	629254	242.400	214.772
33)	L7 AR-1260-3	7.460	6.330	573216	571789	174.957	215.840
34)	L7 AR-1260-4	7.687	6.793	575552	396361	187.395	175.181
35)	L7 AR-1260-5	7.995	7.036	1156601	971560	163.341	176.541
36)	L8 AR-1262-1	7.518	6.537	290705	505224	59.903	135.538 #
37)	L8 AR-1262-2	7.995	7.036	1156601	971560	149.899	165.258
38)	L8 AR-1262-3	8.296	7.309	482683	201660	94.601	79.363
39)	L8 AR-1262-4	8.341	7.376	421841	676900	109.435	153.181 #
40)	L8 AR-1262-5	8.949	7.867	267559	184138	104.248	101.984
41)	L9 AR-1268-1	8.296	7.309	482683	201660	51.291	28.319 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082719\
Data File : P0059923.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Aug 2019 1:14
Operator : SM/AJ
Sample : K4551-20MSD
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 04:35:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.341	7.376	421841	676900	53.515	104.978 #
44)	L9 AR-1268-4	8.949	7.867	267559	184138	96.243	90.206



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082719\
Data File : PO059923.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Aug 2019 1:14
Operator : SM/AJ
Sample : K4551-20MSD
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 04:35:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 17 03:09:44 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.124 min
Delta R.T.: -0.025 min
Response: 1027225
Conc: 29.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#1 Tetrachloro-m-xylene

R.T.: 3.500 min
Delta R.T.: -0.010 min
Response: 845595
Conc: 23.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.595 min
Delta R.T.: -0.008 min
Response: 853425
Conc: 16.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.358 min
Delta R.T.: -0.026 min
Response: 650368
Conc: 16.13 ng/ml



#3 AR-1016-1


R.T.: 5.274 min
Delta R.T.: -0.006 min
Response: 473709
Conc: 295.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#3 AR-1016-1

R.T.: 4.550 min
Delta R.T.: 0.000 min
Response: 296924
Conc: 219.63 ng/ml




#4 AR-1016-2

R.T.: 5.295 min
Delta R.T.: -0.005 min
Response: 706129
Conc: 295.14 ng/ml



#4 AR-1016-2


R.T.: 4.566 min
Delta R.T.: -0.004 min
Response: 435575
Conc: 217.46 ng/ml



#5 AR-1016-3


R.T.: 5.356 min
Delta R.T.: -0.004 min
Response: 431067
Conc: 289.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#5 AR-1016-3

R.T.: 4.738 min
Delta R.T.: -0.002 min
Response: 237719
Conc: 219.73 ng/ml




#6 AR-1016-4

R.T.: 5.454 min
Delta R.T.: -0.006 min
Response: 343199
Conc: 283.69 ng/ml



#6 AR-1016-4


R.T.: 4.779 min
Delta R.T.: -0.001 min
Response: 168031
Conc: 186.08 ng/ml



#7 AR-1016-5


R.T.: 5.741 min
Delta R.T.: 0.001 min
Response: 343783
Conc: 267.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#7 AR-1016-5

R.T.: 4.987 min
Delta R.T.: -0.003 min
Response: 227068
Conc: 197.12 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.360 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.710 min
Delta R.T.: -0.007 min
Response: 111347
Conc: 277.41 ng/ml



#9 AR-1221-2


R.T.: 4.497 min
Delta R.T.: 0.051 min
Response: 210974
Conc: 649.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#9 AR-1221-2

R.T.: 3.790 min
Delta R.T.: -0.011 min
Response: 177345
Conc: 582.18 ng/ml




#10 AR-1221-3

R.T.: 4.497 min
Delta R.T.: -0.025 min
Response: 210974
Conc: 200.57 ng/ml



#10 AR-1221-3


R.T.: 3.862 min
Delta R.T.: -0.012 min
Response: 199709
Conc: 204.11 ng/ml



#11 AR-1232-1


R.T.: 4.497 min
Delta R.T.: -0.024 min
Response: 210974
Conc: 169.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#11 AR-1232-1

R.T.: 3.862 min
Delta R.T.: -0.012 min
Response: 199709
Conc: 175.16 ng/ml




#12 AR-1232-2

R.T.: 5.011 min
Delta R.T.: -0.026 min
Response: 321791
Conc: 455.54 ng/ml



#12 AR-1232-2


R.T.: 4.566 min
Delta R.T.: -0.015 min
Response: 435575
Conc: 342.01 ng/ml



#13 AR-1232-3


R.T.: 5.295 min
Delta R.T.: -0.027 min
Response: 706129
Conc: 461.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#13 AR-1232-3

R.T.: 4.738 min
Delta R.T.: -0.015 min
Response: 237719
Conc: 349.12 ng/ml




#14 AR-1232-4

R.T.: 5.454 min
Delta R.T.: -0.028 min
Response: 343199
Conc: 438.94 ng/ml



#14 AR-1232-4


R.T.: 4.819 min
Delta R.T.: -0.015 min
Response: 187867
Conc: 307.95 ng/ml



#15 AR-1232-5


R.T.: 5.584 min
Delta R.T.: 0.016 min
Response: 311480
Conc: 518.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#15 AR-1232-5

R.T.: 4.987 min
Delta R.T.: -0.016 min
Response: 227068
Conc: 333.56 ng/ml




#16 AR-1242-1

R.T.: 5.295 min
Delta R.T.: -0.006 min
Response: 706129
Conc: 556.22 ng/ml



#16 AR-1242-1


R.T.: 4.566 min
Delta R.T.: 0.002 min
Response: 435575
Conc: 423.44 ng/ml



#17 AR-1242-2


R.T.: 5.295 min
Delta R.T.: -0.028 min
Response: 706129
Conc: 380.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#17 AR-1242-2

R.T.: 4.566 min
Delta R.T.: -0.015 min
Response: 435575
Conc: 285.98 ng/ml




#18 AR-1242-3

R.T.: 5.356 min
Delta R.T.: -0.027 min
Response: 431067
Conc: 368.80 ng/ml



#18 AR-1242-3


R.T.: 4.738 min
Delta R.T.: -0.016 min
Response: 237719
Conc: 284.52 ng/ml



#19 AR-1242-4


R.T.: 5.454 min
Delta R.T.: -0.028 min
Response: 343199
Conc: 354.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#19 AR-1242-4

R.T.: 4.819 min
Delta R.T.: -0.015 min
Response: 187867
Conc: 225.16 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 6.206 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.328 min
Delta R.T.: -0.020 min
Response: 198223
Conc: 177.21 ng/ml



#21 AR-1248-1


R.T.: 5.274 min
Delta R.T.: -0.006 min
Response: 473709
Conc: 472.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#21 AR-1248-1

R.T.: 4.550 min
Delta R.T.: 0.000 min
Response: 296924
Conc: 358.59 ng/ml




#22 AR-1248-2

R.T.: 5.540 min
Delta R.T.: 0.000 min
Response: 294795
Conc: 203.18 ng/ml



#22 AR-1248-2


R.T.: 4.779 min
Delta R.T.: -0.001 min
Response: 168031
Conc: 145.72 ng/ml



#23 AR-1248-3


R.T.: 5.741 min
Delta R.T.: 0.001 min
Response: 343783
Conc: 206.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#23 AR-1248-3

R.T.: 4.819 min
Delta R.T.: -0.001 min
Response: 187867
Conc: 158.17 ng/ml




#24 AR-1248-4

R.T.: 6.111 min
Delta R.T.: -0.029 min
Response: 213474
Conc: 104.43 ng/ml



#24 AR-1248-4


R.T.: 4.987 min
Delta R.T.: -0.003 min
Response: 227068
Conc: 156.27 ng/ml



#25 AR-1248-5


R.T.: 6.111 min
Delta R.T.: -0.069 min
Response: 213474
Conc: 107.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#25 AR-1248-5

R.T.: 5.328 min
Delta R.T.: -0.042 min
Response: 198223
Conc: 138.49 ng/ml




#26 AR-1254-1

R.T.: 6.111 min
Delta R.T.: 0.000 min
Response: 213474
Conc: 102.32 ng/ml



#26 AR-1254-1


R.T.: 5.328 min
Delta R.T.: -0.002 min
Response: 198223
Conc: 83.57 ng/ml



#27 AR-1254-2


R.T.: 6.325 min
Delta R.T.: -0.005 min
Response: 341207
Conc: 102.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#27 AR-1254-2

R.T.: 5.474 min
Delta R.T.: -0.006 min
Response: 216209
Conc: 103.04 ng/ml




#28 AR-1254-3

R.T.: 6.693 min
Delta R.T.: -0.007 min
Response: 158856
Conc: 44.77 ng/ml



#28 AR-1254-3


R.T.: 5.880 min
Delta R.T.: 0.000 min
Response: 347720
Conc: 103.41 ng/ml



#29 AR-1254-4


R.T.: 6.973 min
Delta R.T.: -0.007 min
Response: 99028
Conc: 33.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#29 AR-1254-4

R.T.: 6.123 min
Delta R.T.: 0.023 min
Response: 117575
Conc: 57.02 ng/ml




#30 AR-1254-5

R.T.: 7.390 min
Delta R.T.: 0.000 min
Response: 949938
Conc: 307.19 ng/ml



#30 AR-1254-5


R.T.: 6.503 min
Delta R.T.: -0.007 min
Response: 688155
Conc: 221.63 ng/ml



#31 AR-1260-1


R.T.: 6.850 min
Delta R.T.: 0.000 min
Response: 676166
Conc: 250.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#31 AR-1260-1

R.T.: 5.995 min
Delta R.T.: -0.005 min
Response: 458110
Conc: 201.02 ng/ml




#32 AR-1260-2

R.T.: 7.106 min
Delta R.T.: -0.004 min
Response: 899420
Conc: 242.40 ng/ml



#32 AR-1260-2


R.T.: 6.182 min
Delta R.T.: 0.002 min
Response: 629254
Conc: 214.77 ng/ml



#33 AR-1260-3


R.T.: 7.460 min
Delta R.T.: -0.010 min
Response: 573216
Conc: 174.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#33 AR-1260-3

R.T.: 6.330 min
Delta R.T.: 0.000 min
Response: 571789
Conc: 215.84 ng/ml




#34 AR-1260-4

R.T.: 7.687 min
Delta R.T.: -0.003 min
Response: 575552
Conc: 187.40 ng/ml



#34 AR-1260-4


R.T.: 6.793 min
Delta R.T.: -0.007 min
Response: 396361
Conc: 175.18 ng/ml



#35 AR-1260-5


R.T.: 7.995 min
Delta R.T.: -0.005 min
Response: 1156601
Conc: 163.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#35 AR-1260-5

R.T.: 7.036 min
Delta R.T.: -0.004 min
Response: 971560
Conc: 176.54 ng/ml




#36 AR-1262-1

R.T.: 7.518 min
Delta R.T.: 0.029 min
Response: 290705
Conc: 59.90 ng/ml



#36 AR-1262-1


R.T.: 6.537 min
Delta R.T.: -0.021 min
Response: 505224
Conc: 135.54 ng/ml



#37 AR-1262-2


R.T.: 7.995 min
Delta R.T.: -0.029 min
Response: 1156601
Conc: 149.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#37 AR-1262-2

R.T.: 7.036 min
Delta R.T.: -0.021 min
Response: 971560
Conc: 165.26 ng/ml




#38 AR-1262-3

R.T.: 8.296 min
Delta R.T.: -0.008 min
Response: 482683
Conc: 94.60 ng/ml



#38 AR-1262-3


R.T.: 7.309 min
Delta R.T.: -0.023 min
Response: 201660
Conc: 79.36 ng/ml



#39 AR-1262-4


R.T.: 8.341 min
Delta R.T.: -0.045 min
Response: 421841
Conc: 109.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#39 AR-1262-4

R.T.: 7.376 min
Delta R.T.: -0.022 min
Response: 676900
Conc: 153.18 ng/ml




#40 AR-1262-5

R.T.: 8.949 min
Delta R.T.: -0.036 min
Response: 267559
Conc: 104.25 ng/ml



#40 AR-1262-5


R.T.: 7.867 min
Delta R.T.: -0.022 min
Response: 184138
Conc: 101.98 ng/ml



#41 AR-1268-1


R.T.: 8.296 min
Delta R.T.: -0.007 min
Response: 482683
Conc: 51.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD




#41 AR-1268-1

R.T.: 7.309 min
Delta R.T.: -0.021 min
Response: 201660
Conc: 28.32 ng/ml




#42 AR-1268-2

R.T.: 8.341 min
Delta R.T.: -0.049 min
Response: 421841
Conc: 53.51 ng/ml



#42 AR-1268-2


R.T.: 7.376 min
Delta R.T.: -0.022 min
Response: 676900
Conc: 104.98 ng/ml



#44 AR-1268-4

R.T.: 8.949 min
Delta R.T.: -0.036 min
Response: 267559
Conc: 96.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
S402-K1-1.0-1.5-082619MSD



#44 AR-1268-4

R.T.: 7.867 min
Delta R.T.: -0.021 min
Response: 184138
Conc: 90.21 ng/ml