

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082720\
 Data File : P0070910.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Aug 2020 12:38
 Operator : DD\AJ
 Sample : L3803-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC260025COK-END-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 27 13:00:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 21 04:17:00 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.377	3.566	1374403	889792	19.730	22.103
2) SA Decachlor...	10.004	8.748	1738921	938269	18.574	16.662

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082720\
Data File : P0070910.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Aug 2020 12:38
Operator : DD\AJ
Sample : L3803-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BC260025COK-END-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 27 13:00:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 21 04:17:00 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.377 min
Delta R.T.: 0.006 min
Response: 1374403
Conc: 19.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC260025COK-END-1




#1 Tetrachloro-m-xylene

R.T.: 3.566 min
Delta R.T.: 0.013 min
Response: 889792
Conc: 22.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.004 min
Delta R.T.: 0.005 min
Response: 1738921
Conc: 18.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.748 min
Delta R.T.: 0.006 min
Response: 938269
Conc: 16.66 ng/ml