

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082720\
 Data File : P0070943.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Aug 2020 23:40
 Operator : DD\AJ
 Sample : L3806-08
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-14-D

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 01:05:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 21 04:17:00 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.372	3.556	2114777	1213470	30.358	30.143
2) SA Decachlor...	9.999	8.740	1949450	1173562	20.823	20.840

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082720\
Data File : P0070943.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Aug 2020 23:40
Operator : DD\AJ
Sample : L3806-08
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-14-D

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 01:05:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 21 04:17:00 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.372 min
Delta R.T.: 0.001 min
Response: 2114777
Conc: 30.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-14-D




#1 Tetrachloro-m-xylene

R.T.: 3.556 min
Delta R.T.: 0.003 min
Response: 1213470
Conc: 30.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.999 min
Delta R.T.: 0.000 min
Response: 1949450
Conc: 20.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.002 min
Response: 1173562
Conc: 20.84 ng/ml