

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082823\
 Data File : P0097451.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Aug 2023 03:58
 Operator : YP/AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 04:48:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 29 04:47:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.446	3.628	172.7E6	33133553	50.000	50.000
2) SA Decachlor...	10.295	8.676	92534614	28907485	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.655	3.845	20123131	4093108	500.000	500.000
9) L2 AR-1221-2	4.743	3.930	15029670	2897864	500.000	500.000
10) L2 AR-1221-3	4.819	4.007	44540570	9194023	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082823\
Data File : PO097451.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Aug 2023 03:58
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 04:48:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 29 04:47:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.446 min
Delta R.T.: 0.000 min
Response: 172681597
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: 0.000 min
Response: 33133553
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.295 min
Delta R.T.: 0.000 min
Response: 92534614
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: 0.000 min
Response: 28907485
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 20123131
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.845 min
Delta R.T.: 0.000 min
Response: 4093108
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 15029670
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.930 min
Delta R.T.: 0.000 min
Response: 2897864
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.819 min
Delta R.T.: 0.000 min
Response: 44540570
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 9194023
Conc: 500.00 ng/ml