

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082824\
 Data File : P0105957.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2024 09:50
 Operator : YP/AJ
 Sample : PB163040BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB163040BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 01:36:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.453	3.719	180.7E6	55425380	19.803	20.515
2)	SA Decachlor...	10.195	8.753	111.9E6	45477602	20.622	20.832
Target Compounds							
3)	L1 AR-1016-1	5.616	4.804	353806	104286	1.119	1.091
4)	L1 AR-1016-2	5.650	4.840	441378	441773	1.004	3.481 #
5)	L1 AR-1016-3	5.704	5.006	105985	181843	0.394	2.599 #
6)	L1 AR-1016-4	5.796	5.048	97484	146058	0.443	2.716 #
7)	L1 AR-1016-5	6.086	5.260	59389	23585	0.279	0.343
8)	L2 AR-1221-1	4.643	3.939	636314	232411	5.780	7.617 #
9)	L2 AR-1221-2	4.734	4.023	170429	1032494	2.027	45.176 #
10)	L2 AR-1221-3	4.822	4.097	95306	125236	0.394	1.755 #
11)	L3 AR-1232-1	4.822	4.097	95306	125236	0.470	2.132 #
12)	L3 AR-1232-2	5.355	4.840	48158	441773	0.443	7.823 #
13)	L3 AR-1232-3	5.650	5.006	441378	181843	2.238	5.885 #
14)	L3 AR-1232-4	5.796	5.102	97484	43868	0.990	1.624 #
15)	L3 AR-1232-5	5.889	5.260	127320	23585	1.672	0.798 #
16)	L4 AR-1242-1	5.616	4.840	353806	441773	1.462	6.212 #
17)	L4 AR-1242-2	5.650	4.866	441378	124644	1.329	1.330
18)	L4 AR-1242-3	5.704	5.048	105985	146058	0.516	2.864 #
19)	L4 AR-1242-4	5.796	5.119	97484	257052	0.581	5.014 #
20)	L4 AR-1242-5	6.534	5.641	451185	454842	2.648	7.037 #
21)	L5 AR-1248-1	5.616	4.804	353806	104286	1.882	1.862
22)	L5 AR-1248-2	5.889	5.048	127320	146058	0.493	1.962 #
23)	L5 AR-1248-3	6.086	5.102	59389	43868	0.208	0.559 #
24)	L5 AR-1248-4	6.514	5.260	160724	23585	0.510	0.253 #
25)	L5 AR-1248-5	6.534	5.641	451185	454842	1.486	4.964 #
26)	L6 AR-1254-1	6.467	5.611	1910463	156585	6.189	1.161 #
27)	L6 AR-1254-2	6.705	5.759	1005970	1120093	2.183	9.229 #
28)	L6 AR-1254-3	7.054	6.152	1248888	160544	2.644	0.833 #
29)	L6 AR-1254-4	7.334	6.416	239777	742084	0.688	6.119 #
30)	L6 AR-1254-5	7.753	6.814	289794	564646	0.794	3.194 #
31)	L7 AR-1260-1	7.211	6.306	271031	60331	0.755	0.458 #
32)	L7 AR-1260-2	7.470	6.470	1168470	583109	2.782	3.553 #
33)	L7 AR-1260-3	7.826	6.644	2042436	4800446	7.637	28.444 #
34)	L7 AR-1260-4	8.069	7.115	3442835	112333	11.105	1.014 #
35)	L7 AR-1260-5	8.372	7.351	552620	180401	0.993	0.665 #
36)	L8 AR-1262-1	7.753	6.814	289794	564646	0.998	5.857 #
37)	L8 AR-1262-2	8.372	7.115	552620	112333	0.769	0.723
38)	L8 AR-1262-3	8.699	7.650	38551	49889	0.080	0.438 #
39)	L8 AR-1262-4	8.778	7.690	45538	113294	0.127	0.475 #
40)	L8 AR-1262-5	9.441	8.196	52112	35027	0.203	0.320 #
41)	L9 AR-1268-1	8.682	7.650	123432	49889	0.151	0.145

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082824\
 Data File : P0105957.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2024 09:50
 Operator : YP/AJ
 Sample : PB163040BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB163040BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 01:36:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation


Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm


	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.778	7.690	45538	113294	0.062	0.352 #
43) L9	AR-1268-3	9.012	7.905	158424	63396	0.258	0.236
44) L9	AR-1268-4	9.441	8.196	52112	35027	0.193	0.309 #
45) L9	AR-1268-5	9.855	8.488	754383	307466	0.374	0.359

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :
ECD_O
ClientSampleId :
PB163040BL

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.453 min
Delta R.T.: 0.000 min
Response: 180669785
Conc: 19.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#1 Tetrachloro-m-xylene

R.T.: 3.719 min
Delta R.T.: -0.002 min
Response: 55425380
Conc: 20.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.195 min
Delta R.T.: 0.000 min
Response: 111915641
Conc: 20.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.753 min
Delta R.T.: -0.004 min
Response: 45477602
Conc: 20.83 ng/ml



#3 AR-1016-1


R.T.: 5.616 min
Delta R.T.: -0.002 min
Response: 353806
Conc: 1.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#3 AR-1016-1

R.T.: 4.804 min
Delta R.T.: -0.008 min
Response: 104286
Conc: 1.09 ng/ml




#4 AR-1016-2

R.T.: 5.650 min
Delta R.T.: 0.010 min
Response: 441378
Conc: 1.00 ng/ml



#4 AR-1016-2


R.T.: 4.840 min
Delta R.T.: 0.008 min
Response: 441773
Conc: 3.48 ng/ml



#5 AR-1016-3


R.T.: 5.704 min
Delta R.T.: 0.001 min
Response: 105985
Conc: 0.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#5 AR-1016-3

R.T.: 5.006 min
Delta R.T.: -0.004 min
Response: 181843
Conc: 2.60 ng/ml




#6 AR-1016-4

R.T.: 5.796 min
Delta R.T.: -0.004 min
Response: 97484
Conc: 0.44 ng/ml



#6 AR-1016-4


R.T.: 5.048 min
Delta R.T.: 0.000 min
Response: 146058
Conc: 2.72 ng/ml



#7 AR-1016-5


R.T.: 6.086 min
Delta R.T.: -0.008 min
Response: 59389
Conc: 0.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#7 AR-1016-5

R.T.: 5.260 min
Delta R.T.: -0.005 min
Response: 23585
Conc: 0.34 ng/ml




#8 AR-1221-1

R.T.: 4.643 min
Delta R.T.: -0.016 min
Response: 636314
Conc: 5.78 ng/ml



#8 AR-1221-1


R.T.: 3.939 min
Delta R.T.: 0.003 min
Response: 232411
Conc: 7.62 ng/ml



#9 AR-1221-2


R.T.: 4.734 min
Delta R.T.: -0.011 min
Response: 170429
Conc: 2.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#9 AR-1221-2

R.T.: 4.023 min
Delta R.T.: 0.001 min
Response: 1032494
Conc: 45.18 ng/ml




#10 AR-1221-3

R.T.: 4.822 min
Delta R.T.: 0.000 min
Response: 95306
Conc: 0.39 ng/ml



#10 AR-1221-3


R.T.: 4.097 min
Delta R.T.: -0.002 min
Response: 125236
Conc: 1.75 ng/ml



#11 AR-1232-1


R.T.: 4.822 min
Delta R.T.: 0.000 min
Response: 95306
Conc: 0.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#11 AR-1232-1

R.T.: 4.097 min
Delta R.T.: -0.002 min
Response: 125236
Conc: 2.13 ng/ml




#12 AR-1232-2

R.T.: 5.355 min
Delta R.T.: 0.004 min
Response: 48158
Conc: 0.44 ng/ml



#12 AR-1232-2


R.T.: 4.840 min
Delta R.T.: 0.008 min
Response: 441773
Conc: 7.82 ng/ml



#13 AR-1232-3


R.T.: 5.650 min
Delta R.T.: 0.010 min
Response: 441378
Conc: 2.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#13 AR-1232-3

R.T.: 5.006 min
Delta R.T.: -0.003 min
Response: 181843
Conc: 5.89 ng/ml




#14 AR-1232-4

R.T.: 5.796 min
Delta R.T.: -0.004 min
Response: 97484
Conc: 0.99 ng/ml



#14 AR-1232-4


R.T.: 5.102 min
Delta R.T.: 0.011 min
Response: 43868
Conc: 1.62 ng/ml



#15 AR-1232-5


R.T.: 5.889 min
Delta R.T.: 0.000 min
Response: 127320
Conc: 1.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#15 AR-1232-5

R.T.: 5.260 min
Delta R.T.: -0.005 min
Response: 23585
Conc: 0.80 ng/ml




#16 AR-1242-1

R.T.: 5.616 min
Delta R.T.: -0.002 min
Response: 353806
Conc: 1.46 ng/ml



#16 AR-1242-1


R.T.: 4.840 min
Delta R.T.: 0.005 min
Response: 441773
Conc: 6.21 ng/ml



#17 AR-1242-2


R.T.: 5.650 min
Delta R.T.: 0.009 min
Response: 441378
Conc: 1.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#17 AR-1242-2

R.T.: 4.866 min
Delta R.T.: 0.011 min
Response: 124644
Conc: 1.33 ng/ml




#18 AR-1242-3

R.T.: 5.704 min
Delta R.T.: 0.000 min
Response: 105985
Conc: 0.52 ng/ml



#18 AR-1242-3


R.T.: 5.048 min
Delta R.T.: 0.017 min
Response: 146058
Conc: 2.86 ng/ml



#19 AR-1242-4


R.T.: 5.796 min
Delta R.T.: -0.005 min
Response: 97484
Conc: 0.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#19 AR-1242-4

R.T.: 5.119 min
Delta R.T.: 0.005 min
Response: 257052
Conc: 5.01 ng/ml




#20 AR-1242-5

R.T.: 6.534 min
Delta R.T.: -0.002 min
Response: 451185
Conc: 2.65 ng/ml



#20 AR-1242-5


R.T.: 5.641 min
Delta R.T.: 0.002 min
Response: 454842
Conc: 7.04 ng/ml



#21 AR-1248-1


R.T.: 5.616 min
Delta R.T.: -0.001 min
Response: 353806
Conc: 1.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#21 AR-1248-1

R.T.: 4.804 min
Delta R.T.: -0.008 min
Response: 104286
Conc: 1.86 ng/ml




#22 AR-1248-2

R.T.: 5.889 min
Delta R.T.: 0.000 min
Response: 127320
Conc: 0.49 ng/ml



#22 AR-1248-2


R.T.: 5.048 min
Delta R.T.: 0.000 min
Response: 146058
Conc: 1.96 ng/ml



#23 AR-1248-3


R.T.: 6.086 min
Delta R.T.: -0.007 min
Response: 59389
Conc: 0.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#23 AR-1248-3

R.T.: 5.102 min
Delta R.T.: 0.011 min
Response: 43868
Conc: 0.56 ng/ml




#24 AR-1248-4

R.T.: 6.514 min
Delta R.T.: 0.018 min
Response: 160724
Conc: 0.51 ng/ml



#24 AR-1248-4


R.T.: 5.260 min
Delta R.T.: -0.004 min
Response: 23585
Conc: 0.25 ng/ml



#25 AR-1248-5


R.T.: 6.534 min
Delta R.T.: 0.000 min
Response: 451185
Conc: 1.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#25 AR-1248-5

R.T.: 5.641 min
Delta R.T.: -0.017 min
Response: 454842
Conc: 4.96 ng/ml




#26 AR-1254-1

R.T.: 6.467 min
Delta R.T.: -0.003 min
Response: 1910463
Conc: 6.19 ng/ml



#26 AR-1254-1


R.T.: 5.611 min
Delta R.T.: -0.005 min
Response: 156585
Conc: 1.16 ng/ml



#27 AR-1254-2


R.T.: 6.705 min
Delta R.T.: 0.016 min
Response: 1005970
Conc: 2.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#27 AR-1254-2

R.T.: 5.759 min
Delta R.T.: -0.005 min
Response: 1120093
Conc: 9.23 ng/ml




#28 AR-1254-3

R.T.: 7.054 min
Delta R.T.: 0.001 min
Response: 1248888
Conc: 2.64 ng/ml



#28 AR-1254-3


R.T.: 6.152 min
Delta R.T.: -0.017 min
Response: 160544
Conc: 0.83 ng/ml



#29 AR-1254-4


R.T.: 7.334 min
Delta R.T.: -0.004 min
Response: 239777
Conc: 0.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#29 AR-1254-4

R.T.: 6.416 min
Delta R.T.: 0.019 min
Response: 742084
Conc: 6.12 ng/ml




#30 AR-1254-5

R.T.: 7.753 min
Delta R.T.: -0.003 min
Response: 289794
Conc: 0.79 ng/ml



#30 AR-1254-5


R.T.: 6.814 min
Delta R.T.: -0.001 min
Response: 564646
Conc: 3.19 ng/ml



#31 AR-1260-1


R.T.: 7.211 min
Delta R.T.: -0.007 min
Response: 271031
Conc: 0.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#31 AR-1260-1

R.T.: 6.306 min
Delta R.T.: 0.006 min
Response: 60331
Conc: 0.46 ng/ml




#32 AR-1260-2

R.T.: 7.470 min
Delta R.T.: -0.004 min
Response: 1168470
Conc: 2.78 ng/ml



#32 AR-1260-2


R.T.: 6.470 min
Delta R.T.: -0.016 min
Response: 583109
Conc: 3.55 ng/ml



#33 AR-1260-3


R.T.: 7.826 min
Delta R.T.: -0.006 min
Response: 2042436
Conc: 7.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#33 AR-1260-3

R.T.: 6.644 min
Delta R.T.: 0.001 min
Response: 4800446
Conc: 28.44 ng/ml




#34 AR-1260-4

R.T.: 8.069 min
Delta R.T.: 0.011 min
Response: 3442835
Conc: 11.11 ng/ml



#34 AR-1260-4


R.T.: 7.115 min
Delta R.T.: 0.000 min
Response: 112333
Conc: 1.01 ng/ml



#35 AR-1260-5


R.T.: 8.372 min
Delta R.T.: -0.005 min
Response: 552620
Conc: 0.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#35 AR-1260-5

R.T.: 7.351 min
Delta R.T.: -0.003 min
Response: 180401
Conc: 0.66 ng/ml




#36 AR-1262-1

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 289794
Conc: 1.00 ng/ml



#36 AR-1262-1


R.T.: 6.814 min
Delta R.T.: 0.000 min
Response: 564646
Conc: 5.86 ng/ml



#37 AR-1262-2


R.T.: 8.372 min
Delta R.T.: -0.005 min
Response: 552620
Conc: 0.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#37 AR-1262-2

R.T.: 7.115 min
Delta R.T.: 0.002 min
Response: 112333
Conc: 0.72 ng/ml




#38 AR-1262-3

R.T.: 8.699 min
Delta R.T.: 0.007 min
Response: 38551
Conc: 0.08 ng/ml



#38 AR-1262-3


R.T.: 7.650 min
Delta R.T.: 0.013 min
Response: 49889
Conc: 0.44 ng/ml



#39 AR-1262-4


R.T.: 8.778 min
Delta R.T.: -0.003 min
Response: 45538
Conc: 0.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#39 AR-1262-4

R.T.: 7.690 min
Delta R.T.: -0.011 min
Response: 113294
Conc: 0.48 ng/ml




#40 AR-1262-5

R.T.: 9.441 min
Delta R.T.: 0.006 min
Response: 52112
Conc: 0.20 ng/ml



#40 AR-1262-5


R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 35027
Conc: 0.32 ng/ml



#41 AR-1268-1


R.T.: 8.682 min
Delta R.T.: -0.006 min
Response: 123432
Conc: 0.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#41 AR-1268-1

R.T.: 7.650 min
Delta R.T.: 0.012 min
Response: 49889
Conc: 0.15 ng/ml




#42 AR-1268-2

R.T.: 8.778 min
Delta R.T.: -0.004 min
Response: 45538
Conc: 0.06 ng/ml



#42 AR-1268-2


R.T.: 7.690 min
Delta R.T.: -0.012 min
Response: 113294
Conc: 0.35 ng/ml



#43 AR-1268-3


R.T.: 9.012 min
Delta R.T.: 0.000 min
Response: 158424
Conc: 0.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL




#43 AR-1268-3

R.T.: 7.905 min
Delta R.T.: -0.005 min
Response: 63396
Conc: 0.24 ng/ml




#44 AR-1268-4

R.T.: 9.441 min
Delta R.T.: 0.006 min
Response: 52112
Conc: 0.19 ng/ml



#44 AR-1268-4


R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 35027
Conc: 0.31 ng/ml



#45 AR-1268-5

R.T.: 9.855 min
Delta R.T.: 0.002 min
Response: 754383
Conc: 0.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB163040BL



#45 AR-1268-5

R.T.: 8.488 min
Delta R.T.: -0.006 min
Response: 307466
Conc: 0.36 ng/ml