

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082919\
 Data File : P0060037.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Aug 2019 21:14
 Operator : SM/AJ
 Sample : K4439-03
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S-2-(1-3)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 01:10:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 28 13:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.121	3.499	1008236	835928	28.796	23.223
2)	SA Decachlor...	9.584	8.354	914969	694536	17.199	17.225
Target Compounds							
3)	L1 AR-1016-1	5.268	0.000	34803	0	21.700	N.D. #
4)	L1 AR-1016-2	5.314	0.000	7868	0	3.289	N.D. #
5)	L1 AR-1016-3	5.351	0.000	10435	0	7.006	N.D. #
6)	L1 AR-1016-4	5.446	4.843	2274	31115	1.880	34.457 #
7)	L1 AR-1016-5	5.729	4.932	15173	55083	11.816	47.818 #
8)	L2 AR-1221-1	4.363	0.000	20814	0	49.150	N.D. #
9)	L2 AR-1221-2	4.421	0.000	34987	0	107.720	N.D. #
10)	L2 AR-1221-3	4.501	3.906	29456	121466	28.004	124.141 #
11)	L3 AR-1232-1	4.501	3.906	29456	121466	23.632	106.537 #
12)	L3 AR-1232-2	5.008	0.000	43101	0	61.015	N.D. #
13)	L3 AR-1232-3	5.314	0.000	7868	0	5.138	N.D. #
14)	L3 AR-1232-4	5.538	4.843	18286	31115	23.387	51.003 #
15)	L3 AR-1232-5	5.587	4.932	13072	55083	21.746	80.917 #
16)	L4 AR-1242-1	5.314	0.000	7868	0	6.198	N.D. #
17)	L4 AR-1242-2	5.314	0.000	7868	0	4.240	N.D. #
18)	L4 AR-1242-3	5.351	0.000	10435	0	8.928	N.D. #
19)	L4 AR-1242-4	5.538	4.843	18286	31115	18.893	37.290 #
20)	L4 AR-1242-5	6.205	5.327	13956	80120	11.107	71.625 #
21)	L5 AR-1248-1	5.268	0.000	34803	0	34.715	N.D. #
22)	L5 AR-1248-2	5.538	4.843	18286	31115	12.603	26.983 #
23)	L5 AR-1248-3	5.691	4.843	41550	31115	24.999	26.196
24)	L5 AR-1248-4	6.136	4.932	26116	55083	12.776	37.908 #
25)	L5 AR-1248-5	6.173	5.327	19934	80120	10.052	55.978 #
26)	L6 AR-1254-1	6.109	5.327	74107	80120	35.519	33.780
27)	L6 AR-1254-2	6.326	5.473	102226	38721	30.641	18.454 #
28)	L6 AR-1254-3	6.689	5.869	76533	73019	21.568	21.716
29)	L6 AR-1254-4	6.969	0.000	48780	0	16.561	N.D. #
30)	L6 AR-1254-5	7.385	6.501	123797	77298	40.033	24.895 #
31)	L7 AR-1260-1	6.844	5.994	97480	30534	36.160	13.398 #
32)	L7 AR-1260-2	7.099	6.181	88951	55184	23.973	18.835
33)	L7 AR-1260-3	7.454	6.328	61514	58905	18.775	22.236
34)	L7 AR-1260-4	7.680	6.790	82680	54431	26.920	24.057
35)	L7 AR-1260-5	7.987	7.034	171907	111805	24.278	20.316
36)	L8 AR-1262-1	7.511	6.535	63625	48775	13.111	13.085
37)	L8 AR-1262-2	8.045	7.034	70613	111805	9.152	19.017 #
38)	L8 AR-1262-3	8.288	7.306	111803	32715	21.912	12.875 #
39)	L8 AR-1262-4	8.335	7.372	192843	89402	50.028	20.232 #
40)	L8 AR-1262-5	8.940	7.866	125247	57691	48.799	31.952 #
41)	L9 AR-1268-1	8.288	7.306	111803	32715	11.880	4.594 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082919\
 Data File : P0060037.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Aug 2019 21:14
 Operator : SM/AJ
 Sample : K4439-03
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S-2-(1-3)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 01:10:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 28 13:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.335	7.372	192843	89402	24.464	13.865 #
43) L9	AR-1268-3	8.598	7.569	26120	32161	3.706	5.878 #
44) L9	AR-1268-4	8.940	7.866	125247	57691	45.052	28.262 #
45) L9	AR-1268-5	9.345	8.131	9204	108277	0.499	7.452 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082919\
Data File : P0060037.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Aug 2019 21:14
Operator : SM/AJ
Sample : K4439-03
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 01:10:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 28 13:30:59 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.121 min
Delta R.T.: -0.003 min
Response: 1008236
Conc: 28.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#1 Tetrachloro-m-xylene

R.T.: 3.499 min
Delta R.T.: 0.000 min
Response: 835928
Conc: 23.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.584 min
Delta R.T.: -0.009 min
Response: 914969
Conc: 17.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.354 min
Delta R.T.: -0.005 min
Response: 694536
Conc: 17.23 ng/ml



#3 AR-1016-1


R.T.: 5.268 min
Delta R.T.: -0.006 min
Response: 34803
Conc: 21.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.550 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.314 min
Delta R.T.: 0.019 min
Response: 7868
Conc: 3.29 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.566 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.351 min
Delta R.T.: -0.005 min
Response: 10435
Conc: 7.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.738 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.446 min
Delta R.T.: -0.014 min
Response: 2274
Conc: 1.88 ng/ml



#6 AR-1016-4


R.T.: 4.843 min
Delta R.T.: 0.064 min
Response: 31115
Conc: 34.46 ng/ml



#7 AR-1016-5


R.T.: 5.729 min
Delta R.T.: -0.011 min
Response: 15173
Conc: 11.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#7 AR-1016-5

R.T.: 4.932 min
Delta R.T.: -0.055 min
Response: 55083
Conc: 47.82 ng/ml




#8 AR-1221-1

R.T.: 4.363 min
Delta R.T.: -0.014 min
Response: 20814
Conc: 49.15 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 3.717 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.421 min
Delta R.T.: 0.000 min
Response: 34987
Conc: 107.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.789 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 4.501 min
Delta R.T.: 0.004 min
Response: 29456
Conc: 28.00 ng/ml



#10 AR-1221-3


R.T.: 3.906 min
Delta R.T.: 0.044 min
Response: 121466
Conc: 124.14 ng/ml



#11 AR-1232-1


R.T.: 4.501 min
Delta R.T.: 0.004 min
Response: 29456
Conc: 23.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#11 AR-1232-1

R.T.: 3.906 min
Delta R.T.: 0.044 min
Response: 121466
Conc: 106.54 ng/ml




#12 AR-1232-2

R.T.: 5.008 min
Delta R.T.: -0.004 min
Response: 43101
Conc: 61.01 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T.: 4.566 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.314 min
Delta R.T.: 0.019 min
Response: 7868
Conc: 5.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.738 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.538 min
Delta R.T.: -0.002 min
Response: 18286
Conc: 23.39 ng/ml



#14 AR-1232-4


R.T.: 4.843 min
Delta R.T.: -0.005 min
Response: 31115
Conc: 51.00 ng/ml



#15 AR-1232-5


R.T.: 5.587 min
Delta R.T.: 0.004 min
Response: 13072
Conc: 21.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#15 AR-1232-5

R.T.: 4.932 min
Delta R.T.: -0.055 min
Response: 55083
Conc: 80.92 ng/ml




#16 AR-1242-1

R.T.: 5.314 min
Delta R.T.: 0.019 min
Response: 7868
Conc: 6.20 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T.: 4.566 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.314 min
Delta R.T.: 0.019 min
Response: 7868
Conc: 4.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.566 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.351 min
Delta R.T.: -0.005 min
Response: 10435
Conc: 8.93 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.738 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.538 min
Delta R.T.: -0.002 min
Response: 18286
Conc: 18.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#19 AR-1242-4

R.T.: 4.843 min
Delta R.T.: -0.005 min
Response: 31115
Conc: 37.29 ng/ml




#20 AR-1242-5

R.T.: 6.205 min
Delta R.T.: -0.015 min
Response: 13956
Conc: 11.11 ng/ml



#20 AR-1242-5


R.T.: 5.327 min
Delta R.T.: -0.002 min
Response: 80120
Conc: 71.63 ng/ml



#21 AR-1248-1


R.T.: 5.268 min
Delta R.T.: -0.006 min
Response: 34803
Conc: 34.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.550 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.538 min
Delta R.T.: -0.002 min
Response: 18286
Conc: 12.60 ng/ml



#22 AR-1248-2


R.T.: 4.843 min
Delta R.T.: 0.064 min
Response: 31115
Conc: 26.98 ng/ml



#23 AR-1248-3


R.T.: 5.691 min
Delta R.T.: -0.014 min
Response: 41550
Conc: 25.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#23 AR-1248-3

R.T.: 4.843 min
Delta R.T.: 0.025 min
Response: 31115
Conc: 26.20 ng/ml




#24 AR-1248-4

R.T.: 6.136 min
Delta R.T.: -0.004 min
Response: 26116
Conc: 12.78 ng/ml



#24 AR-1248-4


R.T.: 4.932 min
Delta R.T.: -0.055 min
Response: 55083
Conc: 37.91 ng/ml



#25 AR-1248-5


R.T.: 6.173 min
Delta R.T.: -0.004 min
Response: 19934
Conc: 10.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#25 AR-1248-5

R.T.: 5.327 min
Delta R.T.: -0.002 min
Response: 80120
Conc: 55.98 ng/ml




#26 AR-1254-1

R.T.: 6.109 min
Delta R.T.: -0.002 min
Response: 74107
Conc: 35.52 ng/ml



#26 AR-1254-1


R.T.: 5.327 min
Delta R.T.: -0.002 min
Response: 80120
Conc: 33.78 ng/ml



#27 AR-1254-2


R.T.: 6.326 min
Delta R.T.: 0.001 min
Response: 102226
Conc: 30.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#27 AR-1254-2

R.T.: 5.473 min
Delta R.T.: 0.000 min
Response: 38721
Conc: 18.45 ng/ml




#28 AR-1254-3

R.T.: 6.689 min
Delta R.T.: -0.003 min
Response: 76533
Conc: 21.57 ng/ml



#28 AR-1254-3


R.T.: 5.869 min
Delta R.T.: -0.012 min
Response: 73019
Conc: 21.72 ng/ml



#29 AR-1254-4


R.T.: 6.969 min
Delta R.T.: -0.005 min
Response: 48780
Conc: 16.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.097 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.385 min
Delta R.T.: -0.006 min
Response: 123797
Conc: 40.03 ng/ml



#30 AR-1254-5


R.T.: 6.501 min
Delta R.T.: -0.002 min
Response: 77298
Conc: 24.89 ng/ml



#31 AR-1260-1


R.T.: 6.844 min
Delta R.T.: -0.006 min
Response: 97480
Conc: 36.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#31 AR-1260-1

R.T.: 5.994 min
Delta R.T.: -0.001 min
Response: 30534
Conc: 13.40 ng/ml




#32 AR-1260-2

R.T.: 7.099 min
Delta R.T.: -0.006 min
Response: 88951
Conc: 23.97 ng/ml



#32 AR-1260-2


R.T.: 6.181 min
Delta R.T.: -0.003 min
Response: 55184
Conc: 18.83 ng/ml



#33 AR-1260-3


R.T.: 7.454 min
Delta R.T.: -0.006 min
Response: 61514
Conc: 18.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#33 AR-1260-3

R.T.: 6.328 min
Delta R.T.: -0.003 min
Response: 58905
Conc: 22.24 ng/ml




#34 AR-1260-4

R.T.: 7.680 min
Delta R.T.: -0.007 min
Response: 82680
Conc: 26.92 ng/ml



#34 AR-1260-4


R.T.: 6.790 min
Delta R.T.: -0.004 min
Response: 54431
Conc: 24.06 ng/ml



#35 AR-1260-5


R.T.: 7.987 min
Delta R.T.: -0.007 min
Response: 171907
Conc: 24.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#35 AR-1260-5

R.T.: 7.034 min
Delta R.T.: -0.003 min
Response: 111805
Conc: 20.32 ng/ml




#36 AR-1262-1

R.T.: 7.511 min
Delta R.T.: -0.006 min
Response: 63625
Conc: 13.11 ng/ml



#36 AR-1262-1


R.T.: 6.535 min
Delta R.T.: -0.003 min
Response: 48775
Conc: 13.08 ng/ml



#37 AR-1262-2


R.T.: 8.045 min
Delta R.T.: 0.007 min
Response: 70613
Conc: 9.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#37 AR-1262-2

R.T.: 7.034 min
Delta R.T.: -0.003 min
Response: 111805
Conc: 19.02 ng/ml




#38 AR-1262-3

R.T.: 8.288 min
Delta R.T.: -0.007 min
Response: 111803
Conc: 21.91 ng/ml



#38 AR-1262-3


R.T.: 7.306 min
Delta R.T.: -0.004 min
Response: 32715
Conc: 12.87 ng/ml



#39 AR-1262-4


R.T.: 8.335 min
Delta R.T.: -0.006 min
Response: 192843
Conc: 50.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#39 AR-1262-4

R.T.: 7.372 min
Delta R.T.: -0.004 min
Response: 89402
Conc: 20.23 ng/ml




#40 AR-1262-5

R.T.: 8.940 min
Delta R.T.: -0.010 min
Response: 125247
Conc: 48.80 ng/ml



#40 AR-1262-5


R.T.: 7.866 min
Delta R.T.: -0.003 min
Response: 57691
Conc: 31.95 ng/ml



#41 AR-1268-1


R.T.: 8.288 min
Delta R.T.: -0.007 min
Response: 111803
Conc: 11.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#41 AR-1268-1

R.T.: 7.306 min
Delta R.T.: -0.004 min
Response: 32715
Conc: 4.59 ng/ml




#42 AR-1268-2

R.T.: 8.335 min
Delta R.T.: -0.006 min
Response: 192843
Conc: 24.46 ng/ml



#42 AR-1268-2


R.T.: 7.372 min
Delta R.T.: -0.004 min
Response: 89402
Conc: 13.87 ng/ml



#43 AR-1268-3


R.T.: 8.598 min
Delta R.T.: -0.004 min
Response: 26120
Conc: 3.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)




#43 AR-1268-3

R.T.: 7.569 min
Delta R.T.: -0.024 min
Response: 32161
Conc: 5.88 ng/ml




#44 AR-1268-4

R.T.: 8.940 min
Delta R.T.: -0.010 min
Response: 125247
Conc: 45.05 ng/ml



#44 AR-1268-4


R.T.: 7.866 min
Delta R.T.: -0.003 min
Response: 57691
Conc: 28.26 ng/ml



#45 AR-1268-5

R.T.: 9.345 min
Delta R.T.: -0.002 min
Response: 9204
Conc: 0.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)



#45 AR-1268-5

R.T.: 8.131 min
Delta R.T.: -0.004 min
Response: 108277
Conc: 7.45 ng/ml