

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082919\
 Data File : P0060039.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Aug 2019 21:47
 Operator : SM/AJ
 Sample : K4439-03MSD
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S-2-(1-3)MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 01:10:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 28 13:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.121	3.499	933738	776509	26.669	21.572
2)	SA Decachlor...	9.581	8.353	865069	662626	16.261	16.434
Target Compounds							
3)	L1 AR-1016-1	5.270	4.548	459189	324800	286.300	240.247
4)	L1 AR-1016-2	5.290	4.565	648157	468170	270.913	233.733
5)	L1 AR-1016-3	5.351	4.736	378463	231629	254.083	214.103
6)	L1 AR-1016-4	5.414	4.777	17176	199943	14.198	221.423 #
7)	L1 AR-1016-5	5.736	4.985	27564	247668	21.465	215.003 #
8)	L2 AR-1221-1	4.382	3.703	5863	29151	13.846	72.625 #
9)	L2 AR-1221-2	4.418	3.789	74328	43994	228.847	144.420 #
10)	L2 AR-1221-3	4.494	3.861	242473	159467	230.518	162.980 #
11)	L3 AR-1232-1	4.494	3.861	242473	159467	194.534	139.868 #
12)	L3 AR-1232-2	5.007	4.565	363236	468170	514.211	367.603 #
13)	L3 AR-1232-3	5.290	4.736	648157	231629	423.254	340.175
14)	L3 AR-1232-4	5.535	4.846	259632	97275	332.061	159.452 #
15)	L3 AR-1232-5	5.579	4.985	287698	247668	478.596	363.824
16)	L4 AR-1242-1	5.290	4.565	648157	468170	510.554	455.126
17)	L4 AR-1242-2	5.290	4.565	648157	468170	349.288	307.382
18)	L4 AR-1242-3	5.351	4.736	378463	231629	323.794	277.233
19)	L4 AR-1242-4	5.535	4.846	259632	97275	268.249	116.582 #
20)	L4 AR-1242-5	6.200	5.326	14131	302883	11.245	270.769 #
21)	L5 AR-1248-1	5.270	4.548	459189	324800	458.020	392.259
22)	L5 AR-1248-2	5.535	4.777	259632	199943	178.944	173.393
23)	L5 AR-1248-3	5.689	4.817	43350	218772	26.082	184.191 #
24)	L5 AR-1248-4	6.135	4.985	81574	247668	39.907	170.443 #
25)	L5 AR-1248-5	6.172	5.326	66625	302883	33.597	211.616 #
26)	L6 AR-1254-1	6.105	5.326	350187	302883	167.841	127.702
27)	L6 AR-1254-2	6.319	5.472	404825	241536	121.341	115.110
28)	L6 AR-1254-3	6.685	5.878	226116	439582	63.722	130.730 #
29)	L6 AR-1254-4	6.966	6.093	156639	57415	53.179	27.847 #
30)	L6 AR-1254-5	7.382	6.501	1040169	754832	336.364	243.104 #
31)	L7 AR-1260-1	6.843	5.992	746295	525816	276.834	230.732
32)	L7 AR-1260-2	7.099	6.180	884104	658637	238.272	224.800
33)	L7 AR-1260-3	7.452	6.328	604840	617391	184.609	233.054 #
34)	L7 AR-1260-4	7.679	6.790	664126	433446	216.235	191.572
35)	L7 AR-1260-5	7.986	7.033	1329439	1082083	187.750	196.624
36)	L8 AR-1262-1	7.510	6.534	351591	528297	72.450	141.727 #
37)	L8 AR-1262-2	8.029	7.033	157914	1082083	20.466	184.058 #
38)	L8 AR-1262-3	8.286	7.306	594367	230783	116.490	90.825
39)	L8 AR-1262-4	8.332	7.372	563106	743288	146.082	168.205
40)	L8 AR-1262-5	8.939	7.863	393293	259342	153.237	143.635
41)	L9 AR-1268-1	8.286	7.306	594367	230783	63.159	32.409 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082919\
Data File : P0060039.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Aug 2019 21:47
Operator : SM/AJ
Sample : K4439-03MSD
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 01:10:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 28 13:30:59 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.332	7.372	563106	743288	71.436	115.274 #
43)	L9 AR-1268-3	8.599	7.566	94676	41968	13.431	7.670 #
44)	L9 AR-1268-4	8.939	7.863	393293	259342	141.470	127.048
45)	L9 AR-1268-5	9.339	8.129	10228	138048	0.555	9.501 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082919\
Data File : PO060039.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Aug 2019 21:47
Operator : SM/AJ
Sample : K4439-03MSD
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S-2-(1-3)MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 01:10:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 28 13:30:59 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.121 min
Delta R.T.: -0.003 min
Response: 933738
Conc: 26.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#1 Tetrachloro-m-xylene

R.T.: 3.499 min
Delta R.T.: 0.000 min
Response: 776509
Conc: 21.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.581 min
Delta R.T.: -0.012 min
Response: 865069
Conc: 16.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.353 min
Delta R.T.: -0.006 min
Response: 662626
Conc: 16.43 ng/ml



#3 AR-1016-1


R.T.: 5.270 min
Delta R.T.: -0.005 min
Response: 459189
Conc: 286.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#3 AR-1016-1

R.T.: 4.548 min
Delta R.T.: -0.002 min
Response: 324800
Conc: 240.25 ng/ml




#4 AR-1016-2

R.T.: 5.290 min
Delta R.T.: -0.005 min
Response: 648157
Conc: 270.91 ng/ml



#4 AR-1016-2


R.T.: 4.565 min
Delta R.T.: -0.002 min
Response: 468170
Conc: 233.73 ng/ml



#5 AR-1016-3


R.T.: 5.351 min
Delta R.T.: -0.005 min
Response: 378463
Conc: 254.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#5 AR-1016-3

R.T.: 4.736 min
Delta R.T.: -0.002 min
Response: 231629
Conc: 214.10 ng/ml




#6 AR-1016-4

R.T.: 5.414 min
Delta R.T.: -0.046 min
Response: 17176
Conc: 14.20 ng/ml



#6 AR-1016-4


R.T.: 4.777 min
Delta R.T.: -0.002 min
Response: 199943
Conc: 221.42 ng/ml



#7 AR-1016-5


R.T.: 5.736 min
Delta R.T.: -0.004 min
Response: 27564
Conc: 21.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#7 AR-1016-5

R.T.: 4.985 min
Delta R.T.: -0.002 min
Response: 247668
Conc: 215.00 ng/ml




#8 AR-1221-1

R.T.: 4.382 min
Delta R.T.: 0.005 min
Response: 5863
Conc: 13.85 ng/ml



#8 AR-1221-1


R.T.: 3.703 min
Delta R.T.: -0.014 min
Response: 29151
Conc: 72.62 ng/ml



#9 AR-1221-2


R.T.: 4.418 min
Delta R.T.: -0.004 min
Response: 74328
Conc: 228.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#9 AR-1221-2

R.T.: 3.789 min
Delta R.T.: 0.000 min
Response: 43994
Conc: 144.42 ng/ml




#10 AR-1221-3

R.T.: 4.494 min
Delta R.T.: -0.003 min
Response: 242473
Conc: 230.52 ng/ml



#10 AR-1221-3


R.T.: 3.861 min
Delta R.T.: -0.001 min
Response: 159467
Conc: 162.98 ng/ml



#11 AR-1232-1


R.T.: 4.494 min
Delta R.T.: -0.003 min
Response: 242473
Conc: 194.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#11 AR-1232-1

R.T.: 3.861 min
Delta R.T.: -0.001 min
Response: 159467
Conc: 139.87 ng/ml




#12 AR-1232-2

R.T.: 5.007 min
Delta R.T.: -0.005 min
Response: 363236
Conc: 514.21 ng/ml



#12 AR-1232-2


R.T.: 4.565 min
Delta R.T.: -0.002 min
Response: 468170
Conc: 367.60 ng/ml



#13 AR-1232-3


R.T.: 5.290 min
Delta R.T.: -0.005 min
Response: 648157
Conc: 423.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#13 AR-1232-3

R.T.: 4.736 min
Delta R.T.: -0.002 min
Response: 231629
Conc: 340.18 ng/ml




#14 AR-1232-4

R.T.: 5.535 min
Delta R.T.: -0.005 min
Response: 259632
Conc: 332.06 ng/ml



#14 AR-1232-4


R.T.: 4.846 min
Delta R.T.: -0.002 min
Response: 97275
Conc: 159.45 ng/ml



#15 AR-1232-5


R.T.: 5.579 min
Delta R.T.: -0.004 min
Response: 287698
Conc: 478.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#15 AR-1232-5

R.T.: 4.985 min
Delta R.T.: -0.002 min
Response: 247668
Conc: 363.82 ng/ml




#16 AR-1242-1

R.T.: 5.290 min
Delta R.T.: -0.005 min
Response: 648157
Conc: 510.55 ng/ml



#16 AR-1242-1


R.T.: 4.565 min
Delta R.T.: -0.002 min
Response: 468170
Conc: 455.13 ng/ml



#17 AR-1242-2


R.T.: 5.290 min
Delta R.T.: -0.005 min
Response: 648157
Conc: 349.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#17 AR-1242-2

R.T.: 4.565 min
Delta R.T.: -0.002 min
Response: 468170
Conc: 307.38 ng/ml




#18 AR-1242-3

R.T.: 5.351 min
Delta R.T.: -0.005 min
Response: 378463
Conc: 323.79 ng/ml



#18 AR-1242-3


R.T.: 4.736 min
Delta R.T.: -0.002 min
Response: 231629
Conc: 277.23 ng/ml



#19 AR-1242-4


R.T.: 5.535 min
Delta R.T.: -0.005 min
Response: 259632
Conc: 268.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#19 AR-1242-4

R.T.: 4.846 min
Delta R.T.: -0.002 min
Response: 97275
Conc: 116.58 ng/ml




#20 AR-1242-5

R.T.: 6.200 min
Delta R.T.: -0.019 min
Response: 14131
Conc: 11.25 ng/ml



#20 AR-1242-5


R.T.: 5.326 min
Delta R.T.: -0.003 min
Response: 302883
Conc: 270.77 ng/ml



#21 AR-1248-1


R.T.: 5.270 min
Delta R.T.: -0.005 min
Response: 459189
Conc: 458.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#21 AR-1248-1

R.T.: 4.548 min
Delta R.T.: -0.002 min
Response: 324800
Conc: 392.26 ng/ml




#22 AR-1248-2

R.T.: 5.535 min
Delta R.T.: -0.005 min
Response: 259632
Conc: 178.94 ng/ml



#22 AR-1248-2


R.T.: 4.777 min
Delta R.T.: -0.002 min
Response: 199943
Conc: 173.39 ng/ml



#23 AR-1248-3


R.T.: 5.689 min
Delta R.T.: -0.016 min
Response: 43350
Conc: 26.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#23 AR-1248-3

R.T.: 4.817 min
Delta R.T.: -0.001 min
Response: 218772
Conc: 184.19 ng/ml




#24 AR-1248-4

R.T.: 6.135 min
Delta R.T.: -0.005 min
Response: 81574
Conc: 39.91 ng/ml



#24 AR-1248-4


R.T.: 4.985 min
Delta R.T.: -0.002 min
Response: 247668
Conc: 170.44 ng/ml



#25 AR-1248-5


R.T.: 6.172 min
Delta R.T.: -0.005 min
Response: 66625
Conc: 33.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#25 AR-1248-5

R.T.: 5.326 min
Delta R.T.: -0.003 min
Response: 302883
Conc: 211.62 ng/ml




#26 AR-1254-1

R.T.: 6.105 min
Delta R.T.: -0.005 min
Response: 350187
Conc: 167.84 ng/ml



#26 AR-1254-1


R.T.: 5.326 min
Delta R.T.: -0.003 min
Response: 302883
Conc: 127.70 ng/ml



#27 AR-1254-2


R.T.: 6.319 min
Delta R.T.: -0.006 min
Response: 404825
Conc: 121.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#27 AR-1254-2

R.T.: 5.472 min
Delta R.T.: -0.002 min
Response: 241536
Conc: 115.11 ng/ml




#28 AR-1254-3

R.T.: 6.685 min
Delta R.T.: -0.007 min
Response: 226116
Conc: 63.72 ng/ml



#28 AR-1254-3


R.T.: 5.878 min
Delta R.T.: -0.003 min
Response: 439582
Conc: 130.73 ng/ml



#29 AR-1254-4


R.T.: 6.966 min
Delta R.T.: -0.008 min
Response: 156639
Conc: 53.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#29 AR-1254-4

R.T.: 6.093 min
Delta R.T.: -0.003 min
Response: 57415
Conc: 27.85 ng/ml




#30 AR-1254-5

R.T.: 7.382 min
Delta R.T.: -0.008 min
Response: 1040169
Conc: 336.36 ng/ml



#30 AR-1254-5


R.T.: 6.501 min
Delta R.T.: -0.003 min
Response: 754832
Conc: 243.10 ng/ml



#31 AR-1260-1


R.T.: 6.843 min
Delta R.T.: -0.007 min
Response: 746295
Conc: 276.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#31 AR-1260-1

R.T.: 5.992 min
Delta R.T.: -0.002 min
Response: 525816
Conc: 230.73 ng/ml




#32 AR-1260-2

R.T.: 7.099 min
Delta R.T.: -0.007 min
Response: 884104
Conc: 238.27 ng/ml



#32 AR-1260-2


R.T.: 6.180 min
Delta R.T.: -0.003 min
Response: 658637
Conc: 224.80 ng/ml



#33 AR-1260-3


R.T.: 7.452 min
Delta R.T.: -0.008 min
Response: 604840
Conc: 184.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#33 AR-1260-3

R.T.: 6.328 min
Delta R.T.: -0.003 min
Response: 617391
Conc: 233.05 ng/ml




#34 AR-1260-4

R.T.: 7.679 min
Delta R.T.: -0.008 min
Response: 664126
Conc: 216.23 ng/ml



#34 AR-1260-4


R.T.: 6.790 min
Delta R.T.: -0.004 min
Response: 433446
Conc: 191.57 ng/ml



#35 AR-1260-5


R.T.: 7.986 min
Delta R.T.: -0.008 min
Response: 1329439
Conc: 187.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#35 AR-1260-5

R.T.: 7.033 min
Delta R.T.: -0.004 min
Response: 1082083
Conc: 196.62 ng/ml




#36 AR-1262-1

R.T.: 7.510 min
Delta R.T.: -0.007 min
Response: 351591
Conc: 72.45 ng/ml



#36 AR-1262-1


R.T.: 6.534 min
Delta R.T.: -0.004 min
Response: 528297
Conc: 141.73 ng/ml



#37 AR-1262-2


R.T.: 8.029 min
Delta R.T.: -0.008 min
Response: 157914
Conc: 20.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#37 AR-1262-2

R.T.: 7.033 min
Delta R.T.: -0.004 min
Response: 1082083
Conc: 184.06 ng/ml




#38 AR-1262-3

R.T.: 8.286 min
Delta R.T.: -0.008 min
Response: 594367
Conc: 116.49 ng/ml



#38 AR-1262-3


R.T.: 7.306 min
Delta R.T.: -0.004 min
Response: 230783
Conc: 90.82 ng/ml



#39 AR-1262-4


R.T.: 8.332 min
Delta R.T.: -0.009 min
Response: 563106
Conc: 146.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#39 AR-1262-4

R.T.: 7.372 min
Delta R.T.: -0.004 min
Response: 743288
Conc: 168.20 ng/ml




#40 AR-1262-5

R.T.: 8.939 min
Delta R.T.: -0.011 min
Response: 393293
Conc: 153.24 ng/ml



#40 AR-1262-5


R.T.: 7.863 min
Delta R.T.: -0.005 min
Response: 259342
Conc: 143.63 ng/ml



#41 AR-1268-1


R.T.: 8.286 min
Delta R.T.: -0.008 min
Response: 594367
Conc: 63.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#41 AR-1268-1

R.T.: 7.306 min
Delta R.T.: -0.004 min
Response: 230783
Conc: 32.41 ng/ml




#42 AR-1268-2

R.T.: 8.332 min
Delta R.T.: -0.009 min
Response: 563106
Conc: 71.44 ng/ml



#42 AR-1268-2


R.T.: 7.372 min
Delta R.T.: -0.004 min
Response: 743288
Conc: 115.27 ng/ml



#43 AR-1268-3


R.T.: 8.599 min
Delta R.T.: -0.004 min
Response: 94676
Conc: 13.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD




#43 AR-1268-3

R.T.: 7.566 min
Delta R.T.: -0.026 min
Response: 41968
Conc: 7.67 ng/ml




#44 AR-1268-4

R.T.: 8.939 min
Delta R.T.: -0.011 min
Response: 393293
Conc: 141.47 ng/ml



#44 AR-1268-4


R.T.: 7.863 min
Delta R.T.: -0.005 min
Response: 259342
Conc: 127.05 ng/ml



#45 AR-1268-5

R.T.: 9.339 min
Delta R.T.: -0.008 min
Response: 10228
Conc: 0.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-2-(1-3)MSD



#45 AR-1268-5

R.T.: 8.129 min
Delta R.T.: -0.005 min
Response: 138048
Conc: 9.50 ng/ml