

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082919\
 Data File : P0060013.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Aug 2019 13:51
 Operator : SM/AJ
 Sample : K4544-19
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 T16-17-E1-E4-(0-2.25)

Manual Integrations
 APPROVED

Ankita
 8/30/2019 5:48:37 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 01:54:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 28 13:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.130	3.503	984208	791193	28.110	21.980
2)	SA Decachlor...	9.591	8.357	1235270	979203	23.219	24.285
Target Compounds							
31)	L7 AR-1260-1	6.850	5.995	320919	177091	119.043	77.709 #
32)	L7 AR-1260-2	7.100	6.187	409063	280274	110.245	95.661
33)	L7 AR-1260-3	7.459	6.330	219929	157002	67.127	59.265
34)	L7 AR-1260-4	7.683	6.793	307763	106751	100.205m	47.181 #
35)	L7 AR-1260-5	7.992	7.037	325347	298188	45.947m	54.183

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082919\
Data File : PO060013.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Aug 2019 13:51
Operator : SM/AJ
Sample : K4544-19
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
T16-17-E1-E4-(0-2.25)


Manual Integrations
APPROVED

Ankita
8/30/2019 5:48:37 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 01:54:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 28 13:30:59 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.130 min
Delta R.T.: 0.006 min
Response: 984208
Conc: 28.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
T16-17-E1-E4-(0-2.25)


Manual Integrations
APPROVED

Ankita
8/30/2019 5:48:37 PM




#1 Tetrachloro-m-xylene

R.T.: 3.503 min
Delta R.T.: 0.003 min
Response: 791193
Conc: 21.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.591 min
Delta R.T.: -0.002 min
Response: 1235270
Conc: 23.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.357 min
Delta R.T.: -0.002 min
Response: 979203
Conc: 24.29 ng/ml




#31 AR-1260-1

R.T.: 6.850 min
Delta R.T.: 0.000 min
Response: 320919
Conc: 119.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
T16-17-E1-E4-(0-2.25)


Manual Integrations
APPROVED

Ankita
8/30/2019 5:48:37 PM




#31 AR-1260-1

R.T.: 5.995 min
Delta R.T.: 0.000 min
Response: 177091
Conc: 77.71 ng/ml




#32 AR-1260-2

R.T.: 7.100 min
Delta R.T.: -0.005 min
Response: 409063
Conc: 110.25 ng/ml



#32 AR-1260-2

R.T.: 6.187 min
Delta R.T.: 0.004 min
Response: 280274
Conc: 95.66 ng/ml




#33 AR-1260-3

R.T.: 7.459 min
Delta R.T.: 0.000 min
Response: 219929
Conc: 67.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
T16-17-E1-E4-(0-2.25)


Manual Integrations
APPROVED

Ankita
8/30/2019 5:48:37 PM




#33 AR-1260-3

R.T.: 6.330 min
Delta R.T.: 0.000 min
Response: 157002
Conc: 59.27 ng/ml




#34 AR-1260-4

R.T.: 7.683 min
Delta R.T.: -0.003 min
Response: 307763
Conc: 100.21 ng/ml m



#34 AR-1260-4

R.T.: 6.793 min
Delta R.T.: -0.001 min
Response: 106751
Conc: 47.18 ng/ml




#35 AR-1260-5

R.T.: 7.992 min
Delta R.T.: -0.003 min
Response: 325347
Conc: 45.95 ng/ml m

Instrument :
ECD_O
ClientSampleId :
T16-17-E1-E4-(0-2.25)

Manual Integrations
APPROVED

Ankita
8/30/2019 5:48:37 PM



#35 AR-1260-5

R.T.: 7.037 min
Delta R.T.: 0.000 min
Response: 298188
Conc: 54.18 ng/ml