

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082919\
 Data File : P0060052.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Aug 2019 1:16
 Operator : SM/AJ
 Sample : K4562-01
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-02-15-17

Manual Integrations
 APPROVED

Ankita
 8/30/2019 5:50:03 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 02:22:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 28 13:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.121	3.499	930490	816085	26.576	22.672
2)	SA Decachlor...	9.581	8.352	753003	491487	14.154	12.189m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082919\
Data File : PO060052.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Aug 2019 1:16
Operator : SM/AJ
Sample : K4562-01
Misc :
ALS Vial : 50 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SB-02-15-17


Manual Integrations
APPROVED

Ankita
8/30/2019 5:50:03 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 02:22:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 28 13:30:59 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.121 min
Delta R.T.: -0.003 min
Response: 930490
Conc: 26.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-02-15-17


Manual Integrations
APPROVED

Ankita
8/30/2019 5:50:03 PM




#1 Tetrachloro-m-xylene

R.T.: 3.499 min
Delta R.T.: 0.000 min
Response: 816085
Conc: 22.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.581 min
Delta R.T.: -0.012 min
Response: 753003
Conc: 14.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.352 min
Delta R.T.: -0.007 min
Response: 491487
Conc: 12.19 ng/ml m