

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082922\
 Data File : P0089094.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Aug 2022 11:53
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 12:53:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 12:52:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.431	3.590	285.5E6	77421958	70.565	66.124
2) SA Decachlor...	10.259	8.574	161.3E6	68875120	69.282	65.564
Target Compounds						
16) L4 AR-1242-1	5.606	4.666	67503561	21636350	685.179	658.269
17) L4 AR-1242-2	5.629	4.685	94443589	29919131	693.211	661.561
18) L4 AR-1242-3	5.691	4.859	58223128	16192479	678.542	653.301
19) L4 AR-1242-4	5.791	4.943	47783542	16230600	688.382	651.167
20) L4 AR-1242-5	6.531	5.465	48877157	21199862	694.319	675.697



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO082922\
Data File : PO089094.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Aug 2022 11:53
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 12:53:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 12:52:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.431 min
Delta R.T.: 0.001 min
Response: 285460755
Conc: 70.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.590 min
Delta R.T.: 0.002 min
Response: 77421958
Conc: 66.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.259 min
Delta R.T.: -0.003 min
Response: 161297047
Conc: 69.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.574 min
Delta R.T.: -0.002 min
Response: 68875120
Conc: 65.56 ng/ml



#16 AR-1242-1


R.T.: 5.606 min
Delta R.T.: 0.000 min
Response: 67503561
Conc: 685.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.666 min
Delta R.T.: 0.001 min
Response: 21636350
Conc: 658.27 ng/ml




#17 AR-1242-2

R.T.: 5.629 min
Delta R.T.: 0.000 min
Response: 94443589
Conc: 693.21 ng/ml



#17 AR-1242-2


R.T.: 4.685 min
Delta R.T.: 0.001 min
Response: 29919131
Conc: 661.56 ng/ml



#18 AR-1242-3


R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 58223128
Conc: 678.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 4.859 min
Delta R.T.: 0.001 min
Response: 16192479
Conc: 653.30 ng/ml




#19 AR-1242-4

R.T.: 5.791 min
Delta R.T.: 0.000 min
Response: 47783542
Conc: 688.38 ng/ml



#19 AR-1242-4


R.T.: 4.943 min
Delta R.T.: 0.001 min
Response: 16230600
Conc: 651.17 ng/ml



#20 AR-1242-5

R.T.: 6.531 min
Delta R.T.: 0.000 min
Response: 48877157
Conc: 694.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.465 min
Delta R.T.: 0.002 min
Response: 21199862
Conc: 675.70 ng/ml