

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082922\
 Data File : P0089127.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Aug 2022 21:24
 Operator : YP/AJ
 Sample : N4401-02 10X
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 59B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 03:06:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.588	4888070	1455324	1.304	1.508
2) SA Decachlor...	10.260	8.572	3540612	1461440	1.538	1.653

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0082922\
Data File : P0089127.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Aug 2022 21:24
Operator : YP/AJ
Sample : N4401-02 10X
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
59B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 03:06:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.432 min
Delta R.T.: 0.002 min
Response: 4888070
Conc: 1.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
59B




#1 Tetrachloro-m-xylene

R.T.: 3.588 min
Delta R.T.: -0.001 min
Response: 1455324
Conc: 1.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.260 min
Delta R.T.: 0.003 min
Response: 3540612
Conc: 1.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.572 min
Delta R.T.: 0.000 min
Response: 1461440
Conc: 1.65 ng/ml