

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0083019\
 Data File : P0060164.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 31 Aug 2019 11:42
 Operator : SM/AJ
 Sample : K4603-02
 Misc :
 ALS Vial : 83 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PAR-E1-E4-(0-3.11)

Manual Integrations
 APPROVED

Ankita
 9/3/2019 8:36:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 13:15:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 30 18:12:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.120	3.496	1050978	883246	30.017m	24.538
2) SA Decachlor...	9.575	8.348	913749	722121	17.176	17.909
Target Compounds						
31) L7 AR-1260-1	6.840	5.989	74573	45530	27.663	19.979 #
32) L7 AR-1260-2	7.095	6.176	50765	32995	13.682	11.262
33) L7 AR-1260-3	7.448	6.324	71429	39953	21.801	15.082 #
34) L7 AR-1260-4	7.675	6.785	63283	38500	20.605m	17.016
35) L7 AR-1260-5	7.981	7.029	102294	97427	14.447m	17.703

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0083019\
Data File : P0060164.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 31 Aug 2019 11:42
Operator : SM/AJ
Sample : K4603-02
Misc :
ALS Vial : 83 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PAR-E1-E4-(0-3.11)


Manual Integrations
APPROVED

Ankita
9/3/2019 8:36:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 13:15:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 30 18:12:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.120 min
Delta R.T.: -0.004 min
Response: 1050978
Conc: 30.02 ng/ml m

Instrument :
ECD_O
ClientSampleId :
PAR-E1-E4-(0-3.11)


Manual Integrations
APPROVED

Ankita
9/3/2019 8:36:59 AM




#1 Tetrachloro-m-xylene

R.T.: 3.496 min
Delta R.T.: -0.004 min
Response: 883246
Conc: 24.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.575 min
Delta R.T.: -0.018 min
Response: 913749
Conc: 17.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.348 min
Delta R.T.: -0.011 min
Response: 722121
Conc: 17.91 ng/ml




#31 AR-1260-1

R.T.: 6.840 min
Delta R.T.: -0.010 min
Response: 74573
Conc: 27.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
PAR-E1-E4-(0-3.11)


Manual Integrations
APPROVED

Ankita
9/3/2019 8:36:59 AM




#31 AR-1260-1

R.T.: 5.989 min
Delta R.T.: -0.006 min
Response: 45530
Conc: 19.98 ng/ml




#32 AR-1260-2

R.T.: 7.095 min
Delta R.T.: -0.010 min
Response: 50765
Conc: 13.68 ng/ml



#32 AR-1260-2

R.T.: 6.176 min
Delta R.T.: -0.007 min
Response: 32995
Conc: 11.26 ng/ml




#33 AR-1260-3

R.T.: 7.448 min
Delta R.T.: -0.012 min
Response: 71429
Conc: 21.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
PAR-E1-E4-(0-3.11)


Manual Integrations
APPROVED

Ankita
9/3/2019 8:36:59 AM




#33 AR-1260-3

R.T.: 6.324 min
Delta R.T.: -0.007 min
Response: 39953
Conc: 15.08 ng/ml




#34 AR-1260-4

R.T.: 7.675 min
Delta R.T.: -0.011 min
Response: 63283
Conc: 20.60 ng/ml m



#34 AR-1260-4

R.T.: 6.785 min
Delta R.T.: -0.009 min
Response: 38500
Conc: 17.02 ng/ml




#35 AR-1260-5

R.T.: 7.981 min
Delta R.T.: -0.013 min
Response: 102294
Conc: 14.45 ng/ml m

Instrument :
ECD_O
ClientSampleId :
PAR-E1-E4-(0-3.11)

Manual Integrations
APPROVED

Ankita
9/3/2019 8:36:59 AM



#35 AR-1260-5

R.T.: 7.029 min
Delta R.T.: -0.008 min
Response: 97427
Conc: 17.70 ng/ml