

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0083122\
 Data File : P0089177.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2022 17:25
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 17:41:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.430	3.588	191.2E6	53171896	50.984	55.085
2) SA Decachlor...	10.265	8.574	114.7E6	49992864	49.804	56.534
Target Compounds						
21) L5 AR-1248-1	5.606	4.663	34920912	11156400	507.103	496.737
22) L5 AR-1248-2	5.881	4.899	50157451	15654556	508.214	478.600
23) L5 AR-1248-3	6.087	4.940	53322361	16472191	508.965	483.198
24) L5 AR-1248-4	6.493	5.111	56706409	19311377	492.868	473.781
25) L5 AR-1248-5	6.532	5.503	55718432	19415936	492.280	488.590



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO083122\
Data File : PO089177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2022 17:25
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 17:41:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.430 min
Delta R.T.: 0.000 min
Response: 191171918
Conc: 50.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.588 min
Delta R.T.: -0.001 min
Response: 53171896
Conc: 55.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.265 min
Delta R.T.: 0.009 min
Response: 114665466
Conc: 49.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.574 min
Delta R.T.: 0.002 min
Response: 49992864
Conc: 56.53 ng/ml



#21 AR-1248-1


R.T.: 5.606 min
Delta R.T.: 0.000 min
Response: 34920912
Conc: 507.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.663 min
Delta R.T.: -0.002 min
Response: 11156400
Conc: 496.74 ng/ml




#22 AR-1248-2

R.T.: 5.881 min
Delta R.T.: 0.000 min
Response: 50157451
Conc: 508.21 ng/ml



#22 AR-1248-2


R.T.: 4.899 min
Delta R.T.: -0.002 min
Response: 15654556
Conc: 478.60 ng/ml



#23 AR-1248-3


R.T.: 6.087 min
Delta R.T.: 0.000 min
Response: 53322361
Conc: 508.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.940 min
Delta R.T.: -0.003 min
Response: 16472191
Conc: 483.20 ng/ml




#24 AR-1248-4

R.T.: 6.493 min
Delta R.T.: 0.000 min
Response: 56706409
Conc: 492.87 ng/ml



#24 AR-1248-4


R.T.: 5.111 min
Delta R.T.: -0.002 min
Response: 19311377
Conc: 473.78 ng/ml



#25 AR-1248-5

R.T.: 6.532 min
Delta R.T.: 0.000 min
Response: 55718432
Conc: 492.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.503 min
Delta R.T.: -0.002 min
Response: 19415936
Conc: 488.59 ng/ml