

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0083122\
 Data File : P0089189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2022 20:54
 Operator : YP/AJ
 Sample : N4446-11
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EQUIP-BLANK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 02:19:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.430	3.587	83751464	24025212	22.336	24.890
2) SA Decachlor...	10.264	8.573	32037540	14629999	13.915	16.544

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO083122\
Data File : PO089189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2022 20:54
Operator : YP/AJ
Sample : N4446-11
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EQUIP-BLANK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 02:19:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.430 min
Delta R.T.: 0.000 min
Response: 83751464
Conc: 22.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
EQUIP-BLANK




#1 Tetrachloro-m-xylene

R.T.: 3.587 min
Delta R.T.: -0.002 min
Response: 24025212
Conc: 24.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.264 min
Delta R.T.: 0.007 min
Response: 32037540
Conc: 13.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.573 min
Delta R.T.: 0.000 min
Response: 14629999
Conc: 16.54 ng/ml