

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0090221\
 Data File : P0080907.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2021 2:31
 Operator : DD\AJ
 Sample : P0090221ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 07:22:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	78.942	84.698 E3	-7.3	107	0.00
2 SA Decachlorobiphenyl	59.376	67.335 E3	-13.4	113	0.00
3 L1 AR-1016-1	2.836	3.120 E3	-10.0	111	0.00
4 L1 AR-1016-2	3.990	4.365 E3	-9.4	110	0.00
5 L1 AR-1016-3	2.531	2.735 E3	-8.1	111	0.00
6 L1 AR-1016-4	2.119	2.252 E3	-6.3	112	0.00
7 L1 AR-1016-5	1.963	2.222 E3	-13.2	114	0.00
31 L7 AR-1260-1	3.476	3.916 E3	-12.7	112	0.00
32 L7 AR-1260-2	4.457	4.864 E3	-9.1	113	0.00
33 L7 AR-1260-3	2.624	2.965 E3	-13.0	117	0.00
34 L7 AR-1260-4	3.201	3.646 E3	-13.9	115	0.00
35 L7 AR-1260-5	5.987	6.664 E3	-11.3	112	0.00

Signal #2

1 SA Tetrachloro-m-xylene	24.712	25.405 E3	-2.8	103	0.00
2 SA Decachlorobiphenyl	23.344	25.561 E3	-9.5	111	0.00
3 L1 AR-1016-1	964.520	1016.393	-5.4	106	0.00
4 L1 AR-1016-2	1.337	1.396 E3	-4.4	106	0.00
5 L1 AR-1016-3	714.896	770.180	-7.7	108	0.00
6 L1 AR-1016-4	601.110	649.962	-8.1	110	0.00
7 L1 AR-1016-5	752.761	810.235	-7.6	108	0.00
31 L7 AR-1260-1	1.468	1.554 E3	-5.9	109	0.00
32 L7 AR-1260-2	1.755	1.851 E3	-5.5	108	0.00
33 L7 AR-1260-3	1.831	1.827 E3	0.2	108	0.00
34 L7 AR-1260-4	1.181	1.312 E3	-11.1	110	0.00
35 L7 AR-1260-5	2.831	2.961 E3	-4.6	109	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	971.703	0.000	100.0#	0#	-5.18#
9 L2 AR-1221-2	735.120	0.000	100.0#	0#	-5.28#
10 L2 AR-1221-3	2.276	0.000 E3	100.0#	0#	-5.37#
11 L3 AR-1232-1	1.720	0.000 E3	100.0#	0#	-5.37#
12 L3 AR-1232-2	981.696	0.000	100.0#	0#	-5.96#
13 L3 AR-1232-3	1.810	0.000 E3	100.0#	0#	-6.28#
14 L3 AR-1232-4	968.881	0.000	100.0#	0#	-6.46#
15 L3 AR-1232-5	669.860	0.000	100.0#	0#	-6.56#
16 L4 AR-1242-1	2.189	0.000 E3	100.0#	0#	-6.26#
17 L4 AR-1242-2	3.059	0.000 E3	100.0#	0#	-6.28#
18 L4 AR-1242-3	1.914	0.000 E3	100.0#	0#	-6.35#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0090221\
 Data File : P0080907.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2021 2:31
 Operator : DD\AJ
 Sample : P0090221ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 07:22:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound			AvgRF	CCRF	%Dev	Area%	Dev(Min)

19	L4	AR-1242-4	1.543	0.000 E3	100.0#	0#	-6.46#
20	L4	AR-1242-5	1.602	0.000 E3	100.0#	0#	-7.25#
21	L5	AR-1248-1	1.725	0.000 E3	100.0#	0#	-6.26#
22	L5	AR-1248-2	2.224	0.000 E3	100.0#	0#	-6.56#
23	L5	AR-1248-3	2.646	0.000 E3	100.0#	0#	-6.77#
24	L5	AR-1248-4	2.880	0.000 E3	100.0#	0#	-7.21#
25	L5	AR-1248-5	2.750	0.000 E3	100.0#	0#	-7.25#
26	L6	AR-1254-1	3.232	0.000 E3	100.0#	0#	-7.18#
27	L6	AR-1254-2	4.977	0.000 E3	100.0#	0#	-7.41#
28	L6	AR-1254-3	5.056	0.000 E3	100.0#	0#	-7.79#
29	L6	AR-1254-4	3.737	0.000 E3	100.0#	0#	-8.09#
30	L6	AR-1254-5	4.137	0.000 E3	100.0#	0#	-8.52#
36	L8	AR-1262-1	4.456	0.000 E3	100.0#	0#	-8.59#
37	L8	AR-1262-2	7.663	0.000 E3	100.0#	0#	-9.15#
38	L8	AR-1262-3	3.409	0.000 E3	100.0#	0#	-9.47#
39	L8	AR-1262-4	2.094	0.000 E3	100.0#	0#	-9.56#
40	L8	AR-1262-5	2.771	0.000 E3	100.0#	0#	-10.24#
41	L9	AR-1268-1	9.230	0.000 E3	100.0#	0#	-9.47#
42	L9	AR-1268-2	8.550	0.000 E3	100.0#	0#	-9.56#
43	L9	AR-1268-3	7.115	0.000 E3	100.0#	0#	-9.79#
44	L9	AR-1268-4	3.149	0.000 E3	100.0#	0#	-10.24#
45	L9	AR-1268-5	22.543	0.000 E3	100.0#	0#	-10.66#

Signal #2

8	L2	AR-1221-1	327.339	0.000	100.0#	0#	-4.32#
9	L2	AR-1221-2	246.513	0.000	100.0#	0#	-4.42#
10	L2	AR-1221-3	783.272	0.000	100.0#	0#	-4.51#
11	L3	AR-1232-1	607.155	0.000	100.0#	0#	-4.51#
12	L3	AR-1232-2	612.942	0.000	100.0#	0#	-5.35#
13	L3	AR-1232-3	324.608	0.000	100.0#	0#	-5.54#
14	L3	AR-1232-4	301.584	0.000	100.0#	0#	-5.64#
15	L3	AR-1232-5	332.708	0.000	100.0#	0#	-5.83#
16	L4	AR-1242-1	744.879	0.000	100.0#	0#	-5.33#
17	L4	AR-1242-2	1.020	0.000 E3	100.0#	0#	-5.35#
18	L4	AR-1242-3	554.925	0.000	100.0#	0#	-5.54#
19	L4	AR-1242-4	565.751	0.000	100.0#	0#	-5.64#
20	L4	AR-1242-5	724.247	0.000	100.0#	0#	-6.21#
21	L5	AR-1248-1	592.038	0.000	100.0#	0#	-5.33#
22	L5	AR-1248-2	830.563	0.000	100.0#	0#	-5.59#
23	L5	AR-1248-3	848.730	0.000	100.0#	0#	-5.64#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0090221\
 Data File : P0080907.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2021 2:31
 Operator : DD\AJ
 Sample : P0090221ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 07:22:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	988.698	0.000	100.0#	0#	-5.83#
25 L5 AR-1248-5	906.621	0.000	100.0#	0#	-6.25#
26 L6 AR-1254-1	1.582	0.000 E3	100.0#	0#	-6.21#
27 L6 AR-1254-2	1.446	0.000 E3	100.0#	0#	-6.37#
28 L6 AR-1254-3	2.237	0.000 E3	100.0#	0#	-6.79#
29 L6 AR-1254-4	1.416	0.000 E3	100.0#	0#	-7.04#
30 L6 AR-1254-5	2.110	0.000 E3	100.0#	0#	-7.47#
36 L8 AR-1262-1	1.062	0.000 E3	100.0#	0#	-7.46#
37 L8 AR-1262-2	3.291	0.000 E3	100.0#	0#	-8.02#
38 L8 AR-1262-3	1.340	0.000 E3	100.0#	0#	-8.31#
39 L8 AR-1262-4	2.372	0.000 E3	100.0#	0#	-8.37#
40 L8 AR-1262-5	1.158	0.000 E3	100.0#	0#	-8.87#
41 L9 AR-1268-1	3.779	0.000 E3	100.0#	0#	-8.31#
42 L9 AR-1268-2	3.394	0.000 E3	100.0#	0#	-8.37#
43 L9 AR-1268-3	2.954	0.000 E3	100.0#	0#	-8.58#
44 L9 AR-1268-4	1.316	0.000 E3	100.0#	0#	-8.87#
45 L9 AR-1268-5	8.504	0.000 E3	100.0#	0#	-9.17#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0