

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0090222\
 Data File : P0089220.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2022 15:27
 Operator : YP/AJ
 Sample : N4479-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 SB-14

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/06/2022
 Supervised By :Ankita Jodhani 09/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 02 16:08:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.430	3.587	81184083	19828358	21.651	20.542
2)	SA Decachlor...	10.261	8.569	47009883	20317162	20.419	22.976
Target Compounds							
16)	L4 AR-1242-1	5.604	4.661	138.7E6	45390880	1569.017	1558.194
17)	L4 AR-1242-2	5.627	4.679	189.4E6	56780627	1530.099	1433.325
18)	L4 AR-1242-3	5.690	4.854	116.6E6	33041528	1516.563	1550.714
19)	L4 AR-1242-4	5.789	4.938	104.0E6	28873904	1673.807	1326.994
20)	L4 AR-1242-5	6.530	5.459	42131448	19582109	654.496	716.840
31)	L7 AR-1260-1	7.217	6.140	8104950	3566790	60.190	63.575m
32)	L7 AR-1260-2	7.475	6.328	9673515	3452688	63.386	50.811
33)	L7 AR-1260-3	7.837	6.480	4183988	3235030	37.444	52.727 #
34)	L7 AR-1260-4	8.061	6.950	7146732	2298554	57.892	45.613m
35)	L7 AR-1260-5	8.390	7.193	8208764	3326963	34.410	29.078m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090222\
Data File : PO089220.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2022 15:27
Operator : YP/AJ
Sample : N4479-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-14

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/06/2022
Supervised By :Ankita Jodhani 09/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 02 16:08:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

