

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090222\
 Data File : P0089237.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2022 20:37
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/06/2022
 Supervised By :Ankita Jodhani 09/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 04:07:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.585	194.5E6	53859903	51.876	55.798
2) SA Decachlor...	10.252	8.566	103.6E6	48271232	44.985m	54.587
Target Compounds						
3) L1 AR-1016-1	5.603	4.661	56361282	18589461	501.275	514.755
4) L1 AR-1016-2	5.625	4.679	78221431	26142625	496.334	529.531
5) L1 AR-1016-3	5.687	4.854	49276319	14187128	505.254	533.408
6) L1 AR-1016-4	5.787	4.896	39541420	11523336	502.518	535.390
7) L1 AR-1016-5	6.083	5.108	38148527	14857617	497.091	509.210
31) L7 AR-1260-1	7.215	6.138	66121973	28788568	491.046	513.135
32) L7 AR-1260-2	7.472	6.326	72366312	34094822	474.180	501.748
33) L7 AR-1260-3	7.833	6.478	54296185	31102713	485.910	506.940
34) L7 AR-1260-4	8.061	6.949	59615319	26035494	482.914	516.655
35) L7 AR-1260-5	8.386	7.192	111.7E6	60683302	468.197	530.375

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090222\
Data File : PO089237.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2022 20:37
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/06/2022
Supervised By :Ankita Jodhani 09/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 04:07:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

