

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090319\
 Data File : P0060291.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Sep 2019 7:00
 Operator : SM/AJ
 Sample : K4627-02 5X
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 6-SWG-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 07:18:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 30 18:12:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.112	3.492	350793	355705	10.019	9.882
2)	SA Decachlor...	9.561	8.342	286542	388405	5.386	9.633 #
Target Compounds							
3)	L1 AR-1016-1	5.277	4.548	53360	210837	33.270	155.951 #
4)	L1 AR-1016-2	5.277	4.548	53360	210837	22.303	105.260 #
5)	L1 AR-1016-3	5.351	4.694	92893	23454	62.364	21.679 #
6)	L1 AR-1016-4	0.000	4.825	0	67490	N.D.	74.740 #
7)	L1 AR-1016-5	5.662	4.927	39774	22838	30.973	19.826 #
8)	L2 AR-1221-1	4.365	3.705	143048	182712	337.799	455.203 #
9)	L2 AR-1221-2	4.416	3.782	469102	296597	1444.305	973.650 #
10)	L2 AR-1221-3	4.463	3.856	303740	567692	288.764	580.197 #
11)	L3 AR-1232-1	4.463	3.856	303740	567692	243.687	497.920 #
12)	L3 AR-1232-2	5.008	4.548	370256	210837	524.148	165.547 #
13)	L3 AR-1232-3	5.277	4.694	53360	23454	34.845	34.445
14)	L3 AR-1232-4	0.000	4.825	0	67490	N.D.	110.629 #
15)	L3 AR-1232-5	0.000	4.927	0	22838	N.D.	33.550 #
16)	L4 AR-1242-1	5.277	4.548	53360	210837	42.032	204.963 #
17)	L4 AR-1242-2	5.277	4.548	53360	210837	28.755	138.427 #
18)	L4 AR-1242-3	5.351	4.694	92893	23454	79.474	28.071 #
19)	L4 AR-1242-4	0.000	4.825	0	67490	N.D.	80.886 #
20)	L4 AR-1242-5	6.239	5.318	1467664	683394	1167.967	610.936 #
21)	L5 AR-1248-1	5.277	4.548	53360	210837	53.224	254.626 #
22)	L5 AR-1248-2	0.000	4.825	0	67490	N.D.	58.528 #
23)	L5 AR-1248-3	5.662	4.825	39774	67490	23.931	56.822 #
24)	L5 AR-1248-4	6.125	4.927	254248	22838	124.381	15.717 #
25)	L5 AR-1248-5	6.125	5.318	254248	683394	128.209	477.469 #
26)	L6 AR-1254-1	6.125	5.318	254248	683394	121.859	288.133 #
27)	L6 AR-1254-2	6.309	5.464	630793	513002	189.073	244.484 #
28)	L6 AR-1254-3	6.673	5.891	617419	157874	173.996	46.951 #
29)	L6 AR-1254-4	6.953	6.093	1895420	352544	643.489	170.986 #
30)	L6 AR-1254-5	7.407	6.496	572800	375179	185.229	120.832 #
31)	L7 AR-1260-1	6.862	6.003	316900	548549	117.552	240.707 #
32)	L7 AR-1260-2	7.152	6.214	2563274	574817	690.820	196.192 #
33)	L7 AR-1260-3	7.407	6.346	572800	364038	174.830	137.418
34)	L7 AR-1260-4	7.706	6.747	2246812	209336	731.546	92.521 #
35)	L7 AR-1260-5	7.936	7.076	2771339	845463	391.382	153.628 #
36)	L8 AR-1262-1	7.544	6.496	3305499	375179	681.139	100.650 #
37)	L8 AR-1262-2	8.082	7.076	2942854	845463	381.402	143.810 #
38)	L8 AR-1262-3	8.294	7.314	1734586	574625	339.961	226.144 #
39)	L8 AR-1262-4	8.314	7.355	1117413	199388	289.881	45.121 #
40)	L8 AR-1262-5	8.950	7.868	866320	676187	337.541	374.503
41)	L9 AR-1268-1	8.294	7.314	1734586	574625	184.321	80.695 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090319\
 Data File : P0060291.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Sep 2019 7:00
 Operator : SM/AJ
 Sample : K4627-02 5X
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 6-SWG-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 07:18:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 30 18:12:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.314	7.355	1117413	199388	141.755	30.922 #
43) L9	AR-1268-3	8.585	7.589	853159	60536	121.036	11.064 #
44) L9	AR-1268-4	8.950	7.868	866320	676187	311.621	331.254
45) L9	AR-1268-5	9.289	8.166	787725	331603	42.713	22.823 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090319\
Data File : P0060291.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Sep 2019 7:00
Operator : SM/AJ
Sample : K4627-02 5X
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
6-SWG-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 07:18:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 30 18:12:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.112 min
Delta R.T.: -0.012 min
Response: 350793
Conc: 10.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#1 Tetrachloro-m-xylene

R.T.: 3.492 min
Delta R.T.: -0.008 min
Response: 355705
Conc: 9.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.561 min
Delta R.T.: -0.032 min
Response: 286542
Conc: 5.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.342 min
Delta R.T.: -0.017 min
Response: 388405
Conc: 9.63 ng/ml



#3 AR-1016-1


R.T.: 5.277 min
Delta R.T.: 0.017 min
Response: 53360
Conc: 33.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#3 AR-1016-1

R.T.: 4.548 min
Delta R.T.: 0.008 min
Response: 210837
Conc: 155.95 ng/ml




#4 AR-1016-2

R.T.: 5.277 min
Delta R.T.: -0.003 min
Response: 53360
Conc: 22.30 ng/ml



#4 AR-1016-2


R.T.: 4.548 min
Delta R.T.: -0.012 min
Response: 210837
Conc: 105.26 ng/ml



#5 AR-1016-3


R.T.: 5.351 min
Delta R.T.: 0.011 min
Response: 92893
Conc: 62.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#5 AR-1016-3

R.T.: 4.694 min
Delta R.T.: -0.036 min
Response: 23454
Conc: 21.68 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 5.440 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 4.825 min
Delta R.T.: 0.055 min
Response: 67490
Conc: 74.74 ng/ml



#7 AR-1016-5


R.T.: 5.662 min
Delta R.T.: -0.068 min
Response: 39774
Conc: 30.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#7 AR-1016-5

R.T.: 4.927 min
Delta R.T.: -0.053 min
Response: 22838
Conc: 19.83 ng/ml




#8 AR-1221-1

R.T.: 4.365 min
Delta R.T.: -0.012 min
Response: 143048
Conc: 337.80 ng/ml



#8 AR-1221-1


R.T.: 3.705 min
Delta R.T.: -0.011 min
Response: 182712
Conc: 455.20 ng/ml



#9 AR-1221-2


R.T.: 4.416 min
Delta R.T.: -0.005 min
Response: 469102
Conc: 1444.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#9 AR-1221-2

R.T.: 3.782 min
Delta R.T.: -0.007 min
Response: 296597
Conc: 973.65 ng/ml




#10 AR-1221-3

R.T.: 4.463 min
Delta R.T.: -0.033 min
Response: 303740
Conc: 288.76 ng/ml



#10 AR-1221-3


R.T.: 3.856 min
Delta R.T.: -0.005 min
Response: 567692
Conc: 580.20 ng/ml



#11 AR-1232-1


R.T.: 4.463 min
Delta R.T.: -0.033 min
Response: 303740
Conc: 243.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#11 AR-1232-1

R.T.: 3.856 min
Delta R.T.: -0.005 min
Response: 567692
Conc: 497.92 ng/ml




#12 AR-1232-2

R.T.: 5.008 min
Delta R.T.: -0.004 min
Response: 370256
Conc: 524.15 ng/ml



#12 AR-1232-2


R.T.: 4.548 min
Delta R.T.: -0.018 min
Response: 210837
Conc: 165.55 ng/ml



#13 AR-1232-3


R.T.: 5.277 min
Delta R.T.: -0.018 min
Response: 53360
Conc: 34.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#13 AR-1232-3

R.T.: 4.694 min
Delta R.T.: -0.044 min
Response: 23454
Conc: 34.44 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 5.540 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 4.825 min
Delta R.T.: -0.023 min
Response: 67490
Conc: 110.63 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 5.583 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#15 AR-1232-5

R.T.: 4.927 min
Delta R.T.: -0.060 min
Response: 22838
Conc: 33.55 ng/ml




#16 AR-1242-1

R.T.: 5.277 min
Delta R.T.: -0.018 min
Response: 53360
Conc: 42.03 ng/ml



#16 AR-1242-1


R.T.: 4.548 min
Delta R.T.: -0.018 min
Response: 210837
Conc: 204.96 ng/ml



#17 AR-1242-2


R.T.: 5.277 min
Delta R.T.: -0.018 min
Response: 53360
Conc: 28.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#17 AR-1242-2

R.T.: 4.548 min
Delta R.T.: -0.018 min
Response: 210837
Conc: 138.43 ng/ml




#18 AR-1242-3

R.T.: 5.351 min
Delta R.T.: -0.005 min
Response: 92893
Conc: 79.47 ng/ml



#18 AR-1242-3


R.T.: 4.694 min
Delta R.T.: -0.044 min
Response: 23454
Conc: 28.07 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 5.540 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#19 AR-1242-4

R.T.: 4.825 min
Delta R.T.: -0.023 min
Response: 67490
Conc: 80.89 ng/ml




#20 AR-1242-5

R.T.: 6.239 min
Delta R.T.: 0.020 min
Response: 1467664
Conc: 1167.97 ng/ml



#20 AR-1242-5


R.T.: 5.318 min
Delta R.T.: -0.011 min
Response: 683394
Conc: 610.94 ng/ml



#21 AR-1248-1


R.T.: 5.277 min
Delta R.T.: 0.002 min
Response: 53360
Conc: 53.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#21 AR-1248-1

R.T.: 4.548 min
Delta R.T.: -0.001 min
Response: 210837
Conc: 254.63 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T. : 5.540 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 4.825 min
Delta R.T.: 0.046 min
Response: 67490
Conc: 58.53 ng/ml



#23 AR-1248-3


R.T.: 5.662 min
Delta R.T.: -0.044 min
Response: 39774
Conc: 23.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#23 AR-1248-3

R.T.: 4.825 min
Delta R.T.: 0.006 min
Response: 67490
Conc: 56.82 ng/ml




#24 AR-1248-4

R.T.: 6.125 min
Delta R.T.: -0.016 min
Response: 254248
Conc: 124.38 ng/ml



#24 AR-1248-4


R.T.: 4.927 min
Delta R.T.: -0.060 min
Response: 22838
Conc: 15.72 ng/ml



#25 AR-1248-5


R.T.: 6.125 min
Delta R.T.: -0.053 min
Response: 254248
Conc: 128.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#25 AR-1248-5

R.T.: 5.318 min
Delta R.T.: -0.011 min
Response: 683394
Conc: 477.47 ng/ml




#26 AR-1254-1

R.T.: 6.125 min
Delta R.T.: 0.014 min
Response: 254248
Conc: 121.86 ng/ml



#26 AR-1254-1


R.T.: 5.318 min
Delta R.T.: -0.011 min
Response: 683394
Conc: 288.13 ng/ml



#27 AR-1254-2


R.T.: 6.309 min
Delta R.T.: -0.016 min
Response: 630793
Conc: 189.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#27 AR-1254-2

R.T.: 5.464 min
Delta R.T.: -0.010 min
Response: 513002
Conc: 244.48 ng/ml




#28 AR-1254-3

R.T.: 6.673 min
Delta R.T.: -0.019 min
Response: 617419
Conc: 174.00 ng/ml



#28 AR-1254-3


R.T.: 5.891 min
Delta R.T.: 0.010 min
Response: 157874
Conc: 46.95 ng/ml



#29 AR-1254-4


R.T.: 6.953 min
Delta R.T.: -0.022 min
Response: 1895420
Conc: 643.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#29 AR-1254-4

R.T.: 6.093 min
Delta R.T.: -0.003 min
Response: 352544
Conc: 170.99 ng/ml




#30 AR-1254-5

R.T.: 7.407 min
Delta R.T.: 0.017 min
Response: 572800
Conc: 185.23 ng/ml



#30 AR-1254-5


R.T.: 6.496 min
Delta R.T.: -0.008 min
Response: 375179
Conc: 120.83 ng/ml



#31 AR-1260-1


R.T.: 6.862 min
Delta R.T.: 0.012 min
Response: 316900
Conc: 117.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#31 AR-1260-1

R.T.: 6.003 min
Delta R.T.: 0.008 min
Response: 548549
Conc: 240.71 ng/ml




#32 AR-1260-2

R.T.: 7.152 min
Delta R.T.: 0.047 min
Response: 2563274
Conc: 690.82 ng/ml



#32 AR-1260-2


R.T.: 6.214 min
Delta R.T.: 0.030 min
Response: 574817
Conc: 196.19 ng/ml



#33 AR-1260-3


R.T.: 7.407 min
Delta R.T.: -0.052 min
Response: 572800
Conc: 174.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#33 AR-1260-3

R.T.: 6.346 min
Delta R.T.: 0.015 min
Response: 364038
Conc: 137.42 ng/ml




#34 AR-1260-4

R.T.: 7.706 min
Delta R.T.: 0.019 min
Response: 2246812
Conc: 731.55 ng/ml



#34 AR-1260-4


R.T.: 6.747 min
Delta R.T.: -0.047 min
Response: 209336
Conc: 92.52 ng/ml



#35 AR-1260-5


R.T.: 7.936 min
Delta R.T.: -0.058 min
Response: 2771339
Conc: 391.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#35 AR-1260-5

R.T.: 7.076 min
Delta R.T.: 0.039 min
Response: 845463
Conc: 153.63 ng/ml




#36 AR-1262-1

R.T.: 7.544 min
Delta R.T.: 0.027 min
Response: 3305499
Conc: 681.14 ng/ml



#36 AR-1262-1


R.T.: 6.496 min
Delta R.T.: -0.042 min
Response: 375179
Conc: 100.65 ng/ml



#37 AR-1262-2


R.T.: 8.082 min
Delta R.T.: 0.044 min
Response: 2942854
Conc: 381.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#37 AR-1262-2

R.T.: 7.076 min
Delta R.T.: 0.039 min
Response: 845463
Conc: 143.81 ng/ml




#38 AR-1262-3

R.T.: 8.294 min
Delta R.T.: 0.000 min
Response: 1734586
Conc: 339.96 ng/ml



#38 AR-1262-3


R.T.: 7.314 min
Delta R.T.: 0.003 min
Response: 574625
Conc: 226.14 ng/ml



#39 AR-1262-4


R.T.: 8.314 min
Delta R.T.: -0.027 min
Response: 1117413
Conc: 289.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#39 AR-1262-4

R.T.: 7.355 min
Delta R.T.: -0.021 min
Response: 199388
Conc: 45.12 ng/ml




#40 AR-1262-5

R.T.: 8.950 min
Delta R.T.: 0.000 min
Response: 866320
Conc: 337.54 ng/ml



#40 AR-1262-5


R.T.: 7.868 min
Delta R.T.: 0.000 min
Response: 676187
Conc: 374.50 ng/ml



#41 AR-1268-1


R.T.: 8.294 min
Delta R.T.: 0.000 min
Response: 1734586
Conc: 184.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#41 AR-1268-1

R.T.: 7.314 min
Delta R.T.: 0.003 min
Response: 574625
Conc: 80.70 ng/ml




#42 AR-1268-2

R.T.: 8.314 min
Delta R.T.: -0.027 min
Response: 1117413
Conc: 141.76 ng/ml



#42 AR-1268-2


R.T.: 7.355 min
Delta R.T.: -0.021 min
Response: 199388
Conc: 30.92 ng/ml



#43 AR-1268-3


R.T.: 8.585 min
Delta R.T.: -0.018 min
Response: 853159
Conc: 121.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1




#43 AR-1268-3

R.T.: 7.589 min
Delta R.T.: -0.004 min
Response: 60536
Conc: 11.06 ng/ml




#44 AR-1268-4

R.T.: 8.950 min
Delta R.T.: 0.000 min
Response: 866320
Conc: 311.62 ng/ml



#44 AR-1268-4


R.T.: 7.868 min
Delta R.T.: 0.000 min
Response: 676187
Conc: 331.25 ng/ml



#45 AR-1268-5

R.T.: 9.289 min
Delta R.T.: -0.059 min
Response: 787725
Conc: 42.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
6-SWG-1



#45 AR-1268-5

R.T.: 8.166 min
Delta R.T.: 0.032 min
Response: 331603
Conc: 22.82 ng/ml