

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090319\
 Data File : P0060294.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Sep 2019 7:48
 Operator : SM/AJ
 Sample : K4627-05 5X
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 6-EWG-1

Manual Integrations
 APPROVED

Ankita
 9/4/2019 4:58:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 08:20:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 30 18:12:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.109	3.491	244866	223443	6.994m	6.208
2) SA Decachlor...	9.559	8.340	507195	390502	9.534	9.685m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090319\
Data File : P0060294.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Sep 2019 7:48
Operator : SM/AJ
Sample : K4627-05 5X
Misc :
ALS Vial : 69 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
6-EWG-1


Manual Integrations
APPROVED

Ankita
9/4/2019 4:58:43 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 08:20:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 30 18:12:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.109 min
Delta R.T.: -0.015 min
Response: 244866
Conc: 6.99 ng/ml m

Instrument :
ECD_O
ClientSampleId :
6-EWG-1


Manual Integrations
APPROVED

Ankita
9/4/2019 4:58:43 PM




#1 Tetrachloro-m-xylene

R.T.: 3.491 min
Delta R.T.: -0.008 min
Response: 223443
Conc: 6.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.559 min
Delta R.T.: -0.034 min
Response: 507195
Conc: 9.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.340 min
Delta R.T.: -0.019 min
Response: 390502
Conc: 9.68 ng/ml m