

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090321\
 Data File : P0080916.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2021 10:19
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 03:58:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.927	4.053	4239683	1285362	53.706	52.014
2) SA Decachlor...	10.991	9.416	3327557	1260960	56.042	54.016
Target Compounds						
21) L5 AR-1248-1	6.258	5.325	905818	298937	524.991	504.928
22) L5 AR-1248-2	6.555	5.591	1197458	420449	538.406	506.222
23) L5 AR-1248-3	6.773	5.636	1416007	427952	535.061	504.227
24) L5 AR-1248-4	7.205	5.823	1538217	503111	534.076	508.862
25) L5 AR-1248-5	7.244	6.249	1480676	476899	538.407	526.019



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090321\
Data File : P0080916.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Sep 2021 10:19
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 03:58:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 4239683
Conc: 53.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.053 min
Delta R.T.: -0.003 min
Response: 1285362
Conc: 52.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.991 min
Delta R.T.: -0.007 min
Response: 3327557
Conc: 56.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.416 min
Delta R.T.: -0.004 min
Response: 1260960
Conc: 54.02 ng/ml



#21 AR-1248-1


R.T.: 6.258 min
Delta R.T.: -0.001 min
Response: 905818
Conc: 524.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.325 min
Delta R.T.: -0.002 min
Response: 298937
Conc: 504.93 ng/ml




#22 AR-1248-2

R.T.: 6.555 min
Delta R.T.: 0.000 min
Response: 1197458
Conc: 538.41 ng/ml



#22 AR-1248-2


R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 420449
Conc: 506.22 ng/ml



#23 AR-1248-3


R.T.: 6.773 min
Delta R.T.: -0.002 min
Response: 1416007
Conc: 535.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.636 min
Delta R.T.: -0.002 min
Response: 427952
Conc: 504.23 ng/ml




#24 AR-1248-4

R.T.: 7.205 min
Delta R.T.: -0.002 min
Response: 1538217
Conc: 534.08 ng/ml



#24 AR-1248-4


R.T.: 5.823 min
Delta R.T.: -0.002 min
Response: 503111
Conc: 508.86 ng/ml



#25 AR-1248-5

R.T.: 7.244 min
Delta R.T.: -0.002 min
Response: 1480676
Conc: 538.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.249 min
Delta R.T.: -0.002 min
Response: 476899
Conc: 526.02 ng/ml