

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090321\
 Data File : P0080965.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Sep 2021 1:55
 Operator : DD\AJ
 Sample : M3627-06
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 5-6-FEEDER

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 07:05:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.927	4.053	1994065	626016	25.260	25.332
2) SA Decachlor...	10.992	9.415	1629915	576278	27.451	24.686
Target Compounds						
26) L6 AR-1254-1	7.176	6.205	41487	17450	12.837	11.031
27) L6 AR-1254-2	7.405	6.365	69593	15522	13.983	10.736
28) L6 AR-1254-3	7.788	6.788	104893	27563	20.745	12.320m#
29) L6 AR-1254-4	8.085	7.030	52326	15765	14.002	11.137
30) L6 AR-1254-5	8.514	7.462	83688	50474	20.231	23.922
31) L7 AR-1260-1	7.955	6.927	45143	29802	12.987	20.307 #
32) L7 AR-1260-2	8.220	7.125	110080	26697	24.700	15.210 #
33) L7 AR-1260-3	8.587	7.286	60337	79370	22.992	43.355 #
34) L7 AR-1260-4	8.818	7.764	68073	17795	21.266	15.063 #
35) L7 AR-1260-5	9.149	8.016	85732	35027	14.321m	12.374m



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090321\
Data File : P0080965.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Sep 2021 1:55
Operator : DD\AJ
Sample : M3627-06
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
5-6-FEEDER

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 07:05:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 1994065
Conc: 25.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
5-6-FEEDER




#1 Tetrachloro-m-xylene

R.T.: 4.053 min
Delta R.T.: -0.002 min
Response: 626016
Conc: 25.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.992 min
Delta R.T.: -0.007 min
Response: 1629915
Conc: 27.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.415 min
Delta R.T.: -0.005 min
Response: 576278
Conc: 24.69 ng/ml



#26 AR-1254-1


R.T.: 7.176 min
Delta R.T.: 0.000 min
Response: 41487
Conc: 12.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
5-6-FEEDER




#26 AR-1254-1

R.T.: 6.205 min
Delta R.T.: -0.004 min
Response: 17450
Conc: 11.03 ng/ml




#27 AR-1254-2

R.T.: 7.405 min
Delta R.T.: -0.003 min
Response: 69593
Conc: 13.98 ng/ml



#27 AR-1254-2


R.T.: 6.365 min
Delta R.T.: -0.004 min
Response: 15522
Conc: 10.74 ng/ml



#28 AR-1254-3


R.T.: 7.788 min
Delta R.T.: -0.004 min
Response: 104893
Conc: 20.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
5-6-FEEDER




#28 AR-1254-3

R.T.: 6.788 min
Delta R.T.: -0.007 min
Response: 27563
Conc: 12.32 ng/ml m




#29 AR-1254-4

R.T.: 8.085 min
Delta R.T.: -0.004 min
Response: 52326
Conc: 14.00 ng/ml



#29 AR-1254-4


R.T.: 7.030 min
Delta R.T.: -0.005 min
Response: 15765
Conc: 11.14 ng/ml



#30 AR-1254-5


R.T.: 8.514 min
Delta R.T.: -0.005 min
Response: 83688
Conc: 20.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
5-6-FEEDER




#30 AR-1254-5

R.T.: 7.462 min
Delta R.T.: -0.005 min
Response: 50474
Conc: 23.92 ng/ml




#31 AR-1260-1

R.T.: 7.955 min
Delta R.T.: -0.003 min
Response: 45143
Conc: 12.99 ng/ml



#31 AR-1260-1


R.T.: 6.927 min
Delta R.T.: -0.004 min
Response: 29802
Conc: 20.31 ng/ml



#32 AR-1260-2


R.T.: 8.220 min
Delta R.T.: -0.005 min
Response: 110080
Conc: 24.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
5-6-FEEDER




#32 AR-1260-2

R.T.: 7.125 min
Delta R.T.: -0.004 min
Response: 26697
Conc: 15.21 ng/ml




#33 AR-1260-3

R.T.: 8.587 min
Delta R.T.: -0.005 min
Response: 60337
Conc: 22.99 ng/ml



#33 AR-1260-3


R.T.: 7.286 min
Delta R.T.: 0.000 min
Response: 79370
Conc: 43.36 ng/ml



#34 AR-1260-4


R.T.: 8.818 min
Delta R.T.: -0.006 min
Response: 68073
Conc: 21.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
5-6-FEEDER




#34 AR-1260-4

R.T.: 7.764 min
Delta R.T.: -0.008 min
Response: 17795
Conc: 15.06 ng/ml



#35 AR-1260-5

R.T.: 9.149 min
Delta R.T.: -0.006 min
Response: 85732
Conc: 14.32 ng/ml m



#35 AR-1260-5

R.T.: 8.016 min
Delta R.T.: -0.004 min
Response: 35027
Conc: 12.37 ng/ml m