

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090324\
 Data File : P0106080.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2024 08:59
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 12:04:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.718	473.3E6	143.9E6	51.877	53.262
2) SA Decachlor...	10.191	8.749	291.3E6	118.6E6	53.671	54.325
Target Compounds						
3) L1 AR-1016-1	5.616	4.807	152.7E6	50044752	483.127	523.474
4) L1 AR-1016-2	5.638	4.827	215.0E6	66789752	489.165	526.299
5) L1 AR-1016-3	5.701	5.004	130.0E6	35153150	483.591	502.511
6) L1 AR-1016-4	5.799	5.044	107.7E6	26964259	489.445	501.424
7) L1 AR-1016-5	6.093	5.260	102.5E6	33878964	481.915	492.348
31) L7 AR-1260-1	7.216	6.294	179.6E6	69044471	500.360	524.088
32) L7 AR-1260-2	7.472	6.480	208.2E6	85397561	495.590	520.368
33) L7 AR-1260-3	7.830	6.635	135.5E6	79536510	506.629	471.281
34) L7 AR-1260-4	8.056	7.107	160.0E6	56334175	516.048	508.335
35) L7 AR-1260-5	8.375	7.346	301.9E6	141.3E6	542.722	520.803



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090324\
Data File : PO106080.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2024 08:59
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 12:04:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 06:50:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: -0.002 min
Response: 473289492
Conc: 51.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.718 min
Delta R.T.: -0.004 min
Response: 143901761
Conc: 53.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.191 min
Delta R.T.: -0.005 min
Response: 291267189
Conc: 53.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.749 min
Delta R.T.: -0.007 min
Response: 118592311
Conc: 54.32 ng/ml



#3 AR-1016-1


R.T.: 5.616 min
Delta R.T.: -0.002 min
Response: 152693573
Conc: 483.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.807 min
Delta R.T.: -0.005 min
Response: 50044752
Conc: 523.47 ng/ml




#4 AR-1016-2

R.T.: 5.638 min
Delta R.T.: -0.002 min
Response: 215031923
Conc: 489.17 ng/ml



#4 AR-1016-2


R.T.: 4.827 min
Delta R.T.: -0.005 min
Response: 66789752
Conc: 526.30 ng/ml



#5 AR-1016-3


R.T.: 5.701 min
Delta R.T.: -0.002 min
Response: 130013516
Conc: 483.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.004 min
Delta R.T.: -0.005 min
Response: 35153150
Conc: 502.51 ng/ml




#6 AR-1016-4

R.T.: 5.799 min
Delta R.T.: -0.002 min
Response: 107744242
Conc: 489.45 ng/ml



#6 AR-1016-4


R.T.: 5.044 min
Delta R.T.: -0.005 min
Response: 26964259
Conc: 501.42 ng/ml



#7 AR-1016-5


R.T.: 6.093 min
Delta R.T.: -0.001 min
Response: 102528746
Conc: 481.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.260 min
Delta R.T.: -0.005 min
Response: 33878964
Conc: 492.35 ng/ml




#31 AR-1260-1

R.T.: 7.216 min
Delta R.T.: -0.001 min
Response: 179559185
Conc: 500.36 ng/ml



#31 AR-1260-1


R.T.: 6.294 min
Delta R.T.: -0.006 min
Response: 69044471
Conc: 524.09 ng/ml



#32 AR-1260-2


R.T.: 7.472 min
Delta R.T.: -0.002 min
Response: 208180823
Conc: 495.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.480 min
Delta R.T.: -0.007 min
Response: 85397561
Conc: 520.37 ng/ml




#33 AR-1260-3

R.T.: 7.830 min
Delta R.T.: -0.002 min
Response: 135498017
Conc: 506.63 ng/ml



#33 AR-1260-3


R.T.: 6.635 min
Delta R.T.: -0.007 min
Response: 79536510
Conc: 471.28 ng/ml



#34 AR-1260-4


R.T.: 8.056 min
Delta R.T.: -0.002 min
Response: 159982978
Conc: 516.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.107 min
Delta R.T.: -0.007 min
Response: 56334175
Conc: 508.33 ng/ml



#35 AR-1260-5

R.T.: 8.375 min
Delta R.T.: -0.001 min
Response: 301931293
Conc: 542.72 ng/ml



#35 AR-1260-5

R.T.: 7.346 min
Delta R.T.: -0.007 min
Response: 141316970
Conc: 520.80 ng/ml