

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090324\
 Data File : P0106082.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2024 09:51
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 12:05:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.718	462.3E6	143.8E6	50.675	53.242
2) SA Decachlor...	10.191	8.749	317.4E6	135.9E6	58.493	62.258
Target Compounds						
26) L6 AR-1254-1	6.469	5.612	156.9E6	71572849	508.234	530.472
27) L6 AR-1254-2	6.687	5.759	230.0E6	63234771	499.145	521.005
28) L6 AR-1254-3	7.052	6.165	232.1E6	102.5E6	491.257	531.706
29) L6 AR-1254-4	7.336	6.392	178.9E6	65541309	513.662	540.445
30) L6 AR-1254-5	7.754	6.811	184.4E6	93003683	504.906	526.015



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090324\
Data File : PO106082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2024 09:51
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 12:05:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 06:50:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: -0.002 min
Response: 462320422
Conc: 50.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.718 min
Delta R.T.: -0.004 min
Response: 143845604
Conc: 53.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.191 min
Delta R.T.: -0.005 min
Response: 317435014
Conc: 58.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.749 min
Delta R.T.: -0.008 min
Response: 135911973
Conc: 62.26 ng/ml



#26 AR-1254-1


R.T.: 6.469 min
Delta R.T.: -0.002 min
Response: 156895380
Conc: 508.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.612 min
Delta R.T.: -0.004 min
Response: 71572849
Conc: 530.47 ng/ml




#27 AR-1254-2

R.T.: 6.687 min
Delta R.T.: -0.001 min
Response: 229968390
Conc: 499.15 ng/ml



#27 AR-1254-2


R.T.: 5.759 min
Delta R.T.: -0.005 min
Response: 63234771
Conc: 521.00 ng/ml



#28 AR-1254-3


R.T.: 7.052 min
Delta R.T.: -0.001 min
Response: 232076872
Conc: 491.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.165 min
Delta R.T.: -0.005 min
Response: 102491015
Conc: 531.71 ng/ml




#29 AR-1254-4

R.T.: 7.336 min
Delta R.T.: -0.002 min
Response: 178933070
Conc: 513.66 ng/ml



#29 AR-1254-4


R.T.: 6.392 min
Delta R.T.: -0.005 min
Response: 65541309
Conc: 540.44 ng/ml



#30 AR-1254-5

R.T.: 7.754 min
Delta R.T.: -0.001 min
Response: 184368590
Conc: 504.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.811 min
Delta R.T.: -0.004 min
Response: 93003683
Conc: 526.01 ng/ml