

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090324\
 Data File : P0106101.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2024 16:24
 Operator : YP/AJ
 Sample : P3798-04 10X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BQ71242-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 16:40:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.719	8958084	4057427	0.982	1.502 #
2) SA Decachlor...	10.193	8.750	7930683	3214398	1.461	1.472
Target Compounds						
26) L6 AR-1254-1	6.470	5.613	25100944	11842702	81.310	87.774
27) L6 AR-1254-2	6.687	5.760	36211058	13117752	78.596	108.080 #
28) L6 AR-1254-3	7.053	6.165	33854102	17135128	71.662	88.894
29) L6 AR-1254-4	7.338	6.392	26693990	9882874	76.630	81.493
30) L6 AR-1254-5	7.755	6.811	29481776	15383629	80.738	87.007



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090324\
Data File : PO106101.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2024 16:24
Operator : YP/AJ
Sample : P3798-04 10X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BQ71242-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 16:40:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 06:50:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: -0.002 min
Response: 8958084
Conc: 0.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
BQ71242-1-2




#1 Tetrachloro-m-xylene

R.T.: 3.719 min
Delta R.T.: -0.003 min
Response: 4057427
Conc: 1.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.193 min
Delta R.T.: -0.003 min
Response: 7930683
Conc: 1.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.750 min
Delta R.T.: -0.007 min
Response: 3214398
Conc: 1.47 ng/ml



#26 AR-1254-1


R.T.: 6.470 min
Delta R.T.: -0.001 min
Response: 25100944
Conc: 81.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
BQ71242-1-2




#26 AR-1254-1

R.T.: 5.613 min
Delta R.T.: -0.003 min
Response: 11842702
Conc: 87.77 ng/ml




#27 AR-1254-2

R.T.: 6.687 min
Delta R.T.: 0.000 min
Response: 36211058
Conc: 78.60 ng/ml



#27 AR-1254-2


R.T.: 5.760 min
Delta R.T.: -0.004 min
Response: 13117752
Conc: 108.08 ng/ml



#28 AR-1254-3


R.T.: 7.053 min
Delta R.T.: 0.000 min
Response: 33854102
Conc: 71.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
BQ71242-1-2




#28 AR-1254-3

R.T.: 6.165 min
Delta R.T.: -0.004 min
Response: 17135128
Conc: 88.89 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: 0.000 min
Response: 26693990
Conc: 76.63 ng/ml



#29 AR-1254-4


R.T.: 6.392 min
Delta R.T.: -0.004 min
Response: 9882874
Conc: 81.49 ng/ml



#30 AR-1254-5

R.T.: 7.755 min
Delta R.T.: 0.000 min
Response: 29481776
Conc: 80.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
BQ71242-1-2



#30 AR-1254-5

R.T.: 6.811 min
Delta R.T.: -0.005 min
Response: 15383629
Conc: 87.01 ng/ml