

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0090418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 05 03:48:40 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0048696.D 1000 =P0048694.D 750 =P0048695.D
 250 =P0048697.D 50 =P0048698.D

Compound		500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.172	1.114	1.150	1.250	1.347	1.207 E5	7.72
2) SA	Decachlorobiphenyl	1.333	1.311	1.304	1.446	1.583	1.395 E5	8.57
3) L1	AR-1016-1	3.175	2.871	3.092	3.468	3.630	3.247 E3	9.32
4) L1	AR-1016-2	3.349	3.067	3.273	3.699	3.808	3.439 E3	8.93
5) L1	AR-1016-3	4.921	4.517	4.835	5.429	5.544	5.049 E3	8.48
6) L1	AR-1016-4	4.488	4.074	4.395	4.879	5.178	4.603 E3	9.37
7) L1	AR-1016-5	3.875	3.444	3.784	4.203	4.100	3.881 E3	7.65
8) L2	AR-1221-1	1.553					1.553 E3	0.00
9) L2	AR-1221-2	1.196					1.196 E3	0.00
10) L2	AR-1221-3	3.773					3.773 E3	0.00
11) L3	AR-1232-1	2.774					2.774 E3	0.00
12) L3	AR-1232-2	1.485					1.485 E3	0.00
13) L3	AR-1232-3	4.298					4.298 E2	0.00
14) L3	AR-1232-4	3.962					3.962 E2	0.00
15) L3	AR-1232-5	2.351					2.351 E2	0.00
16) L4	AR-1242-1	1.490					1.490 E3	0.00
17) L4	AR-1242-2	2.873					2.873 E3	0.00
18) L4	AR-1242-3	1.431					1.431 E3	0.00
19) L4	AR-1242-4	1.104					1.104 E3	0.00
20) L4	AR-1242-5	3.801					3.801 E3	0.00
21) L5	AR-1248-1	4.994					4.994 E3	0.00
22) L5	AR-1248-2	5.233					5.233 E3	0.00
23) L5	AR-1248-3	6.807					6.807 E3	0.00
24) L5	AR-1248-4	6.584					6.584 E3	0.00
25) L5	AR-1248-5	4.547					4.547 E3	0.00
26) L6	AR-1254-1	1.133					1.133 E4	0.00
27) L6	AR-1254-2	7.345					7.345 E3	0.00
28) L6	AR-1254-3	1.274					1.274 E4	0.00
29) L6	AR-1254-4	9.695					9.695 E3	0.00
30) L6	AR-1254-5	7.729					7.729 E3	0.00
31) L7	AR-1260-1	7.947	7.329	7.814	8.751	9.355	8.239 E3	9.79
32) L7	AR-1260-2	0.949	0.883	0.931	1.046	1.104	0.983 E4	9.17
33) L7	AR-1260-3	1.154	1.096	1.134	1.253	1.329	1.193 E4	7.99
34) L7	AR-1260-4	8.301	7.949	8.143	9.047	9.475	8.583 E3	7.56
35) L7	AR-1260-5	1.593	1.561	1.578	1.718	1.850	1.660 E4	7.41
36) L8	AR-1262-1	4.883					4.883 E3	0.00
37) L8	AR-1262-2	4.345					4.345 E3	0.00
38) L8	AR-1262-3	4.858					4.858 E3	0.00
39) L8	AR-1262-4	5.264					5.264 E3	0.00
40) L8	AR-1262-5	6.716					6.716 E3	0.00
41) L9	AR-1268-1	4.271					4.271 E3	0.00
42) L9	AR-1268-2	5.455					5.455 E3	0.00
43) L9	AR-1268-3	2.211					2.211 E4	0.00
44) L9	AR-1268-4	1.703					1.703 E4	0.00
45) L9	AR-1268-5	5.184					5.184 E4	0.00

Signal #2 Calibration Files

500 =P0048696.D 1000 =P0048694.D 750 =P0048695.D
 250 =P0048697.D 50 =P0048698.D

Compound		500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.535	1.323	1.449	1.614	1.895	1.563 E5	13.73
2) SA	Decachlorobiphenyl	1.410	1.289	1.328	1.540	1.826	1.479 E5	14.65
3) L1	AR-1016-1	4.278	3.552	3.962	4.607	5.269	4.334 E3	15.05
4) L1	AR-1016-2	6.387	5.297	5.975	6.880	7.656	6.439 E3	13.89
5) L1	AR-1016-3	8.368	7.139	7.955	9.080	9.677	8.444 E3	11.65

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0090418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 05 03:48:40 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0048696.D 1000 =P0048694.D 750 =P0048695.D
 250 =P0048697.D 50 =P0048698.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	5.992	5.092	5.682	6.464	6.143	5.875	E3	8.86
7)	L1 AR-1016-5	5.078	4.290	4.793	5.564	6.124	5.170	E3	13.64
8)	L2 AR-1221-1	2.141					2.141	E3	0.00
9)	L2 AR-1221-2	1.621					1.621	E3	0.00
10)	L2 AR-1221-3	4.925					4.925	E3	0.00
11)	L3 AR-1232-1	1.928					1.928	E3	0.00
12)	L3 AR-1232-2	6.644					6.644	E2	0.00
13)	L3 AR-1232-3	1.742					1.742	E3	0.00
14)	L3 AR-1232-4	1.849					1.849	E3	0.00
15)	L3 AR-1232-5	2.505					2.505	E2	0.00
16)	L4 AR-1242-1	1.963					1.963	E3	0.00
17)	L4 AR-1242-2	4.046					4.046	E3	0.00
18)	L4 AR-1242-3	1.432					1.432	E3	0.00
19)	L4 AR-1242-4	1.210					1.210	E3	0.00
20)	L4 AR-1242-5	3.092					3.092	E3	0.00
21)	L5 AR-1248-1	5.310					5.310	E3	0.00
22)	L5 AR-1248-2	9.603					9.603	E3	0.00
23)	L5 AR-1248-3	6.806					6.806	E3	0.00
24)	L5 AR-1248-4	6.075					6.075	E3	0.00
25)	L5 AR-1248-5	4.654					4.654	E3	0.00
26)	L6 AR-1254-1	1.015					1.015	E4	0.00
27)	L6 AR-1254-2	8.821					8.821	E3	0.00
28)	L6 AR-1254-3	1.102					1.102	E4	0.00
29)	L6 AR-1254-4	8.122					8.122	E3	0.00
30)	L6 AR-1254-5	1.516					1.516	E4	0.00
31)	L7 AR-1260-1	0.997	0.866	0.939	1.108	1.295	1.041	E4	16.06
32)	L7 AR-1260-2	1.195	1.051	1.127	1.324	1.565	1.252	E4	16.08
33)	L7 AR-1260-3	1.141	1.017	1.080	1.258	1.437	1.186	E4	13.97
34)	L7 AR-1260-4	0.909	0.820	0.855	1.001	1.145	0.946	E4	13.80
35)	L7 AR-1260-5	2.079	1.930	1.983	2.273	2.593	2.171	E4	12.40
36)	L8 AR-1262-1	5.505					5.505	E3	0.00
37)	L8 AR-1262-2	5.415					5.415	E3	0.00
38)	L8 AR-1262-3	5.965					5.965	E3	0.00
39)	L8 AR-1262-4	8.877					8.877	E3	0.00
40)	L8 AR-1262-5	7.564					7.564	E3	0.00
41)	L9 AR-1268-1	5.457					5.457	E3	0.00
42)	L9 AR-1268-2	2.377					2.377	E4	0.00
43)	L9 AR-1268-3	1.958					1.958	E4	0.00
44)	L9 AR-1268-4	5.617					5.617	E3	0.00
45)	L9 AR-1268-5	5.453					5.453	E4	0.00

(#) = Out of Range