

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090419\
 Data File : P0060360.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Sep 2019 4:13
 Operator : SM/AJ
 Sample : K4672-04 5X
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EXJC-1

Manual Integrations
 APPROVED

Ankita
 9/6/2019 9:08:13 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 05 05:08:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 30 18:12:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.106	3.488	270477	227830	7.725	6.329m
2) SA Decachlor...	9.608	8.378	1385066	15829905	26.035	392.597m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090419\
Data File : P0060360.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Sep 2019 4:13
Operator : SM/AJ
Sample : K4672-04 5X
Misc :
ALS Vial : 48 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
EXJC-1


Manual Integrations
APPROVED

Ankita
9/6/2019 9:08:13 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 05 05:08:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 30 18:12:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.106 min
Delta R.T.: -0.018 min
Response: 270477
Conc: 7.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
EXJC-1


Manual Integrations
APPROVED

Ankita
9/6/2019 9:08:13 AM




#1 Tetrachloro-m-xylene

R.T.: 3.488 min
Delta R.T.: -0.012 min
Response: 227830
Conc: 6.33 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.608 min
Delta R.T.: 0.015 min
Response: 1385066
Conc: 26.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.378 min
Delta R.T.: 0.019 min
Response: 15829905
Conc: 392.60 ng/ml m