

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090419\
 Data File : P0060368.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Sep 2019 6:23
 Operator : SM/AJ
 Sample : K4676-02 5X
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 DC-1

Manual Integrations
 APPROVED

Ankita
 9/6/2019 9:08:30 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 05 07:00:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 30 18:12:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.105	3.487	254750	218176	7.276	6.061
2) SA Decachlor...	9.554	8.335	5097313	303571	95.814m	7.529m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090419\
Data File : PO060368.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Sep 2019 6:23
Operator : SM/AJ
Sample : K4676-02 5X
Misc :
ALS Vial : 56 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
DC-1


Manual Integrations
APPROVED

Ankita
9/6/2019 9:08:30 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 05 07:00:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 30 18:12:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.105 min
Delta R.T.: -0.019 min
Response: 254750
Conc: 7.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
DC-1


Manual Integrations
APPROVED

Ankita
9/6/2019 9:08:30 AM




#1 Tetrachloro-m-xylene

R.T.: 3.487 min
Delta R.T.: -0.012 min
Response: 218176
Conc: 6.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.554 min
Delta R.T.: -0.039 min
Response: 5097313
Conc: 95.81 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.335 min
Delta R.T.: -0.024 min
Response: 303571
Conc: 7.53 ng/ml m