

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090519\
 Data File : P0060411.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Sep 2019 21:08
 Operator : SM/AJ
 Sample : K4694-10
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 T2-1-4-(0-1.5)

Manual Integrations
 APPROVED

Ankita
 9/6/2019 2:35:11 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 01:39:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 05 17:14:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.100	3.485	1206613	953940	34.462m	26.502
2)	SA Decachlor...	9.545	8.347	1164711	606583	21.893m	15.044 #
Target Compounds							
31)	L7 AR-1260-1	6.820	5.973	8890908	6202247	3298.036m	2721.590m
32)	L7 AR-1260-2	7.074	6.161	15476815	10685316	4171.109m	3647.021m
33)	L7 AR-1260-3	7.428	6.309	16514830	7730363	5040.645m	2918.069m#
34)	L7 AR-1260-4	7.655	6.770	15866938	10387000	5166.161m	4590.780m
35)	L7 AR-1260-5	7.961	7.014	48799418	39282998	6891.690m	7138.057m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090519\
Data File : PO060411.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Sep 2019 21:08
Operator : SM/AJ
Sample : K4694-10
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
T2-1-4-(0-1.5)

Manual Integrations
APPROVED

Ankita
9/6/2019 2:35:11 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 01:39:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 05 17:14:47 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

