

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0090523\
 Data File : P0097734.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Sep 2023 21:24
 Operator : YP/AJ
 Sample : PB155318BL
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 02:02:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0082823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 29 10:29:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.446	3.626	73988333	16679408	22.168	24.724
2)	SA Decachlor...	10.299	8.665	44204147	16715882	24.085	29.120
Target Compounds							
3)	L1 AR-1016-1	5.626	4.715	34463	115731	0.383	5.316 #
4)	L1 AR-1016-2	5.654	4.715	196793	115731	1.418	3.835 #
5)	L1 AR-1016-3	5.701	4.911	64860	11243	0.715	0.696
6)	L1 AR-1016-4	5.805	4.930	93684	27025	1.351	1.883 #
7)	L1 AR-1016-5	6.109	5.187	60421	122612	0.851	6.704 #
8)	L2 AR-1221-1	4.671	3.833	85517	15117	2.125	1.847
9)	L2 AR-1221-2	4.757	3.927	1626665	207982	54.115	35.885 #
10)	L2 AR-1221-3	4.829	4.015	48520	1013	0.545	0.055 #
11)	L3 AR-1232-1	4.829	4.015	48520	1013	0.658	0.067 #
12)	L3 AR-1232-2	5.345	4.715	45078	115731	1.176	8.290 #
13)	L3 AR-1232-3	5.654	4.911	196793	11243	3.043	1.601 #
14)	L3 AR-1232-4	5.805	5.000	93684	152353	2.941	21.548 #
15)	L3 AR-1232-5	5.901	5.187	103858	122612	3.530	15.577 #
16)	L4 AR-1242-1	5.626	4.715	34463	115731	0.426	5.999 #
17)	L4 AR-1242-2	5.654	4.715	196793	115731	1.576	4.252 #
18)	L4 AR-1242-3	5.701	4.911	64860	11243	0.789	0.783
19)	L4 AR-1242-4	5.805	5.000	93684	152353	1.492	9.619 #
20)	L4 AR-1242-5	6.587f	5.519	1076560	170147	17.117	9.641 #
21)	L5 AR-1248-1	5.626	4.715	34463	115731	0.584	8.422 #
22)	L5 AR-1248-2	5.901	4.930	103858	27025	1.050	1.219
23)	L5 AR-1248-3	6.109	5.000	60421	152353	0.583	6.626 #
24)	L5 AR-1248-4	6.506	5.187	41934	122612	0.458	4.518 #
25)	L5 AR-1248-5	6.587f	5.519f	1076560	170147	10.605	7.526 #
26)	L6 AR-1254-1	6.493	5.519	117491	170147	1.069	4.408 #
27)	L6 AR-1254-2	6.712	5.672	95593	672329	0.587	19.739 #
28)	L6 AR-1254-3	7.081	6.095	998816	182923	6.133	3.448 #
29)	L6 AR-1254-4	7.366	6.319	185081	138370	1.899	5.074 #
30)	L6 AR-1254-5	7.800	6.744	248323	266791	2.199	6.051 #
31)	L7 AR-1260-1	7.247	6.204	100880	71803	0.790	1.922 #
32)	L7 AR-1260-2	7.499	6.381	91955	247193	0.675	5.842 #
33)	L7 AR-1260-3	7.843	6.565	33836	130071	0.320	3.107 #
34)	L7 AR-1260-4	8.096	6.996f	65208	94408	0.576	2.919 #
35)	L7 AR-1260-5	8.427	7.271	86345	179453	0.459	2.718 #
36)	L8 AR-1262-1	7.843	6.824	33836	213261	0.225	10.626 #
37)	L8 AR-1262-2	8.427	6.996f	86345	94408	0.394	2.161 #
38)	L8 AR-1262-3	8.735	7.550	62435	152560	0.403	4.895 #
39)	L8 AR-1262-4	8.841	7.632	54914	89496	0.441	1.591 #
40)	L8 AR-1262-5	9.506	0.000	116267	0	1.691	N.D. #
41)	L9 AR-1268-1	8.735	7.550	62435	152560	0.220	1.594 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090523\
Data File : P0097734.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2023 21:24
Operator : YP/AJ
Sample : PB155318BL
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 02:02:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 29 10:29:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.841	7.632	54914	89496	0.216	1.080 #
43)	L9 AR-1268-3	9.072	0.000	158131	0	0.701	N.D. #
44)	L9 AR-1268-4	9.506	0.000	116267	0	1.511	N.D. #
45)	L9 AR-1268-5	9.950	8.407	484886	260214	0.702	1.141 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090523\
Data File : PO097734.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2023 21:24
Operator : YP/AJ
Sample : PB155318BL
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 02:02:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 29 10:29:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.446 min
Delta R.T.: -0.001 min
Response: 73988333
Conc: 22.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.626 min
Delta R.T.: -0.002 min
Response: 16679408
Conc: 24.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.299 min
Delta R.T.: 0.006 min
Response: 44204147
Conc: 24.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.665 min
Delta R.T.: -0.011 min
Response: 16715882
Conc: 29.12 ng/ml



#3 AR-1016-1


R.T.: 5.626 min
Delta R.T.: 0.002 min
Response: 34463
Conc: 0.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.715 min
Delta R.T.: -0.006 min
Response: 115731
Conc: 5.32 ng/ml




#4 AR-1016-2

R.T.: 5.654 min
Delta R.T.: 0.007 min
Response: 196793
Conc: 1.42 ng/ml



#4 AR-1016-2


R.T.: 4.715 min
Delta R.T.: -0.025 min
Response: 115731
Conc: 3.84 ng/ml



#5 AR-1016-3


R.T.: 5.701 min
Delta R.T.: -0.008 min
Response: 64860
Conc: 0.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.911 min
Delta R.T.: -0.007 min
Response: 11243
Conc: 0.70 ng/ml




#6 AR-1016-4

R.T.: 5.805 min
Delta R.T.: -0.002 min
Response: 93684
Conc: 1.35 ng/ml



#6 AR-1016-4


R.T.: 4.930 min
Delta R.T.: -0.030 min
Response: 27025
Conc: 1.88 ng/ml



#7 AR-1016-5


R.T.: 6.109 min
Delta R.T.: 0.004 min
Response: 60421
Conc: 0.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.187 min
Delta R.T.: 0.013 min
Response: 122612
Conc: 6.70 ng/ml




#8 AR-1221-1

R.T.: 4.671 min
Delta R.T.: 0.016 min
Response: 85517
Conc: 2.12 ng/ml



#8 AR-1221-1


R.T.: 3.833 min
Delta R.T.: -0.012 min
Response: 15117
Conc: 1.85 ng/ml



#9 AR-1221-2


R.T.: 4.757 min
Delta R.T.: 0.014 min
Response: 1626665
Conc: 54.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.927 min
Delta R.T.: -0.003 min
Response: 207982
Conc: 35.89 ng/ml




#10 AR-1221-3

R.T.: 4.829 min
Delta R.T.: 0.010 min
Response: 48520
Conc: 0.54 ng/ml



#10 AR-1221-3


R.T.: 4.015 min
Delta R.T.: 0.008 min
Response: 1013
Conc: 0.06 ng/ml



#11 AR-1232-1


R.T.: 4.829 min
Delta R.T.: 0.009 min
Response: 48520
Conc: 0.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.015 min
Delta R.T.: 0.007 min
Response: 1013
Conc: 0.07 ng/ml




#12 AR-1232-2

R.T.: 5.345 min
Delta R.T.: -0.008 min
Response: 45078
Conc: 1.18 ng/ml



#12 AR-1232-2


R.T.: 4.715 min
Delta R.T.: -0.025 min
Response: 115731
Conc: 8.29 ng/ml



#13 AR-1232-3


R.T.: 5.654 min
Delta R.T.: 0.007 min
Response: 196793
Conc: 3.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.911 min
Delta R.T.: -0.006 min
Response: 11243
Conc: 1.60 ng/ml




#14 AR-1232-4

R.T.: 5.805 min
Delta R.T.: -0.003 min
Response: 93684
Conc: 2.94 ng/ml



#14 AR-1232-4


R.T.: 5.000 min
Delta R.T.: -0.002 min
Response: 152353
Conc: 21.55 ng/ml



#15 AR-1232-5


R.T.: 5.901 min
Delta R.T.: 0.002 min
Response: 103858
Conc: 3.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.187 min
Delta R.T.: 0.013 min
Response: 122612
Conc: 15.58 ng/ml




#16 AR-1242-1

R.T.: 5.626 min
Delta R.T.: 0.002 min
Response: 34463
Conc: 0.43 ng/ml



#16 AR-1242-1


R.T.: 4.715 min
Delta R.T.: -0.006 min
Response: 115731
Conc: 6.00 ng/ml



#17 AR-1242-2


R.T.: 5.654 min
Delta R.T.: 0.007 min
Response: 196793
Conc: 1.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.715 min
Delta R.T.: -0.026 min
Response: 115731
Conc: 4.25 ng/ml




#18 AR-1242-3

R.T.: 5.701 min
Delta R.T.: -0.008 min
Response: 64860
Conc: 0.79 ng/ml



#18 AR-1242-3


R.T.: 4.911 min
Delta R.T.: -0.007 min
Response: 11243
Conc: 0.78 ng/ml



#19 AR-1242-4


R.T.: 5.805 min
Delta R.T.: -0.002 min
Response: 93684
Conc: 1.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.000 min
Delta R.T.: -0.001 min
Response: 152353
Conc: 9.62 ng/ml




#20 AR-1242-5

R.T.: 6.587 min
Delta R.T.: 0.036 min
Response: 1076560
Conc: 17.12 ng/ml



#20 AR-1242-5


R.T.: 5.519 min
Delta R.T.: -0.009 min
Response: 170147
Conc: 9.64 ng/ml



#21 AR-1248-1


R.T.: 5.626 min
Delta R.T.: 0.002 min
Response: 34463
Conc: 0.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.715 min
Delta R.T.: -0.006 min
Response: 115731
Conc: 8.42 ng/ml




#22 AR-1248-2

R.T.: 5.901 min
Delta R.T.: 0.002 min
Response: 103858
Conc: 1.05 ng/ml



#22 AR-1248-2


R.T.: 4.930 min
Delta R.T.: -0.029 min
Response: 27025
Conc: 1.22 ng/ml



#23 AR-1248-3


R.T.: 6.109 min
Delta R.T.: 0.004 min
Response: 60421
Conc: 0.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.000 min
Delta R.T.: 0.000 min
Response: 152353
Conc: 6.63 ng/ml




#24 AR-1248-4

R.T.: 6.506 min
Delta R.T.: -0.007 min
Response: 41934
Conc: 0.46 ng/ml



#24 AR-1248-4


R.T.: 5.187 min
Delta R.T.: 0.013 min
Response: 122612
Conc: 4.52 ng/ml



#25 AR-1248-5


R.T.: 6.587 min
Delta R.T.: 0.035 min
Response: 1076560
Conc: 10.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.519 min
Delta R.T.: -0.050 min
Response: 170147
Conc: 7.53 ng/ml




#26 AR-1254-1

R.T.: 6.493 min
Delta R.T.: 0.007 min
Response: 117491
Conc: 1.07 ng/ml



#26 AR-1254-1


R.T.: 5.519 min
Delta R.T.: -0.010 min
Response: 170147
Conc: 4.41 ng/ml



#27 AR-1254-2


R.T.: 6.712 min
Delta R.T.: 0.005 min
Response: 95593
Conc: 0.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.672 min
Delta R.T.: -0.005 min
Response: 672329
Conc: 19.74 ng/ml




#28 AR-1254-3

R.T.: 7.081 min
Delta R.T.: 0.005 min
Response: 998816
Conc: 6.13 ng/ml



#28 AR-1254-3


R.T.: 6.095 min
Delta R.T.: 0.012 min
Response: 182923
Conc: 3.45 ng/ml



#29 AR-1254-4


R.T.: 7.366 min
Delta R.T.: 0.002 min
Response: 185081
Conc: 1.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.319 min
Delta R.T.: 0.006 min
Response: 138370
Conc: 5.07 ng/ml




#30 AR-1254-5

R.T.: 7.800 min
Delta R.T.: 0.013 min
Response: 248323
Conc: 2.20 ng/ml



#30 AR-1254-5


R.T.: 6.744 min
Delta R.T.: 0.010 min
Response: 266791
Conc: 6.05 ng/ml



#31 AR-1260-1


R.T.: 7.247 min
Delta R.T.: 0.006 min
Response: 100880
Conc: 0.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.204 min
Delta R.T.: -0.010 min
Response: 71803
Conc: 1.92 ng/ml




#32 AR-1260-2

R.T.: 7.499 min
Delta R.T.: -0.002 min
Response: 91955
Conc: 0.67 ng/ml



#32 AR-1260-2


R.T.: 6.381 min
Delta R.T.: -0.023 min
Response: 247193
Conc: 5.84 ng/ml



#33 AR-1260-3


R.T.: 7.843 min
Delta R.T.: -0.021 min
Response: 33836
Conc: 0.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.565 min
Delta R.T.: 0.006 min
Response: 130071
Conc: 3.11 ng/ml




#34 AR-1260-4

R.T.: 8.096 min
Delta R.T.: 0.003 min
Response: 65208
Conc: 0.58 ng/ml



#34 AR-1260-4


R.T.: 6.996 min
Delta R.T.: -0.037 min
Response: 94408
Conc: 2.92 ng/ml



#35 AR-1260-5


R.T.: 8.427 min
Delta R.T.: 0.007 min
Response: 86345
Conc: 0.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.271 min
Delta R.T.: -0.004 min
Response: 179453
Conc: 2.72 ng/ml




#36 AR-1262-1

R.T.: 7.843 min
Delta R.T.: -0.021 min
Response: 33836
Conc: 0.22 ng/ml



#36 AR-1262-1


R.T.: 6.824 min
Delta R.T.: -0.002 min
Response: 213261
Conc: 10.63 ng/ml



#37 AR-1262-2


R.T.: 8.427 min
Delta R.T.: 0.007 min
Response: 86345
Conc: 0.39 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.996 min
Delta R.T.: -0.037 min
Response: 94408
Conc: 2.16 ng/ml




#38 AR-1262-3

R.T.: 8.735 min
Delta R.T.: -0.007 min
Response: 62435
Conc: 0.40 ng/ml



#38 AR-1262-3


R.T.: 7.550 min
Delta R.T.: -0.012 min
Response: 152560
Conc: 4.90 ng/ml



#39 AR-1262-4


R.T.: 8.841 min
Delta R.T.: 0.007 min
Response: 54914
Conc: 0.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.632 min
Delta R.T.: 0.007 min
Response: 89496
Conc: 1.59 ng/ml




#40 AR-1262-5

R.T.: 9.506 min
Delta R.T.: -0.004 min
Response: 116267
Conc: 1.69 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.122 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.735 min
Delta R.T.: -0.004 min
Response: 62435
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.550 min
Delta R.T.: -0.011 min
Response: 152560
Conc: 1.59 ng/ml




#42 AR-1268-2

R.T.: 8.841 min
Delta R.T.: 0.004 min
Response: 54914
Conc: 0.22 ng/ml



#42 AR-1268-2


R.T.: 7.632 min
Delta R.T.: 0.006 min
Response: 89496
Conc: 1.08 ng/ml



#43 AR-1268-3


R.T.: 9.072 min
Delta R.T.: 0.000 min
Response: 158131
Conc: 0.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.926 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.506 min
Delta R.T.: -0.004 min
Response: 116267
Conc: 1.51 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.122 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.950 min
Delta R.T.: 0.009 min
Response: 484886
Conc: 0.70 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.407 min
Delta R.T.: -0.009 min
Response: 260214
Conc: 1.14 ng/ml