

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090622\
 Data File : P0089296.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2022 12:29
 Operator : YP/AJ
 Sample : N4488-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OR-3-090122

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/07/2022
 Supervised By :Ankita Jodhani 09/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 13:18:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.585	66549159	20497562	17.748	21.235
2) SA Decachlor...	10.260	8.569	41275185	17120944	17.928	19.361

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090622\
Data File : PO089296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2022 12:29
Operator : YP/AJ
Sample : N4488-01
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
OR-3-090122


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/07/2022
Supervised By :Ankita Jodhani 09/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 13:18:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.428 min
Delta R.T.: -0.002 min
Response: 66549159
Conc: 17.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
OR-3-090122


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/07/2022
Supervised By :Ankita Jodhani 09/07/2022




#1 Tetrachloro-m-xylene

R.T.: 3.585 min
Delta R.T.: -0.004 min
Response: 20497562
Conc: 21.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.260 min
Delta R.T.: 0.003 min
Response: 41275185
Conc: 17.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.569 min
Delta R.T.: -0.003 min
Response: 17120944
Conc: 19.36 ng/ml