

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0090623.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 07 02:34:45 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0097777.D 1000 =P0097775.D 750 =P0097776.D
 250 =P0097778.D 50 =P0097779.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.188	4.065	4.077	4.226	3.864	4.084	E6	3.46
2) SA	Decachlorobiphenyl	3.354	3.178	3.219	3.433	3.120	3.261	E6	3.96
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	2.912	2.773	2.797	2.988	2.721	2.838	E5	3.85
42) L9	AR-1268-2	2.636	2.516	2.535	2.682	2.402	2.554	E5	4.29
43) L9	AR-1268-3	2.274	2.206	2.222	2.321	2.151	2.235	E5	2.93
44) L9	AR-1268-4	8.556	8.191	8.212	8.525	7.377	8.172	E4	5.83
45) L9	AR-1268-5	6.829	6.664	6.666	6.889	6.118	6.633	E5	4.59

Signal #2 Calibration Files

500 =P0097777.D 1000 =P0097775.D 750 =P0097776.D
 250 =P0097778.D 50 =P0097779.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.024	1.000	1.006	1.021	0.894	0.989	E6	5.46
2) SA	Decachlorobiphenyl	1.535	1.494	1.494	1.545	1.426	1.499	E6	3.14
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0090623.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 07 02:34:45 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0097777.D 1000 =P0097775.D 750 =P0097776.D
 250 =P0097778.D 50 =P0097779.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.339	1.314	1.314	1.349	1.266	1.316 E5	2.43
42) L9	AR-1268-2	1.179	1.165	1.159	1.186	1.073	1.153 E5	3.96
43) L9	AR-1268-3	2.826	2.700	2.714	2.896	2.686	2.764 E4	3.33
44) L9	AR-1268-4	3.646	3.571	3.525	3.719	3.152	3.523 E4	6.24
45) L9	AR-1268-5	3.226	3.229	3.183	3.200	2.838	3.135 E5	5.33

(#) = Out of Range