

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090623\
 Data File : P0097762.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2023 19:32
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 01:15:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 01:05:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.443	3.623	104.3E6	25577468	25.555	25.474
2)	SA Decachlor...	10.295	8.662	49434569	22393480	26.017	25.830
Target Compounds							
16)	L4 AR-1242-1	5.621	4.714	24624357	7462192	264.133	264.820
17)	L4 AR-1242-2	5.644	4.733	37522814	10456202	267.564	262.749
18)	L4 AR-1242-3	5.706	4.910	23354024	5714675	268.095	265.755
19)	L4 AR-1242-4	5.805	4.994	17981444	6219014	263.244	271.978
20)	L4 AR-1242-5	6.549	5.520	19057377	7001766	277.154	267.291



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090623\
Data File : PO097762.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2023 19:32
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 01:15:36 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 01:05:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.443 min
Delta R.T.: 0.000 min
Response: 104292991
Conc: 25.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: 0.000 min
Response: 25577468
Conc: 25.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.295 min
Delta R.T.: 0.000 min
Response: 49434569
Conc: 26.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.662 min
Delta R.T.: 0.000 min
Response: 22393480
Conc: 25.83 ng/ml



#16 AR-1242-1


R.T.: 5.621 min
Delta R.T.: 0.000 min
Response: 24624357
Conc: 264.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.714 min
Delta R.T.: 0.000 min
Response: 7462192
Conc: 264.82 ng/ml




#17 AR-1242-2

R.T.: 5.644 min
Delta R.T.: 0.000 min
Response: 37522814
Conc: 267.56 ng/ml



#17 AR-1242-2


R.T.: 4.733 min
Delta R.T.: 0.000 min
Response: 10456202
Conc: 262.75 ng/ml



#18 AR-1242-3


R.T.: 5.706 min
Delta R.T.: 0.000 min
Response: 23354024
Conc: 268.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.910 min
Delta R.T.: 0.000 min
Response: 5714675
Conc: 265.75 ng/ml




#19 AR-1242-4

R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 17981444
Conc: 263.24 ng/ml



#19 AR-1242-4


R.T.: 4.994 min
Delta R.T.: 0.000 min
Response: 6219014
Conc: 271.98 ng/ml



#20 AR-1242-5

R.T.: 6.549 min
Delta R.T.: 0.000 min
Response: 19057377
Conc: 277.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 7001766
Conc: 267.29 ng/ml