

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090624\
 Data File : P0106297.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2024 08:04
 Operator : YP/AJ
 Sample : P3851-02
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC122669-1-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/09/2024
 Supervised By :Ankita Jodhani 09/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 08:58:05 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.456	3.722	105.8E6	37766757	11.601	13.979
2) SA Decachlor...	10.205	8.758	79246498	27397655	14.602	12.550
Target Compounds						
31) L7 AR-1260-1	7.220	6.299	5230035	2763856	14.574	20.979 #
32) L7 AR-1260-2	7.477	6.486	8409132	2550581	20.019	15.542
33) L7 AR-1260-3	7.833	6.639	6759119	3500238	25.272	20.740
34) L7 AR-1260-4	8.058	7.112	3387868	1174289	10.928m	10.596
35) L7 AR-1260-5	8.380	7.351	11270631	3745292	20.259	13.803 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090624\
Data File : PO106297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2024 08:04
Operator : YP/AJ
Sample : P3851-02
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BC122669-1-2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/09/2024
Supervised By :Ankita Jodhani 09/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 08:58:05 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 06:50:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

